

Data Path : Z:\HPCHEM1\BNA M\DATA\BM110515\
 Data File : BM003060.D
 Acq On : 05 Nov 2015 12:34
 Operator : UM/IZ
 Sample : SSTDC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDC001

Quant Time: Nov 16 18:54:24 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM110415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 16 18:52:15 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	120	0.00
2	1,4-Dioxane	0.510	0.454	11.0	116	0.00
3	n-Nitrosodimethylamine	0.469	0.440	6.2	117	0.00
4 S	2-Fluorophenol	1.058	0.994	6.0	118	0.00
5 S	Phenol-d6	1.234	1.165	5.6	120	0.00
6	bis(2-Chloroethyl)ether	1.183	1.129	4.6	118	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	120	0.00
8 S	Nitrobenzene-d5	0.231	0.222	3.9	122	0.00
9	Nitrobenzene	0.249	0.240	3.6	122	0.00
10	Naphthalene	1.075	1.044	2.9	118	0.00
11	Hexachlorobutadiene	0.166	0.165	0.6	120	0.00
12	2-Methylnaphthalene	0.695	0.671	3.5	117	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	120	0.00
14 S	2,4,6-Tribromophenol	0.129	0.122	5.4	115	0.00
15 S	2-Fluorobiphenyl	1.576	1.509	4.3	116	0.00
16	Acenaphthylene	1.946	1.819	6.5	117	0.00
17	Acenaphthene	1.339	1.284	4.1	125	0.00
18	Fluorene	1.426	1.341	6.0	110	0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	150#	0.00
20	4-Bromophenyl-phenylether	0.173	0.141	18.5	94	0.00
21	Hexachlorobenzene	0.247	0.210	15.0	104	0.00
22	Pentachlorophenol	0.076	0.062	18.4	112	0.00
23	Phenanthrene	1.134	0.922	18.7	95	0.00
24	Anthracene	1.060	0.990	6.6	140	0.00
25	Fluoranthene	1.127	0.983	12.8	121	0.01
26 I	Chrysene-d12	1.000	1.000	0.0	133	0.00
27	Benzydine	0.454	0.281	38.1#	86	0.00
28	Pvrene	1.452	1.294	10.9	122	0.00
29 S	Terphenyl-d14	0.721	0.662	8.2	119	0.00
30	Benzo(a)anthracene	1.406	1.188	15.5	127	0.00
31	3,3'-Dichlorobenzidine	0.367	0.266	27.5#	97	0.00
32	Chrysene	1.562	1.260	19.3	110	0.01
33	Bis(2-ethylhexyl)phthalate	0.562	0.475	15.5	132	0.01
34	Indeno(1,2,3-cd)pyrene	1.385	1.248	9.9	120	0.00
35 I	Perylene-d12	1.000	1.000	0.0	121	0.00
36	Benzo(b)fluoranthene	1.707	1.420	16.8	115	0.00
37	Benzo(k)fluoranthene	1.542	1.297	15.9	115	0.00
38 C	Benzo(a)pyrene	1.283	1.145	10.8	119	0.00
39	Dibenzo(a,h)anthracene	1.249	1.194	4.4	121	0.00
40	Benzo(a,h,i)perylene	1.349	1.244	7.8	119	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM110515\
Data File : BM003060.D
Acq On : 05 Nov 2015 12:34
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Nov 16 18:54:24 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM110415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 16 18:52:15 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				