

Data Path : Z:\HPCHEM1\BNA M\DATA\BM110515\
 Data File : BM003068.D
 Acq On : 05 Nov 2015 17:34
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Nov 17 13:18:23 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM110415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 17 12:50:51 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	133	0.00
2	1,4-Dioxane	0.510	0.448	12.2	126	0.00
3	n-Nitrosodimethylamine	0.469	0.449	4.3	132	0.00
4 S	2-Fluorophenol	1.058	1.040	1.7	137	0.00
5 S	Phenol-d6	1.234	1.229	0.4	141	0.00
6	bis(2-Chloroethyl)ether	1.183	1.144	3.3	133	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	136	0.00
8 S	Nitrobenzene-d5	0.231	0.235	-1.7	146	0.00
9	Nitrobenzene	0.249	0.252	-1.2	145	0.00
10	Naphthalene	1.075	1.044	2.9	134	0.00
11	Hexachlorobutadiene	0.166	0.165	0.6	136	0.00
12	2-Methylnaphthalene	0.695	0.682	1.9	135	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	138	0.00
14 S	2,4,6-Tribromophenol	0.129	0.130	-0.8	141	0.00
15 S	2-Fluorobiphenyl	1.576	1.515	3.9	134	0.00
16	Acenaphthylene	1.946	1.942	0.2	144	0.00
17	Acenaphthene	1.339	1.305	2.5	146	0.00
18	Fluorene	1.426	1.355	5.0	128	0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	169#	0.00
20	4-Bromophenyl-phenylether	0.173	0.147	15.0	109	0.00
21	Hexachlorobenzene	0.247	0.213	13.8	119	0.00
22	Pentachlorophenol	0.076	0.069	9.2	141	0.00
23	Phenanthrene	1.134	0.914	19.4	106	0.00
24	Anthracene	1.060	1.044	1.5	166#	0.00
25	Fluoranthene	1.127	0.992	12.0	137	0.01
26 I	Chrysene-d12	1.000	1.000	0.0	151#	0.00
27	Benzydine	0.454	0.375	17.4	130	0.00
28	Pvrene	1.452	1.315	9.4	140	0.00
29 S	Terphenyl-d14	0.721	0.661	8.3	134	0.00
30	Benzo(a)anthracene	1.406	1.222	13.1	148	0.00
31	3,3'-Dichlorobenzidine	0.367	0.312	15.0	130	0.00
32	Chrysene	1.562	1.215	22.2	120	0.01
33	Bis(2-ethylhexyl)phthalate	0.562	0.532	5.3	167#	0.01
34	Indeno(1,2,3-cd)pyrene	1.385	1.333	3.8	145	0.00
35 I	Perylene-d12	1.000	1.000	0.0	141	0.00
36	Benzo(b)fluoranthene	1.707	1.449	15.1	137	0.00
37	Benzo(k)fluoranthene	1.542	1.273	17.4	132	0.00
38 C	Benzo(a)pyrene	1.283	1.190	7.2	144	0.00
39	Dibenzo(a,h)anthracene	1.249	1.235	1.1	147	0.00
40	Benzo(a,h,i)perylene	1.349	1.281	5.0	143	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM110515\
Data File : BM003068.D
Acq On : 05 Nov 2015 17:34
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Nov 17 13:18:23 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM110415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 17 12:50:51 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				