


Data Path : Z:\HPCHEM1\BNA_M\Data\BM110517\
Data File : BM012741.D
Acq On : 05 Nov 2017 15:44
Operator : SJ/JU
Sample : DFTPP42
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP42

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM110417MA.M
Title : SVOA CALIBRATION
Last Update : Tue Nov 07 04:13:22 2017



AutoFind: Scans 2473, 2474, 2475; Background Corrected with Scan 2467

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	24.9	50728	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	13.8	28024	PASS
70	69	0.00	2	0.4	109	PASS
127	198	10	80	39.3	80133	PASS
197	198	0.00	2	0.8	1596	PASS
198	198	100	100	100.0	203690	PASS
199	198	5	9	7.0	14190	PASS
275	198	10	60	33.1	67357	PASS
365	198	1	100	4.2	8479	PASS
441	442	0.01	100	15.0	26592	PASS
442	198	50	100	86.9	177002	PASS
443	442	15	24	19.1	33832	PASS

m/z	Abundance
38.00	933.0
39.10	4501.0
40.90	392.0
43.10	492.0
43.90	376.0
49.10	545.0
50.00	16576.0
51.00	58416.0
52.00	3295.0
55.00	760.0
56.00	2257.0
57.00	4200.0
58.00	342.0
60.80	723.0
62.00	801.0
63.00	1683.0
64.00	302.0
64.90	706.0
69.00	32192.0
73.00	599.0
74.00	4336.0
75.00	6456.0
76.20	2225.0
77.00	41296.0
78.00	3016.0
79.00	3986.0
80.00	2803.0
81.00	3655.0
82.00	872.0
82.90	955.0
85.00	1013.0
85.90	1094.0
87.00	620.0
90.90	1377.0
92.00	1410.0
93.00	8384.0
94.00	787.0
95.10	471.0
96.00	572.0
96.90	427.0
98.00	7157.0
99.00	5666.0
99.90	402.0
100.90	3194.0
102.90	990.0
104.00	2651.0
105.00	2393.0
105.90	481.0
107.00	25280.0
108.00	4374.0
109.10	989.0
110.00	48888.0
111.00	7674.0
112.00	1066.0
112.90	311.0
116.10	1426.0
117.00	23184.0
118.00	1622.0
120.00	335.0
122.00	2082.0
123.00	2981.0
124.00	1482.0
125.00	1134.0
127.00	88504.0
128.00	6612.0
129.00	35848.0
130.10	2881.0
131.00	679.0
133.00	307.0
133.90	967.0
135.00	3470.0
136.00	1052.0
136.90	1700.0
137.90	605.0
139.00	312.0
140.10	561.0
141.00	4939.0
142.00	1559.0
143.00	1157.0
144.10	372.0
145.10	363.0
146.00	737.0
147.00	2387.0
148.00	5662.0
148.90	1204.0
150.00	438.0
151.10	542.0
152.00	532.0
153.00	1672.0
154.00	1249.0
155.00	3391.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42

156.00	4169.0
157.20	754.0
157.90	921.0
158.90	707.0
160.00	2126.0
161.00	2273.0
161.70	778.0
162.90	374.0
164.90	1793.0
166.10	1614.0
167.00	10089.0
168.00	4198.0
169.00	1004.0
169.90	368.0
170.80	469.0
171.90	1016.0
172.90	1214.0
174.00	2293.0
175.00	3829.0
175.90	982.0
177.00	2011.0
177.90	705.0
179.00	8163.0
180.00	5112.0
180.90	2495.0
181.90	419.0
182.80	326.0
183.90	516.0
185.00	3630.0
186.00	27656.0
187.00	8624.0
188.10	975.0
188.90	2260.0
190.00	492.0
190.90	1129.0
191.90	2387.0
193.00	2869.0
194.00	537.0
196.00	5997.0
197.10	1787.0
197.90	209984.0
198.90	14901.0
199.90	1289.0
201.40	873.0
203.00	1539.0
204.00	7801.0
205.00	12389.0
206.00	55864.0
207.10	9902.0
208.00	2360.0
208.90	854.0
210.30	940.0
211.00	2262.0
215.00	538.0
216.00	1044.0
216.90	14931.0
218.00	1960.0
221.00	9325.0
221.80	2214.0
223.00	3273.0
224.00	32040.0
225.00	7724.0
226.00	842.0
227.00	13408.0
228.00	2214.0
229.00	2582.0
230.00	332.0
231.00	1482.0
232.00	392.0
234.10	935.0
235.00	955.0
236.00	881.0
236.90	1165.0
238.90	806.0
240.00	407.0
240.90	680.0
241.90	1791.0
243.00	1749.0
244.00	24552.0
245.00	3698.0
246.00	5389.0
246.90	816.0
249.00	794.0
250.00	365.0
253.00	521.0
254.20	1154.0
255.00	136448.0
256.00	20448.0
256.90	1531.0
257.90	7980.0
258.90	1193.0
265.00	3657.0
266.00	632.0
272.10	375.0
272.90	4205.0

Instrument :

BNA_M

ClientSampleId :

DFTPP42

274.00	10302.0
275.00	63936.0
276.00	8916.0
277.00	5164.0
277.90	1056.0
281.00	538.0
283.00	738.0
283.90	441.0
285.00	895.0
293.00	1102.0
295.00	412.0
296.00	17304.0
297.00	2920.0
301.90	372.0
303.00	2287.0
304.00	547.0
307.90	301.0
310.00	365.0
314.10	588.0
315.00	2118.0
316.00	1135.0
321.00	490.0
323.00	6261.0
324.00	1108.0
326.90	1352.0
327.90	466.0
332.00	509.0
333.00	609.0
334.00	3730.0
335.00	1031.0
341.10	979.0
346.10	975.0
352.00	1673.0
353.10	916.0
354.10	2111.0
354.90	447.0
365.00	7255.0
366.00	847.0
371.00	504.0
372.10	3022.0
373.10	575.0
382.90	627.0
390.00	376.0
402.00	1302.0
403.00	1612.0
404.00	795.0
421.10	1109.0
422.10	922.0
423.10	7836.0
424.10	1604.0
441.10	20240.0
442.10	141504.0
443.10	27104.0
444.00	2310.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42

m/z	Abundance
38.00	921.0
39.00	4103.0
40.00	386.0
41.00	399.0
43.10	568.0
44.00	388.0
48.90	545.0
50.00	15420.0
51.00	60776.0
52.00	2897.0
55.10	872.0
56.00	2110.0
57.00	4156.0
58.00	345.0
60.90	703.0
62.00	832.0
63.10	1963.0
65.00	835.0
69.00	33360.0
70.00	329.0
73.00	533.0
74.00	4555.0
75.00	6264.0
76.10	2266.0
77.00	43048.0
78.00	2715.0
79.00	3923.0
80.00	2755.0
81.00	4203.0
82.00	1059.0
83.10	920.0
85.00	959.0
86.00	1632.0
87.10	665.0
90.90	1447.0
92.00	1404.0
93.00	8587.0
93.90	1003.0
95.00	330.0
95.90	633.0
97.10	380.0
98.00	7533.0
99.00	5605.0
99.90	623.0
101.00	3158.0
102.90	1227.0
104.00	2064.0
105.00	2411.0
106.10	951.0
107.00	28888.0
108.00	4345.0
109.00	1221.0
110.00	52808.0
111.00	7717.0
112.00	1050.0
112.90	464.0
116.00	1972.0
117.00	26592.0
118.00	1972.0
118.90	441.0
119.90	471.0
122.00	2275.0
123.00	3113.0
124.00	1726.0
125.10	1474.0
126.20	818.0
127.00	94904.0
128.00	8042.0
129.00	39632.0
130.00	3251.0
130.90	837.0
133.00	323.0
134.00	1419.0
135.00	3562.0
135.90	1372.0
137.10	1683.0
137.70	422.0
140.10	543.0
141.00	5732.0
142.00	1535.0
142.90	1208.0
143.90	373.0
145.00	372.0
146.10	939.0
147.00	3356.0
148.00	6208.0
148.90	1109.0
150.00	359.0
151.00	754.0
152.00	488.0
153.00	1875.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42

153.90	1184.0
155.00	2964.0
156.00	4789.0
157.00	912.0
158.00	1195.0
159.00	804.0
160.00	1862.0
161.00	2654.0
162.00	657.0
164.00	365.0
165.00	2721.0
166.00	1730.0
167.00	10983.0
168.00	5244.0
169.00	1013.0
170.10	493.0
170.90	582.0
172.00	1013.0
173.00	1223.0
174.00	2314.0
175.10	3958.0
176.00	1224.0
177.00	2671.0
178.00	784.0
179.00	9223.0
180.00	5926.0
181.00	3032.0
182.10	477.0
183.90	802.0
185.00	4705.0
186.00	33256.0
187.00	9822.0
187.90	1058.0
188.90	2292.0
190.00	531.0
190.90	1025.0
191.90	3313.0
193.00	3119.0
193.90	889.0
195.10	308.0
195.90	6621.0
197.10	1877.0
197.90	244224.0
199.00	16440.0
199.90	1282.0
201.40	1289.0
203.00	1801.0
204.00	8679.0
205.00	14706.0
206.00	66512.0
207.00	9203.0
208.00	2732.0
209.10	820.0
210.20	1267.0
210.90	2818.0
215.00	757.0
216.00	1446.0
217.00	17248.0
218.00	2128.0
221.00	10551.0
221.80	2484.0
223.00	3939.0
224.00	39472.0
225.00	9801.0
226.00	1087.0
227.00	15486.0
228.00	2171.0
228.90	3140.0
229.90	473.0
231.00	1522.0
234.00	1103.0
234.90	1189.0
235.90	982.0
237.00	1264.0
239.00	895.0
240.00	596.0
240.90	1014.0
242.00	1976.0
243.00	2132.0
244.00	30528.0
245.10	3966.0
246.00	6500.0
246.90	1293.0
248.90	1393.0
250.00	342.0
250.90	313.0
253.10	720.0
254.10	1878.0
255.00	167424.0
256.00	26408.0
257.10	1697.0
258.00	9727.0
258.90	1895.0
264.00	301.0
265.00	4104.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42

265.90	762.0
269.90	307.0
271.00	416.0
271.90	432.0
273.00	5085.0
274.00	13770.0
275.00	79480.0
276.00	10207.0
277.00	6760.0
277.90	933.0
281.00	797.0
281.90	340.0
283.00	549.0
284.10	638.0
284.90	1387.0
290.00	333.0
291.90	432.0
293.00	1407.0
294.00	344.0
295.00	637.0
296.00	22992.0
297.00	2956.0
301.90	387.0
303.00	2670.0
304.00	822.0
307.90	375.0
314.00	1117.0
315.00	2468.0
316.00	1390.0
321.00	649.0
322.00	429.0
323.00	7647.0
324.00	1294.0
327.00	1470.0
328.00	1219.0
332.00	497.0
333.00	733.0
334.00	5006.0
335.00	1353.0
341.00	1078.0
345.90	1648.0
352.00	2273.0
353.00	1428.0
354.00	2487.0
355.10	589.0
365.00	10359.0
366.00	1515.0
371.00	597.0
372.10	3504.0
373.00	1014.0
383.00	744.0
390.00	391.0
402.00	1373.0
403.00	2014.0
404.00	806.0
421.00	1916.0
422.00	1562.0
423.10	12407.0
424.00	2232.0
425.10	319.0
441.10	31912.0
442.10	206912.0
443.10	39968.0
444.00	3775.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42

m/z	Abundance
38.00	335.0
39.00	2528.0
41.00	591.0
43.00	561.0
43.90	349.0
50.10	7718.0
51.00	32992.0
52.00	1572.0
55.00	757.0
55.90	1235.0
57.00	2372.0
60.90	323.0
62.00	393.0
63.00	1175.0
65.10	418.0
69.00	18520.0
73.00	530.0
74.10	2310.0
74.90	3881.0
76.00	960.0
77.00	24664.0
78.00	1792.0
79.00	1991.0
80.00	1245.0
81.00	2180.0
82.00	727.0
82.90	657.0
85.00	672.0
86.00	687.0
86.80	365.0
91.00	837.0
92.00	1036.0
93.00	5015.0
93.90	371.0
94.90	338.0
96.90	387.0
98.00	4488.0
99.00	2744.0
100.90	2163.0
102.90	481.0
103.90	1316.0
104.90	1330.0
106.00	444.0
107.00	15564.0
108.00	2144.0
109.00	920.0
110.00	29672.0
111.00	5142.0
112.10	545.0
116.00	1411.0
117.00	14587.0
118.00	1043.0
119.00	357.0
121.90	1261.0
123.00	2112.0
124.00	769.0
125.00	810.0
127.00	56992.0
128.00	4277.0
129.00	22664.0
130.00	1477.0
131.00	326.0
133.90	542.0
135.00	2094.0
136.10	761.0
136.90	838.0
140.00	360.0
141.10	3183.0
142.00	1266.0
142.90	1032.0
146.00	652.0
147.00	1822.0
148.00	3329.0
149.00	795.0
151.10	502.0
153.00	1233.0
154.00	756.0
155.00	1834.0
156.00	2774.0
157.00	637.0
158.00	597.0
159.00	687.0
160.00	1152.0
161.00	1664.0
162.00	528.0
165.00	1605.0
166.00	1002.0
167.00	6795.0
168.00	2837.0
169.00	625.0
170.90	346.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42

172.00	801.0
173.00	813.0
174.00	1289.0
175.00	2471.0
176.00	772.0
177.00	1614.0
178.00	593.0
179.00	5732.0
180.00	3752.0
181.00	1444.0
182.00	336.0
184.10	566.0
185.00	2733.0
186.00	20576.0
187.00	5906.0
188.00	694.0
189.00	1513.0
190.10	318.0
191.00	628.0
191.90	1644.0
193.10	2360.0
194.00	471.0
194.90	424.0
196.00	4592.0
197.10	1126.0
197.90	156864.0
198.90	11230.0
200.00	839.0
201.50	961.0
203.00	1053.0
204.00	5855.0
205.00	10559.0
206.00	43336.0
207.10	6698.0
208.00	1928.0
209.00	814.0
210.00	717.0
211.00	1793.0
214.90	472.0
216.00	1143.0
217.00	11318.0
217.90	1566.0
220.10	302.0
221.00	6416.0
221.90	1447.0
223.00	2564.0
224.00	24064.0
225.00	6049.0
226.20	488.0
227.00	10847.0
228.00	1313.0
229.00	2005.0
230.10	398.0
230.90	1041.0
234.00	736.0
235.00	866.0
235.90	494.0
237.00	949.0
239.10	597.0
239.90	460.0
241.00	752.0
242.00	1437.0
243.00	1498.0
244.00	20296.0
245.00	3024.0
246.00	4549.0
247.00	1095.0
248.90	951.0
252.10	349.0
252.90	537.0
253.90	856.0
254.20	948.0
255.00	118032.0
256.00	16302.0
257.00	1146.0
258.00	6749.0
258.90	636.0
265.00	2743.0
265.90	606.0
272.00	326.0
273.00	3813.0
274.00	9907.0
275.00	58656.0
276.00	7560.0
277.00	5285.0
278.00	552.0
281.10	571.0
282.10	458.0
283.00	641.0
283.90	456.0
284.90	642.0
292.90	1286.0
295.10	412.0
296.00	16552.0
297.00	1991.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42

303.00	2101.0
303.90	524.0
314.00	911.0
315.00	2100.0
316.10	912.0
321.10	475.0
322.00	355.0
323.00	5715.0
324.00	804.0
327.00	926.0
328.00	476.0
331.90	546.0
333.00	645.0
334.00	3872.0
335.00	1038.0
340.90	633.0
346.00	1751.0
347.10	335.0
352.00	1806.0
353.00	1279.0
354.00	1615.0
355.10	632.0
365.00	7825.0
365.90	966.0
371.00	527.0
372.10	2707.0
373.10	690.0
383.00	1022.0
390.00	418.0
402.10	1020.0
403.00	1848.0
404.00	689.0
421.00	1410.0
422.00	1194.0
423.00	9869.0
424.10	2107.0
441.10	27624.0
442.10	182592.0
443.10	34424.0
444.10	3456.0

Instrument :

BNA_M

ClientSampleId :

DFTPP42

Scan 2467 (17.598 min): BM012741.D
DFTPP42

m/z	Abundance
41.00	355.0
43.00	533.0
43.90	390.0
55.00	398.0
57.00	369.0
109.00	424.0
207.00	1404.0
281.10	391.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42