

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM110624\  
 Data File : BM048478.D  
 Acq On : 06 Nov 2024 15:21  
 Operator : RC/JU  
 Sample : PB164609BL  
 Misc :  
 ALS Vial : 10 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SBLK609

Manual Integrations  
 APPROVED

Reviewed By :Yogesh Patel 11/07/2024  
 Supervised By :mohammad ahmed 11/11/2024

Quant Time: Nov 06 15:56:45 2024  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_M\Methods\SFAM-EPA-SIM-BM110624.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Wed Nov 06 14:39:10 2024  
 Response via : Initial Calibration

| Compound                  | R.T.   | QIon | Response | Conc  | Units | Dev(Min) |
|---------------------------|--------|------|----------|-------|-------|----------|
| -----                     |        |      |          |       |       |          |
| Internal Standards        |        |      |          |       |       |          |
| 1) 1,4-Dichlorobenzene-d4 | 7.678  | 152  | 4054     | 0.400 | ng/ul | 0.00     |
| 4) Naphthalene-d8         | 10.464 | 136  | 12792    | 0.400 | ng/ul | 0.00     |
| 9) Acenaphthene-d10       | 14.322 | 164  | 6778     | 0.400 | ng/ul | 0.00     |
| 13) Phenanthrene-d10      | 17.069 | 188  | 13202m   | 0.400 | ng/ul | -0.01    |
| 17) Chrysene-d12          | 21.277 | 240  | 10258m   | 0.400 | ng/ul | 0.01     |
| 23) Perylene-d12          | 23.487 | 264  | 8337m    | 0.400 | ng/ul | 0.00     |

|                             |        |     |       |       |       |      |
|-----------------------------|--------|-----|-------|-------|-------|------|
| System Monitoring Compounds |        |     |       |       |       |      |
| 3) 1,4-Dioxane-d8           | 3.134  | 96  | 23589 | 4.823 | ng/ul | 0.00 |
| 6) 2-Methylnaphthalene-d10  | 12.086 | 152 | 4435  | 0.268 | ng/ul | 0.01 |
| 18) Fluoranthene-d10        | 19.100 | 212 | 8865  | 0.308 | ng/ul | 0.00 |

|                  |        |
|------------------|--------|
| Target Compounds | Qvalue |
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM110624\  
Data File : BM048478.D  
Acq On : 06 Nov 2024 15:21  
Operator : RC/JU  
Sample : PB164609BL  
Misc :  
ALS Vial : 10 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
SBLK609

Manual Integrations  
APPROVED

Reviewed By :Yogesh Patel 11/07/2024  
Supervised By :mohammad ahmed 11/11/2024

Quant Time: Nov 06 15:56:45 2024  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_M\Methods\SFAM-EPA-SIM-BM110624.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Wed Nov 06 14:39:10 2024  
Response via : Initial Calibration

