

Data Path : Z:\HPCHEM1\BNA M\DATA\BM110716\
 Data File : BM007764.D
 Acq On : 07 Nov 2016 09:12
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02036

Quant Time: Nov 07 10:11:58 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM102016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 24 12:55:22 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
2	1,4-Dioxane	0.510	0.503	1.4	89	0.00
3 S	1,4-Dioxane-d8	0.467	0.479	-2.6	90	0.00
4	Benzaldehyde	0.976	1.154	-18.2	95	0.00
5 S	Phenol-d5	1.544	1.569	-1.6	95	0.00
6	Phenol	1.584	1.607	-1.5	93	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.916	0.937	-2.3	93	0.00
8	Bis(2-Chloroethyl)ether	1.221	1.256	-2.9	93	0.00
9 S	2-Chlorophenol-d4	1.184	1.270	-7.3	96	0.00
10	2-Chlorophenol	1.215	1.268	-4.4	93	0.00
11	2-Methylphenol	1.153	1.168	-1.3	95	0.00
12	2,2'-oxybis(1-Chloropropane	1.353	1.432	-5.8	95	0.00
13 S	4-Methylphenol-d8	1.208	1.223	-1.2	93	0.00
14	Acetophenone	1.894	1.964	-3.7	95	0.00
15 P	N-Nitroso-di-n-propylamine	0.940	1.022	-8.7	99	0.00
16	4-Methylphenol	1.243	1.286	-3.5	95	0.00
17	Hexachloroethane	0.537	0.549	-2.2	90	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	91	0.00
19 S	Nitrobenzene-d5	0.144	0.152	-5.6	94	0.00
20	Nitrobenzene	0.370	0.395	-6.8	97	0.00
21	Isophorone	0.633	0.688	-8.7	99	0.00
22 S	2-Nitrophenol-d4	0.153	0.160	-4.6	96	0.00
23 C	2-Nitrophenol	0.159	0.173	-8.8	96	0.00
24	2,4-Dimethylphenol	0.364	0.387	-6.3	96	0.00
25	Bis(2-Chloroethoxy)methane	0.387	0.403	-4.1	94	0.00
26 S	2,4-Dichlorophenol-d3	0.300	0.321	-7.0	97	0.00
27 C	2,4-Dichlorophenol	0.295	0.315	-6.8	96	0.00
28	Naphthalene	0.973	1.001	-2.9	94	0.00
29 S	4-Chloroaniline-d4	0.342	0.384	-12.3	99	0.00
30	4-Chloroaniline	0.347	0.387	-11.5	97	0.00
31 C	Hexachlorobutadiene	0.241	0.247	-2.5	93	0.00
32	Caprolactam	0.082	0.084	-2.4	103	0.00
33 C	4-Chloro-3-methylphenol	0.310	0.334	-7.7	98	0.00
34	2-Methylnaphthalene	0.692	0.723	-4.5	95	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
36	1,2,4,5-Tetrachlorobenzene	0.646	0.651	-0.8	94	0.00
37	Hexachlorocyclopentadiene	0.380	0.282	25.8#	76	0.00
38 C	2,4,6-Trichlorophenol	0.359	0.376	-4.7	98	0.00
39	2,4,5-Trichlorophenol	0.393	0.399	-1.5	96	0.00
40	1,1'-Biphenyl	1.410	1.414	-0.3	93	0.00
41	2-Chloronaphthalene	1.079	1.107	-2.6	96	0.00
42	2-Nitroaniline	0.286	0.312	-9.1	102	0.00
43 S	Dimethylphthalate-d6	1.362	1.405	-3.2	98	0.00
44	Dimethylphthalate	1.322	1.373	-3.9	98	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM110716\
 Data File : BM007764.D
 Acq On : 07 Nov 2016 09:12
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02036

Quant Time: Nov 07 10:11:58 2016
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM102016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 24 12:55:22 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.249	0.258	-3.6	99	0.00
46 S	Acenaphthylene-d8	1.639	1.714	-4.6	97	0.00
47	Acenaphthylene	1.662	1.722	-3.6	97	0.00
48	3-Nitroaniline	0.246	0.265	-7.7	103	0.00
49 C	Acenaphthene	1.147	1.155	-0.7	96	0.00
50	2,4-Dinitrophenol	0.153	0.125	18.3	93	0.00
51 S	4-Nitrophenol-d4	0.215	0.204	5.1	99	0.00
52	4-Nitrophenol	0.214	0.204	4.7	95	0.00
53	Dibenzofuran	1.647	1.675	-1.7	96	0.00
54	2,4-Dinitrotoluene	0.359	0.386	-7.5	100	0.00
55	2,3,4,6-Tetrachlorophenol	0.351	0.369	-5.1	99	0.00
56	Diethylphthalate	1.309	1.345	-2.8	98	0.00
57 S	Fluorene-d10	1.194	1.225	-2.6	96	0.00
58	Fluorene	1.315	1.369	-4.1	98	0.00
59	4-Chlorophenyl-phenylether	0.701	0.721	-2.9	97	0.00
60	4-Nitroaniline	0.248	0.243	2.0	106	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.119	0.108	9.2	96	0.00
63	4,6-Dinitro-2-methylphenol	0.123	0.116	5.7	98	0.00
64	N-Nitrosodiphenylamine	0.572	0.597	-4.4	98	0.00
65	4-Bromophenyl-phenylether	0.228	0.236	-3.5	98	0.00
66	Hexachlorobenzene	0.253	0.258	-2.0	96	0.00
67	Atrazine	0.220	0.224	-1.8	100	0.00
68 C	Pentachlorophenol	0.142	0.135	4.9	101	0.00
69	Phenanthrene	1.063	1.105	-4.0	99	0.00
70 S	Anthracene-d10	0.901	0.961	-6.7	100	0.00
71	Anthracene	1.083	1.150	-6.2	100	0.00
72	Carbazole	0.937	0.981	-4.7	100	0.00
73	Di-n-butylphthalate	1.044	1.132	-8.4	101	0.00
74 C	Fluoranthene	1.243	1.322	-6.4	101	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	97	0.00
76 S	Pyrene-d10	0.822	0.847	-3.0	100	0.00
77	Pyrene	1.033	1.065	-3.1	100	0.00
78	Butylbenzylphthalate	0.368	0.404	-9.8	107	0.00
79	3,3'-Dichlorobenzidine	0.355	0.375	-5.6	102	0.00
80	Benzo(a)anthracene	1.069	1.128	-5.5	102	0.00
81	Bis(2-ethylhexyl)phthalate	0.541	0.602	-11.3	107	0.00
82	Chrysene	1.027	1.079	-5.1	102	0.00
83 I	Perylene-d12	1.000	1.000	0.0	101	0.00
84	Di-n-octyl phthalate	1.014	1.033	-1.9	108	0.00
85	Benzo(b)fluoranthene	1.143	1.142	0.1	101	0.00
86	Benzo(k)fluoranthene	1.051	1.103	-4.9	105	0.00
87 S	Benzo(a)pyrene-d12	0.873	0.912	-4.5	105	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM110716\
Data File : BM007764.D
Acq On : 07 Nov 2016 09:12
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02036

Quant Time: Nov 07 10:11:58 2016
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM102016.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Oct 24 12:55:22 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.076	1.117	-3.8	104	0.00
89	Indeno(1,2,3-cd)pyrene	1.289	1.339	-3.9	104	0.00
90	Dibenzo(a,h)anthracene	1.088	1.118	-2.8	103	0.00
91	Benzo(a,h,i)perylene	1.083	1.124	-3.8	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0