

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110720\
 Data File : BM027747.D
 Acq On : 07 Nov 2020 07:36
 Operator : JU/CG
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.4112

Quant Time: Nov 07 08:06:53 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM110520.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Nov 07 02:24:33 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.393	152	431	0.40	ng/ul	0.00
4) Naphthalene-d8	11.233	136	1460	0.40	ng/ul	0.00
9) Acenaphthene-d10	15.005	164	926	0.40	ng/ul	0.00
13) Phenanthrene-d10	17.717	188	2063	0.40	ng/ul	0.00
19) Chrysene-d12	21.815	240	1804	0.40	ng/ul	0.00
23) Perylene-d12	24.257	264	1662	0.40	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.539	96	180	0.41	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.794	152	1014	0.42	ng/ul	0.00
17) Fluoranthene-d10	19.702	212	2330	0.42	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.577	88	197	0.41	ng/ul	92
5) Naphthalene	11.283	128	1721	0.42	ng/ul	99
7) 2-Methylnaphthalene	12.866	142	1188	0.42	ng/ul	100
8) 1-Methylnaphthalene	13.082	142	1110	0.41	ng/ul	100
10) Acenaphthylene	14.732	152	1660	0.40	ng/ul	98
11) Acenaphthene	15.065	153	1299	0.41	ng/ul	98
12) Fluorene	16.040	166	1545	0.42	ng/ul	97
14) Pentachlorophenol	17.360	266	309	0.45	ng/ul	94
15) Phenanthrene	17.759	178	2468	0.41	ng/ul	98
16) Anthracene	17.849	178	2419	0.41	ng/ul	99
18) Fluoranthene	19.733	202	2886	0.41	ng/ul	98
20) Pyrene	20.087	202	2890	0.41	ng/ul	99
21) Benzo(a)anthracene	21.795	228	2676	0.41	ng/ul	99
22) Chrysene	21.851	228	2737	0.40	ng/ul	98
24) Benzo(b)fluoranthene	23.508	252	2706	0.40	ng/ul	99
25) Benzo(k)fluoranthene	23.559	252	2657	0.40	ng/ul	99
26) Benzo(a)pyrene	24.150	252	2381	0.39	ng/ul	97
27) Indeno(1,2,3-cd)pyrene	26.788	276	3035	0.43	ng/ul#	97
28) Dibenzo(a,h)anthracene	26.801	278	2527	0.43	ng/ul	96
29) Benzo(g,h,i)perylene	27.566	276	2584	0.43	ng/ul	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110720\
Data File : BM027747.D
Acq On : 07 Nov 2020 07:36
Operator : JU/CG
Sample : SSTDCCC0.4
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4112

Quant Time: Nov 07 08:06:53 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM110520.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Nov 07 02:24:33 2020
Response via : Initial Calibration

