

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110722\
 Data File : BM037388.D
 Acq On : 07 Nov 2022 12:47
 Operator : CG/JU
 Sample : N5276-05
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EW4W5

Quant Time: Nov 07 23:06:54 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM110522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 07 00:30:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.837	152	6181	0.400	ng/ul	0.00
4) Naphthalene-d8	10.633	136	20314	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.473	164	12305	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.212	188	24828	0.400	ng/ul	0.00
17) Chrysene-d12	21.386	240	19146	0.400	ng/ul	0.00
23) Perylene-d12	23.721	264	18466	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.255	96	27577	3.514	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.228	152	8769	0.287	ng/ul	0.00
18) Fluoranthene-d10	19.238	212	22050	0.368	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110722\
Data File : BM037388.D
Acq On : 07 Nov 2022 12:47
Operator : CG/JU
Sample : N5276-05
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
EW4W5

Quant Time: Nov 07 23:06:54 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM110522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 07 00:30:18 2022
Response via : Initial Calibration

