

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110722\
 Data File : BM037412.D
 Acq On : 08 Nov 2022 04:00
 Operator : CG/JU
 Sample : N5455-02
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BFZ12

Quant Time: Nov 08 04:41:29 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM110522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 07 00:30:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.833	152	5552	0.400	ng/u1	-0.01
4) Naphthalene-d8	10.628	136	17979	0.400	ng/u1	-0.01
9) Acenaphthene-d10	14.469	164	10738	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.208	188	21274	0.400	ng/u1	0.00
17) Chrysene-d12	21.383	240	13366	0.400	ng/u1	0.00
23) Perylene-d12	23.713	264	14219	0.400	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.251	96	26018	3.691	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.228	152	8186	0.303	ng/u1	0.00
18) Fluoranthene-d10	19.234	212	17992	0.430	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110722\
Data File : BM037412.D
Acq On : 08 Nov 2022 04:00
Operator : CG/JU
Sample : N5455-02
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BFZ12

Quant Time: Nov 08 04:41:29 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM110522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 07 00:30:18 2022
Response via : Initial Calibration

