

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110722\
 Data File : BM037416.D
 Acq On : 08 Nov 2022 06:21
 Operator : CG/JU
 Sample : PB148839BL
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK839

Quant Time: Nov 08 06:54:00 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM110522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 07 00:30:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.833	152	5375	0.400	ng/u1	-0.01
4) Naphthalene-d8	10.627	136	17273	0.400	ng/u1	-0.01
9) Acenaphthene-d10	14.465	164	10331	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.208	188	19874	0.400	ng/u1	0.00
17) Chrysene-d12	21.382	240	11860	0.400	ng/u1	# 0.00
23) Perylene-d12	23.712	264	12293	0.400	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.251	96	40496	5.934	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.228	152	8376	0.323	ng/u1	0.00
18) Fluoranthene-d10	19.234	212	15799	0.426	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110722\
Data File : BM037416.D
Acq On : 08 Nov 2022 06:21
Operator : CG/JU
Sample : PB148839BL
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK839

Quant Time: Nov 08 06:54:00 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM110522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 07 00:30:18 2022
Response via : Initial Calibration

