

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110724\
 Data File : BM048511.D
 Acq On : 07 Nov 2024 14:28
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD020196

Quant Time: Nov 07 22:20:12 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM110724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 07 22:15:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.669	152	127198	20.000	ng/u1	0.00
20) Naphthalene-d8	10.451	136	507470	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.315	164	306001	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.062	188	606984	20.000	ng/u1	0.00
79) Chrysene-d12	21.286	240	597241	20.000	ng/u1	0.00
88) Perylene-d12	24.197	264	711856	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.128	96	28714	8.887	ng/uL	0.00
4) Pyridine-d5	3.546	84	203766	21.759	ng/u1	0.00
7) Phenol-d5	6.845	99	219773	20.112	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.004	67	134518	19.872	ng/u1	0.00
11) 2-Chlorophenol-d4	7.204	132	188435	21.204	ng/u1	0.00
15) 4-Methylphenol-d8	8.381	113	170147	19.393	ng/u1	0.00
21) Nitrobenzene-d5	8.822	128	87542	21.041	ng/u1	0.00
24) 2-Nitrophenol-d4	9.545	143	92549	20.534	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.080	165	174828	21.256	ng/u1	0.00
31) 4-Chloroaniline-d4	10.598	131	226729	19.852	ng/u1	0.00
46) Dimethylphthalate-d6	13.727	166	460464	20.136	ng/u1	0.00
49) Acenaphthylene-d8	14.004	160	545626	21.161	ng/u1	0.00
54) 4-Nitrophenol-d4	14.533	143	81662	19.527	ng/u1	0.00
60) Fluorene-d10	15.310	176	406529	21.039	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.439	200	66456	16.567	ng/u1	0.00
73) Anthracene-d10	17.156	188	618051	21.533	ng/u1	0.00
81) Pyrene-d10	19.450	212	714610	19.718	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.997	264	804530	21.427	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.163	88	29585	8.662	ng/uL	99
5) Pyridine	3.563	79	205767	21.439	ng/u1	99
6) Benzaldehyde	6.810	77	123597	22.853	ng/u1	97
8) Phenol	6.875	94	229962	20.206	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.092	93	181447	19.957	ng/u1	99
12) 2-Chlorophenol	7.234	128	193511	21.209	ng/u1	100
13) 2-Methylphenol	8.116	108	171254	20.003	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.186	45	271030	19.504	ng/u1	99
16) Acetophenone	8.486	105	271350	19.637	ng/u1	100
17) N-Nitroso-di-n-propyla...	8.475	70	131858	18.605	ng/u1	99
18) 4-Methylphenol	8.445	108	188099	19.907	ng/u1	98
19) Hexachloroethane	8.739	117	81057	21.075	ng/u1	99
22) Nitrobenzene	8.863	77	198245	20.506	ng/u1	97
23) Isophorone	9.392	82	366565	19.565	ng/u1	99
25) 2-Nitrophenol	9.575	139	103382	20.926	ng/u1	98
26) 2,4-Dimethylphenol	9.639	107	190007	20.950	ng/u1	97
27) Bis(2-Chloroethoxy)met...	9.875	93	231529	19.925	ng/u1	99
29) 2,4-Dichlorophenol	10.110	162	178609	21.223	ng/u1	96
30) Naphthalene	10.504	128	597979	21.098	ng/u1	98
32) 4-Chloroaniline	10.622	127	219540	19.839	ng/u1	100
33) Hexachlorobutadiene	10.786	225	117413	20.936	ng/u1	99
34) Caprolactam	11.398	113	50518	18.821	ng/u1	97
35) 4-Chloro-3-methylphenol	11.757	107	170549	20.141	ng/u1	98
36) 2-Methylnaphthalene	12.122	142	391242	20.478	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110724\
 Data File : BM048511.D
 Acq On : 07 Nov 2024 14:28
 Operator : RC/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 SSTD020196


Quant Time: Nov 07 22:20:12 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM110724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 07 22:15:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.345	142	395948	20.430	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.498	216	208821	21.555	ng/ul	99
40) Hexachlorocyclopentadiene	12.474	237	98604	17.003	ng/ul	99
41) 2,4,6-Trichlorophenol	12.745	196	132946	21.260	ng/ul	99
42) 2,4,5-Trichlorophenol	12.821	196	143694	21.756	ng/ul	99
43) 1,1'-Biphenyl	13.145	154	504428	21.493	ng/ul	100
44) 2-Chloronaphthalene	13.186	162	397116	21.836	ng/ul	99
45) 2-Nitroaniline	13.398	65	102744	20.413	ng/ul	98
47) Dimethylphthalate	13.774	163	465464	20.295	ng/ul	99
48) 2,6-Dinitrotoluene	13.898	165	90536	20.302	ng/ul	99
50) Acenaphthylene	14.033	152	608374	21.601	ng/ul	99
51) 3-Nitroaniline	14.233	138	92112	19.623	ng/ul	100
52) Acenaphthene	14.380	153	427459	21.389	ng/ul	99
53) 2,4-Dinitrophenol	14.445	184	37933	12.507	ng/ul	96
55) 4-Nitrophenol	14.551	109	58396	19.129	ng/ul	97
56) Dibenzofuran	14.715	168	587644	21.372	ng/ul	99
57) 2,4-Dinitrotoluene	14.692	165	134342	20.606	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	14.945	232	118599	20.944	ng/ul#	99
59) Diethylphthalate	15.145	149	468486	20.304	ng/ul	99
61) Fluorene	15.362	166	472089	21.562	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.362	204	233203	21.217	ng/ul	98
63) 4-Nitroaniline	15.398	138	78900	18.563	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.457	198	73820	17.166	ng/ul	99
67) N-Nitrosodiphenylamine	15.574	169	392336	21.649	ng/ul	99
68) 4-Bromophenyl-phenylether	16.257	248	138449	20.906	ng/ul	96
69) Hexachlorobenzene	16.368	284	168458	21.260	ng/ul	98
70) Atrazine	16.533	200	131824	21.245	ng/ul	99
71) Pentachlorophenol	16.715	266	99833	19.913	ng/ul	97
72) Phenanthrene	17.104	178	721347	21.462	ng/ul	99
74) Anthracene	17.192	178	736101	21.881	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.110	216	206687	21.517	ng/uL	97
76) Pentachlorobenzene	14.639	250	205507	21.022	ng/uL	100
77) Carbazole	17.468	167	660213	22.351	ng/ul	99
78) Di-n-butylphthalate	18.033	149	793894	20.452	ng/ul	99
80) Fluoranthene	19.121	202	831264	19.712	ng/ul	99
82) Pyrene	19.480	202	906996	20.060	ng/ul	99
83) Butylbenzylphthalate	20.386	149	355624	19.033	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.197	252	271654	20.846	ng/ul	100
85) Benzo(a)anthracene	21.268	228	896639	20.818	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.192	149	531858	19.754	ng/ul	100
87) Chrysene	21.327	228	860052	21.235	ng/ul	99
89) Di-n-octyl phthalate	22.291	149	919873	17.273	ng/ul	100
90) Benzo(b)fluoranthene	23.280	252	946879	20.844	ng/ul	99
91) Benzo(k)fluoranthene	23.338	252	927726	20.397	ng/ul	100
93) Benzo(a)pyrene	24.062	252	887535	21.476	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	27.479	276	1156629	23.239	ng/ul	97
95) Dibenzo(a,h)anthracene	27.521	278	946188	23.245	ng/ul	99
96) Benzo(g,h,i)perylene	28.485	276	901333	23.599	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 07 22:20:12 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM110724.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Nov 07 22:15:00 2024
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.662 ng/uL
RT: 3.163 min Scan# 30
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28




Instrument :
BNA_M
ClientSampleId :
SSTD020196


Tgt Ion: 88 Resp: 29585
Ion Ratio Lower Upper
88 100
43 27.8 22.1 33.1
58 77.3 60.7 91.1



#5
Pyridine
Concen: 21.439 ng/uL
RT: 3.563 min Scan# 98
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28

Tgt Ion: 79 Resp: 205767
Ion Ratio Lower Upper
79 100
52 78.3 62.6 94.0
51 36.7 27.8 41.8









#6
Benzaldehyde
Concen: 22.853 ng/ul
RT: 6.810 min Scan# 610
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28




Instrument :
BNA_M
ClientSampleId :
SSTD020196


Tgt Ion: 77 Resp: 123597
Ion Ratio Lower Upper
77 100
105 101.5 78.0 117.0
106 97.5 76.6 115.0



#8
Phenol
Concen: 20.206 ng/ul
RT: 6.875 min Scan# 661
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28

Tgt Ion: 94 Resp: 229962
Ion Ratio Lower Upper
94 100
65 25.8 20.8 31.2
66 33.5 26.1 39.1









#10
Bis(2-Chloroethyl)ether
Concen: 19.957 ng/ul
RT: 7.092 min Scan# 698
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28




Instrument :
BNA_M
ClientSampleId :
SSTD020196


Tgt Ion: 93 Resp: 181447
Ion Ratio Lower Upper
93 100
63 74.7 59.6 89.4
95 30.6 25.0 37.6



#12
2-Chlorophenol
Concen: 21.209 ng/ul
RT: 7.234 min Scan# 722
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28

Tgt Ion:128 Resp: 193511
Ion Ratio Lower Upper
128 100
64 43.8 35.2 52.8
130 32.0 25.8 38.6









#13
2-Methylphenol
Concen: 20.003 ng/ul
RT: 8.116 min Scan# 81
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28




Instrument :
BNA_M
ClientSampleId :
SSTD020196


Tgt Ion:108 Resp: 171254
Ion Ratio Lower Upper
108 100
107 86.7 71.5 107.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 19.504 ng/ul
RT: 8.186 min Scan# 884
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28

Tgt Ion: 45 Resp: 271030
Ion Ratio Lower Upper
45 100
77 13.1 10.6 16.0
79 9.9 7.4 11.2









#16
Acetophenone
Concen: 19.637 ng/ul
RT: 8.486 min Scan# 911
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28




Instrument :
BNA_M
ClientSampleId :
SSTD020196


Tgt Ion:105 Resp: 271350
Ion Ratio Lower Upper
105 100
77 74.9 59.7 89.5
51 25.6 20.6 31.0



#17
N-Nitroso-di-n-propylamine
Concen: 18.605 ng/ul
RT: 8.475 min Scan# 933
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28

Tgt Ion: 70 Resp: 131858
Ion Ratio Lower Upper
70 100
42 57.7 47.0 70.4
101 11.1 8.5 12.7
130 23.6 18.4 27.6





#18

4-Methylphenol

Concen: 19.907 ng/ul

RT: 8.445 min Scan# 91

Delta R.T. 0.000 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28

Instrument :

BNA_M

ClientSampleId :





SSTD020196

Tgt Ion:108 Resp: 188099

Ion Ratio Lower Upper

108 100

107 112.9 88.8 133.2



#19

Hexachloroethane

Concen: 21.075 ng/ul

RT: 8.739 min Scan# 978

Delta R.T. 0.000 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28




Tgt Ion:117 Resp: 81057


Ion Ratio Lower Upper

117 100

201 106.2 83.8 125.8

199 64.2 52.1 78.1









#22
Nitrobenzene
Concen: 20.506 ng/ul
RT: 8.863 min Scan# 999
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28




Instrument :
BNA_M
ClientSampleId :
SSTD020196


Tgt Ion: 77 Resp: 198245
Ion Ratio Lower Upper
77 100
123 47.9 36.6 55.0
65 12.9 9.5 14.3



#23
Isophorone
Concen: 19.565 ng/ul
RT: 9.392 min Scan# 1089
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28

Tgt Ion: 82 Resp: 366565
Ion Ratio Lower Upper
82 100
95 6.7 5.0 7.6
138 18.0 14.2 21.4









#25
2-Nitrophenol
Concen: 20.926 ng/ul
RT: 9.575 min Scan# 1120
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28




Instrument :
BNA_M
ClientSampleId :
SSTD020196


Tgt Ion:139 Resp: 103382
Ion Ratio Lower Upper
139 100
65 46.5 37.7 56.5
109 25.6 22.3 33.5



#26
2,4-Dimethylphenol
Concen: 20.950 ng/ul
RT: 9.639 min Scan# 1131
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28

Tgt Ion:107 Resp: 190007
Ion Ratio Lower Upper
107 100
121 50.3 42.9 64.3
122 88.9 69.8 104.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.925 ng/ul
 RT: 9.875 min Scan# 1171
 Delta R.T. 0.000 min
 Lab File: BM048511.D
 Acq: 07 Nov 2024 14:28

Instrument :
 BNA_M
 ClientSampleId :
 SST020196

Tgt Ion: 93 Resp: 231529




Ion	Ratio	Lower	Upper
93	100		
95	32.4	25.3	37.9
123	11.9	9.5	14.3




#29
 2,4-Dichlorophenol
 Concen: 21.223 ng/ul
 RT: 10.110 min Scan# 1211
 Delta R.T. 0.000 min
 Lab File: BM048511.D
 Acq: 07 Nov 2024 14:28

Tgt Ion: 162 Resp: 178609

Ion	Ratio	Lower	Upper
162	100		
164	64.7	54.6	82.0
98	33.7	28.2	42.2









#30
Naphthalene
Concen: 21.098 ng/ul
RT: 10.504 min Scan# 1278
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28




Instrument :
BNA_M
ClientSampleId :
SSTD020196


Tgt Ion:128 Resp: 597979
Ion Ratio Lower Upper
128 100
129 11.2 8.5 12.7
127 13.1 10.0 15.0



#32
4-Chloroaniline
Concen: 19.839 ng/ul
RT: 10.622 min Scan# 1298
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28

Tgt Ion:127 Resp: 219540
Ion Ratio Lower Upper
127 100
129 32.6 26.2 39.4









#33
Hexachlorobutadiene
Concen: 20.936 ng/ul
RT: 10.786 min Scan# 117413
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28




Instrument :
BNA_M
ClientSampleId :
SSTD020196


Tgt Ion:225 Resp: 117413
Ion Ratio Lower Upper
225 100
223 61.5 48.5 72.7
227 63.7 51.5 77.3



#34
Caprolactam
Concen: 18.821 ng/ul
RT: 11.398 min Scan# 1430
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28

Tgt Ion:113 Resp: 50518
Ion Ratio Lower Upper
113 100
55 172.7 138.5 207.7
56 122.4 104.1 156.1









#35
4-Chloro-3-methylphenol
Concen: 20.141 ng/ul
RT: 11.757 min Scan# 1491
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28




Instrument :
BNA_M
ClientSampleId :
SSTD020196


Tgt Ion:107 Resp: 170549
Ion Ratio Lower Upper
107 100
144 27.6 20.6 31.0
142 85.2 66.7 100.1



#36
2-Methylnaphthalene
Concen: 20.478 ng/ul
RT: 12.122 min Scan# 1553
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28

Tgt Ion:142 Resp: 391242
Ion Ratio Lower Upper
142 100
141 87.2 71.6 107.4









#37
 1-Methylnaphthalene
 Concen: 20.430 ng/ul
 RT: 12.345 min Scan# 1591
 Delta R.T. 0.000 min
 Lab File: BM048511.D
 Acq: 07 Nov 2024 14:28




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020196


Tgt Ion:142 Resp: 395948
 Ion Ratio Lower Upper
 142 100
 141 93.3 73.6 110.4
 116 3.9 2.7 4.1



#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 21.555 ng/ul
 RT: 12.498 min Scan# 1617
 Delta R.T. 0.000 min
 Lab File: BM048511.D
 Acq: 07 Nov 2024 14:28

Tgt Ion:216 Resp: 208821
 Ion Ratio Lower Upper
 216 100
 214 79.4 62.6 93.8
 179 20.5 16.5 24.7
 108 15.7 12.9 19.3









#40
Hexachlorocyclopentadiene
Concen: 17.003 ng/ul
RT: 12.474 min Scan# 1613
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28




Instrument :
BNA_M
ClientSampleId :
SSTD020196


Tgt Ion:237 Resp: 98604
Ion Ratio Lower Upper
237 100
235 61.2 50.0 75.0
272 13.3 10.6 16.0



#41
2,4,6-Trichlorophenol
Concen: 21.260 ng/ul
RT: 12.745 min Scan# 1659
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28

Tgt Ion:196 Resp: 132946
Ion Ratio Lower Upper
196 100
198 94.5 76.6 114.8
200 31.1 25.8 38.6





#42

2,4,5-Trichlorophenol

Concen: 21.756 ng/ul

RT: 12.821 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28

Instrument :

BNA_M

ClientSampleId :

SSTD020196





Tgt Ion:196 Resp: 143694

Ion Ratio Lower Upper

196 100

198 96.8 78.7 118.1

200 31.2 24.6 37.0



#43

1,1'-Biphenyl

Concen: 21.493 ng/ul

RT: 13.145 min Scan# 1727

Delta R.T. 0.000 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28




Tgt Ion:154 Resp: 504428


Ion Ratio Lower Upper

154 100

153 39.6 31.4 47.2

76 12.5 9.9 14.9





#44

2-Chloronaphthalene

Concen: 21.836 ng/ul

RT: 13.186 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28

Instrument :

BNA_M

ClientSampleId :

SSTD020196





Tgt Ion:162 Resp: 397116

Ion Ratio Lower Upper

162 100

164 31.9 25.8 38.6

127 37.3 29.5 44.3



#45

2-Nitroaniline

Concen: 20.413 ng/ul

RT: 13.398 min Scan# 1770

Delta R.T. 0.000 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28




Tgt Ion: 65 Resp: 102744


Ion Ratio Lower Upper

65 100

92 68.3 56.2 84.2

138 110.5 89.6 134.4









#47
Dimethylphthalate
Concen: 20.295 ng/ul
RT: 13.774 min Scan# 1834
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28




Instrument :
BNA_M
ClientSampleId :
SSTD020196


Tgt Ion:163 Resp: 465464
Ion Ratio Lower Upper
163 100
194 3.9 3.0 4.4
164 10.2 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 20.302 ng/ul
RT: 13.898 min Scan# 1855
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28

Tgt Ion:165 Resp: 90536
Ion Ratio Lower Upper
165 100
89 55.8 45.4 68.0
121 21.3 17.1 25.7





#50

Acenaphthylene

Concen: 21.601 ng/ul

RT: 14.033 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28

Instrument :

BNA_M

ClientSampleId :

SSTD020196





Tgt Ion:152 Resp: 608374

Ion Ratio Lower Upper

152 100

151 20.0 15.8 23.6

153 13.4 10.5 15.7



#51

3-Nitroaniline

Concen: 19.623 ng/ul

RT: 14.233 min Scan# 1912

Delta R.T. 0.000 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28




Tgt Ion:138 Resp: 92112


Ion Ratio Lower Upper

138 100

108 10.1 7.8 11.8

92 113.8 91.0 136.4









#52
Acenaphthene
Concen: 21.389 ng/ul
RT: 14.380 min Scan# 1937
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28




Instrument :
BNA_M
ClientSampleId :
SSTD020196


Tgt Ion:153 Resp: 427459
Ion Ratio Lower Upper
153 100
152 47.3 38.3 57.5
154 88.4 71.1 106.7



#53
2,4-Dinitrophenol
Concen: 12.507 ng/ul
RT: 14.445 min Scan# 1948
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28

Tgt Ion:184 Resp: 37933
Ion Ratio Lower Upper
184 100
63 68.9 52.5 78.7
154 66.5 51.4 77.0









#55
4-Nitrophenol
Concen: 19.129 ng/ul
RT: 14.551 min Scan# 1966
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28




Instrument :
BNA_M
ClientSampleId :
SSTD020196


Tgt Ion:109 Resp: 58396
Ion Ratio Lower Upper
109 100
139 149.7 115.9 173.9
65 135.3 107.0 160.4



#56
Dibenzofuran
Concen: 21.372 ng/ul
RT: 14.715 min Scan# 1994
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28

Tgt Ion:168 Resp: 587644
Ion Ratio Lower Upper
168 100
139 36.4 29.7 44.5









#57
2,4-Dinitrotoluene
Concen: 20.606 ng/ul
RT: 14.692 min Scan# 1990
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28




Instrument :
BNA_M
ClientSampleId :
SSTD020196


Tgt Ion:165 Resp: 134342
Ion Ratio Lower Upper
165 100
63 41.7 34.6 51.8
182 2.7 2.4 3.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.944 ng/ul
RT: 14.945 min Scan# 2033
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28

Tgt Ion:232 Resp: 118599
Ion Ratio Lower Upper
232 100
131 45.8 36.3 54.5
130 2.8 1.8 2.8#
166 30.4 24.2 36.2









#59
Diethylphthalate
Concen: 20.304 ng/ul
RT: 15.145 min Scan# 2067
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28




Instrument :
BNA_M
ClientSampleId :
SSTD020196


Tgt Ion:149 Resp: 468486
Ion Ratio Lower Upper
149 100
177 22.5 17.8 26.8
150 12.5 9.6 14.4



#61
Fluorene
Concen: 21.562 ng/ul
RT: 15.362 min Scan# 2104
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28

Tgt Ion:166 Resp: 472089
Ion Ratio Lower Upper
166 100
165 97.4 78.7 118.1
167 13.3 11.0 16.6









#62
4-Chlorophenyl-phenylether
Concen: 21.217 ng/ul
RT: 15.362 min Scan# 2104
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28




Instrument :
BNA_M
ClientSampleId :
SSTD020196


Tgt Ion:204 Resp: 233203
Ion Ratio Lower Upper
204 100
206 32.9 26.0 39.0
141 54.8 45.4 68.2



#63
4-Nitroaniline
Concen: 18.563 ng/ul
RT: 15.398 min Scan# 2110
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28

Tgt Ion:138 Resp: 78900
Ion Ratio Lower Upper
138 100
92 47.8 39.7 59.5
108 69.6 58.5 87.7









#66
 4,6-Dinitro-2-methylphenol
 Concen: 17.166 ng/ul
 RT: 15.457 min Scan# 2120
 Delta R.T. 0.000 min
 Lab File: BM048511.D
 Acq: 07 Nov 2024 14:28




Instrument :
 BNA_M
 ClientSampleId :
 SSTD020196


Tgt Ion:198 Resp: 73820
 Ion Ratio Lower Upper
 198 100
 182 5.8 5.0 7.6
 77 24.8 20.1 30.1



#67
 N-Nitrosodiphenylamine
 Concen: 21.649 ng/ul
 RT: 15.574 min Scan# 2140
 Delta R.T. 0.000 min
 Lab File: BM048511.D
 Acq: 07 Nov 2024 14:28

Tgt Ion:169 Resp: 392336
 Ion Ratio Lower Upper
 169 100
 168 65.5 52.6 79.0
 167 36.5 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 20.906 ng/ul
RT: 16.257 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28




Instrument :
BNA_M
ClientSampleId :
SSTD020196


Tgt Ion:248 Resp: 138449
Ion Ratio Lower Upper
248 100
250 95.9 79.4 119.2
141 70.5 53.8 80.6



#69
Hexachlorobenzene
Concen: 21.260 ng/ul
RT: 16.368 min Scan# 2275
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28

Tgt Ion:284 Resp: 168458
Ion Ratio Lower Upper
284 100
142 34.9 27.5 41.3
249 30.5 23.2 34.8





#70

Atrazine

Concen: 21.245 ng/ul

RT: 16.533 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28

Instrument :

BNA_M

ClientSampleId :

SSTD020196





Tgt Ion:200 Resp: 131824

Ion Ratio Lower Upper

200 100

173 26.0 20.2 30.2

215 46.8 38.2 57.2



#71

Pentachlorophenol

Concen: 19.913 ng/ul

RT: 16.715 min Scan# 2334

Delta R.T. 0.000 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28




Tgt Ion:266 Resp: 99833


Ion Ratio Lower Upper

266 100

264 60.6 51.1 76.7

268 64.8 53.2 79.8





#72

Phenanthrene

Concen: 21.462 ng/ul

RT: 17.104 min Scan# 2400

Delta R.T. 0.000 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28

Instrument :

BNA_M

ClientSampleId :

SSTD020196





Tgt Ion:178 Resp: 721347

Ion Ratio Lower Upper

178 100

179 14.8 12.0 18.0

176 19.2 15.1 22.7



#74

Anthracene

Concen: 21.881 ng/ul

RT: 17.192 min Scan# 2415

Delta R.T. 0.000 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28




Tgt Ion:178 Resp: 736101


Ion Ratio Lower Upper

178 100

179 14.8 12.6 19.0

176 18.9 15.0 22.4





#75

1,2,3,4-Tetrachlorobenzene

Concen: 21.517 ng/uL

RT: 13.110 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28

Instrument :

BNA_M

ClientSampleId :

SSTD020196





Tgt Ion:216 Resp: 206687

Ion Ratio Lower Upper

216 100

214 80.5 61.9 92.9

218 49.0 38.6 58.0



#76

Pentachlorobenzene

Concen: 21.022 ng/uL

RT: 14.639 min Scan# 1981

Delta R.T. 0.006 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28




Tgt Ion:250 Resp: 205507


Ion Ratio Lower Upper

250 100

248 62.7 49.7 74.5

252 62.9 50.2 75.2





#77

Carbazole

Concen: 22.351 ng/ul

RT: 17.468 min Scan# 2462

Delta R.T. 0.000 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28

Instrument :

BNA_M

ClientSampleId :

SSTD020196





Tgt Ion:167 Resp: 660213

Ion Ratio Lower Upper

167 100

166 21.2 16.4 24.6

139 12.7 10.3 15.5



#78

Di-n-butylphthalate

Concen: 20.452 ng/ul

RT: 18.033 min Scan# 2558

Delta R.T. 0.006 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28




Tgt Ion:149 Resp: 793894


Ion Ratio Lower Upper

149 100

150 8.9 7.4 11.2

104 4.6 4.0 6.0









#80
Fluoranthene
Concen: 19.712 ng/ul
RT: 19.121 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28




Instrument :
BNA_M
ClientSampleId :
SSTD020196


Tgt Ion:202 Resp: 831264
Ion Ratio Lower Upper
202 100
101 11.3 8.5 12.7
100 8.5 7.0 10.4



#82
Pyrene
Concen: 20.060 ng/ul
RT: 19.480 min Scan# 2804
Delta R.T. 0.000 min
Lab File: BM048511.D
Acq: 07 Nov 2024 14:28

Tgt Ion:202 Resp: 906996
Ion Ratio Lower Upper
202 100
101 13.1 10.6 15.8
100 10.6 8.9 13.3





#83

Butylbenzylphthalate

Concen: 19.033 ng/ul

RT: 20.386 min Scan# 2958

Delta R.T. 0.000 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28

Instrument :

BNA_M

ClientSampleId :

SSTD020196





Tgt Ion:149 Resp: 355624

Ion Ratio Lower Upper

149 100

91 66.8 53.1 79.7

206 20.3 16.6 25.0



#84

3,3'-Dichlorobenzidine

Concen: 20.846 ng/ul

RT: 21.197 min Scan# 3096

Delta R.T. 0.000 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28




Tgt Ion:252 Resp: 271654


Ion Ratio Lower Upper

252 100

254 64.9 52.1 78.1

126 10.4 8.3 12.5





#85

Benzo(a)anthracene

Concen: 20.818 ng/ul

RT: 21.268 min Scan# 3107

Delta R.T. 0.006 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28

Instrument :

BNA_M

ClientSampleId :

SSTD020196





Tgt Ion:228 Resp: 896639

Ion Ratio Lower Upper

228 100

229 19.5 15.7 23.5

226 26.2 21.7 32.5



#86

Bis(2-ethylhexyl)phthalate

Concen: 19.754 ng/ul

RT: 21.192 min Scan# 3095

Delta R.T. 0.000 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28




Tgt Ion:149 Resp: 531858


Ion Ratio Lower Upper

149 100

167 27.3 22.0 33.0

279 3.7 3.1 4.7





#87

Chrysene

Concen: 21.235 ng/ul

RT: 21.327 min Scan# 3118

Delta R.T. 0.000 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28

Instrument :

BNA_M

ClientSampleId :

SSTD020196





Tgt Ion:228 Resp: 860052

Ion Ratio Lower Upper

228 100

226 29.6 23.4 35.2

229 19.8 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 17.273 ng/ul




RT: 22.291 min Scan# 3282


Delta R.T. 0.000 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28

Tgt Ion:149 Resp: 919873





#90

Benzo(b)fluoranthene

Concen: 20.844 ng/ul

RT: 23.280 min Scan# 3449

Delta R.T. 0.006 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28

Instrument :

BNA_M

ClientSampleId :

SSTD020196





Tgt Ion:252 Resp: 946879

Ion Ratio Lower Upper

252 100

253 21.8 17.7 26.5

125 10.5 8.7 13.1



#91

Benzo(k)fluoranthene

Concen: 20.397 ng/ul

RT: 23.338 min Scan# 3460

Delta R.T. 0.000 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28




Tgt Ion:252 Resp: 927726


Ion Ratio Lower Upper

252 100

253 21.9 17.4 26.0

125 10.0 8.0 12.0





#93

Benzo(a)pyrene

Concen: 21.476 ng/ul

RT: 24.062 min Scan# 3

Delta R.T. 0.006 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28

Instrument :

BNA_M

ClientSampleId :

SSTD020196





Tgt Ion:252 Resp: 887535

Ion Ratio Lower Upper

252 100

253 21.8 17.8 26.8

125 11.4 9.1 13.7



#94

Indeno(1,2,3-cd)pyrene

Concen: 23.239 ng/ul

RT: 27.479 min Scan# 4164

Delta R.T. 0.012 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28




Tgt Ion:276 Resp: 1156629


Ion Ratio Lower Upper

276 100

138 22.3 16.2 24.2

277 24.9 19.5 29.3





#95

Dibenzo(a,h)anthracene

Concen: 23.245 ng/ul

RT: 27.521 min Scan# 4171

Delta R.T. 0.006 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28

Instrument :

BNA_M

ClientSampleId :

SSTD020196





Tgt Ion:278 Resp: 946188

Ion Ratio Lower Upper

278 100

139 15.6 12.6 18.8

279 23.4 19.0 28.6



#96

Benzo(g,h,i)perylene

Concen: 23.599 ng/ul

RT: 28.485 min Scan# 4335

Delta R.T. 0.000 min

Lab File: BM048511.D

Acq: 07 Nov 2024 14:28

Tgt Ion:276 Resp: 901333

Ion Ratio Lower Upper

276 100

138 22.1 17.6 26.4

277 23.8 18.6 28.0

