

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110724\
 Data File : BM048512.D
 Acq On : 07 Nov 2024 15:07
 Operator : RC/JU
 Sample : SSTD00554
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD005041

Quant Time: Nov 07 22:20:50 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM110724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 07 22:15:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.669	152	128238	20.000	ng/u1	0.00
20) Naphthalene-d8	10.451	136	510138	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.316	164	305786	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.057	188	603038	20.000	ng/u1	0.00
79) Chrysene-d12	21.286	240	585027	20.000	ng/u1	0.00
88) Perylene-d12	24.198	264	709490	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.128	96	6215	1.908	ng/uL	0.00
4) Pyridine-d5	3.558	84	36025	3.816	ng/u1	0.01
7) Phenol-d5	6.852	99	40884	3.711	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.005	67	25924	3.799	ng/u1	0.00
11) 2-Chlorophenol-d4	7.205	132	35950	4.013	ng/u1	0.00
15) 4-Methylphenol-d8	8.387	113	31160	3.523	ng/u1	0.00
21) Nitrobenzene-d5	8.828	128	16193	3.872	ng/u1	0.00
24) 2-Nitrophenol-d4	9.546	143	15939	3.518	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.093	165	31789	3.845	ng/u1	0.01
31) 4-Chloroaniline-d4	10.610	131	38815	3.381	ng/u1	0.01
46) Dimethylphthalate-d6	13.728	166	90594	3.964	ng/u1	0.00
49) Acenaphthylene-d8	14.004	160	103361	4.012	ng/u1	0.00
54) 4-Nitrophenol-d4	14.557	143	5899	1.412	ng/u1	0.02
60) Fluorene-d10	15.310	176	79484	4.116	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.451	200	7046	1.768	ng/u1	0.01
73) Anthracene-d10	17.157	188	118150	4.143	ng/u1	0.00
81) Pyrene-d10	19.457	212	135184	3.808	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.998	264	148286	3.963	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.164	88	7064	2.051	ng/uL	93
5) Pyridine	3.575	79	42684	4.411	ng/u1#	84
6) Benzaldehyde	6.816	77	25652	4.705	ng/u1	99
8) Phenol	6.875	94	43770	3.815	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.099	93	36157	3.945	ng/u1	98
12) 2-Chlorophenol	7.234	128	37762	4.105	ng/u1	99
13) 2-Methylphenol	8.122	108	32708	3.789	ng/u1	94
14) 2,2'-oxybis(1-Chloropr...	8.187	45	57410	4.098	ng/u1	96
16) Acetophenone	8.493	105	53218	3.820	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.475	70	25065	3.508	ng/u1	97
18) 4-Methylphenol	8.457	108	34631	3.635	ng/u1	94
19) Hexachloroethane	8.740	117	16796	4.332	ng/u1	96
22) Nitrobenzene	8.869	77	36602	3.766	ng/u1	100
23) Isophorone	9.393	82	69961	3.715	ng/u1	99
25) 2-Nitrophenol	9.581	139	17393	3.502	ng/u1	93
26) 2,4-Dimethylphenol	9.646	107	36373	3.989	ng/u1	96
27) Bis(2-Chloroethoxy)met...	9.875	93	42337	3.624	ng/u1	98
29) 2,4-Dichlorophenol	10.122	162	32665	3.861	ng/u1	98
30) Naphthalene	10.504	128	126312	4.433	ng/u1	99
32) 4-Chloroaniline	10.634	127	37969	3.413	ng/u1	93
33) Hexachlorobutadiene	10.787	225	24086	4.272	ng/u1	94
34) Caprolactam	11.410	113	7906	2.930	ng/u1	87
35) 4-Chloro-3-methylphenol	11.763	107	30583	3.593	ng/u1	94
36) 2-Methylnaphthalene	12.122	142	77349	4.027	ng/u1	94


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110724\
 Data File : BM048512.D
 Acq On : 07 Nov 2024 15:07
 Operator : RC/JU
 Sample : SSTD00554
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD005041

Quant Time: Nov 07 22:20:50 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM110724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 07 22:15:00 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.345	142	80585	4.136	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.498	216	40807	4.215	ng/ul	95
40) Hexachlorocyclopentadiene	12.475	237	13540	2.336	ng/ul	92
41) 2,4,6-Trichlorophenol	12.751	196	24217	3.875	ng/ul	99
42) 2,4,5-Trichlorophenol	12.834	196	25005	3.789	ng/ul	94
43) 1,1'-Biphenyl	13.145	154	97984	4.178	ng/ul	97
44) 2-Chloronaphthalene	13.187	162	77816	4.282	ng/ul	98
45) 2-Nitroaniline	13.410	65	15412	3.064	ng/ul	92
47) Dimethylphthalate	13.775	163	91550	3.995	ng/ul	99
48) 2,6-Dinitrotoluene	13.898	165	14516	3.257	ng/ul	90
50) Acenaphthylene	14.034	152	116920	4.154	ng/ul	99
51) 3-Nitroaniline	14.245	138	12668	2.701	ng/ul#	91
52) Acenaphthene	14.381	153	85444	4.278	ng/ul	98
53) 2,4-Dinitrophenol	14.463	184	2524	0.833	ng/ul	93
55) 4-Nitrophenol	14.575	109	5323	1.745	ng/ul	89
56) Dibenzofuran	14.716	168	116357	4.235	ng/ul	96
57) 2,4-Dinitrotoluene	14.698	165	20261	3.110	ng/ul#	95
58) 2,3,4,6-Tetrachlorophenol	14.951	232	21115	3.731	ng/ul#	96
59) Diethylphthalate	15.145	149	87424	3.792	ng/ul	100
61) Fluorene	15.363	166	93629	4.279	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.363	204	46207	4.207	ng/ul	97
63) 4-Nitroaniline	15.410	138	8459	1.992	ng/ul#	85
66) 4,6-Dinitro-2-methylph...	15.463	198	8398	1.966	ng/ul#	87
67) N-Nitrosodiphenylamine	15.581	169	71708	3.983	ng/ul	98
68) 4-Bromophenyl-phenylether	16.251	248	26571	4.039	ng/ul	92
69) Hexachlorobenzene	16.369	284	32702	4.154	ng/ul	97
70) Atrazine	16.533	200	23988	3.891	ng/ul	98
71) Pentachlorophenol	16.722	266	16745	3.362	ng/ul	91
72) Phenanthrene	17.104	178	143116	4.286	ng/ul	100
74) Anthracene	17.192	178	139592	4.177	ng/ul	98
75) 1,2,3,4-Tetrachloroben...	13.110	216	42279	4.430	ng/ul	98
76) Pentachlorobenzene	14.639	250	40962	4.218	ng/ul	97
77) Carbazole	17.475	167	110016	3.749	ng/ul	100
78) Di-n-butylphthalate	18.033	149	135610	3.516	ng/ul	100
80) Fluoranthene	19.122	202	158235	3.831	ng/ul	98
82) Pyrene	19.486	202	175896	3.971	ng/ul	100
83) Butylbenzylphthalate	20.386	149	57855	3.161	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.204	252	48593	3.807	ng/ul	96
85) Benzo(a)anthracene	21.268	228	172600	4.091	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.192	149	86282	3.271	ng/ul	98
87) Chrysene	21.327	228	170947	4.309	ng/ul	98
89) Di-n-octyl phthalate	22.292	149	141816	2.672	ng/ul	100
90) Benzo(b)fluoranthene	23.280	252	180517	3.987	ng/ul	98
91) Benzo(k)fluoranthene	23.339	252	176559	3.895	ng/ul	98
93) Benzo(a)pyrene	24.062	252	165743	4.024	ng/ul	97
94) Indeno(1,2,3-cd)pyrene	27.468	276	214610	4.326	ng/ul	97
95) Dibenzo(a,h)anthracene	27.521	278	170612	4.205	ng/ul	99
96) Benzo(g,h,i)perylene	28.486	276	78251	2.056	ng/ul	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 2.051 ng/uL
RT: 3.164 min Scan# 30
Delta R.T. 0.001 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07




Instrument :
BNA_M
ClientSampleId :
SSTD005041


Tgt Ion: 88 Resp: 7064
Ion Ratio Lower Upper
88 100
43 29.9 22.1 33.1
58 68.6 60.7 91.1



#5
Pyridine
Concen: 4.411 ng/uL
RT: 3.575 min Scan# 100
Delta R.T. 0.012 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07

Tgt Ion: 79 Resp: 42684
Ion Ratio Lower Upper
79 100
52 89.8 62.6 94.0
51 47.8 27.8 41.8









#6
Benzaldehyde
Concen: 4.705 ng/ul
RT: 6.816 min Scan# 651
Delta R.T. 0.006 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07




Instrument :
BNA_M
ClientSampleId :
SSTD005041


Tgt Ion: 77 Resp: 25652
Ion Ratio Lower Upper
77 100
105 95.6 78.0 117.0
106 96.8 76.6 115.0



#8
Phenol
Concen: 3.815 ng/ul
RT: 6.875 min Scan# 661
Delta R.T. 0.001 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07

Tgt Ion: 94 Resp: 43770
Ion Ratio Lower Upper
94 100
65 25.5 20.8 31.2
66 31.3 26.1 39.1









#10
Bis(2-Chloroethyl)ether
Concen: 3.945 ng/ul
RT: 7.099 min Scan# 698
Delta R.T. 0.006 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07




Instrument :
BNA_M
ClientSampleId :
SSTD005041


Tgt Ion: 93	Resp:	36157
Ion Ratio	Lower	Upper
93	100	
63	73.1	59.6 89.4
95	29.5	25.0 37.6



#12
2-Chlorophenol
Concen: 4.105 ng/ul
RT: 7.234 min Scan# 722
Delta R.T. 0.000 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07

Tgt Ion: 128	Resp:	37762
Ion Ratio	Lower	Upper
128	100	
64	44.6	35.2 52.8
130	33.1	25.8 38.6









#13
2-Methylphenol
Concen: 3.789 ng/ul
RT: 8.122 min Scan# 81
Delta R.T. 0.006 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07




Instrument :
BNA_M
ClientSampleId :
SSTD005041


Tgt Ion:108 Resp: 32708
Ion Ratio Lower Upper
108 100
107 83.7 71.5 107.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 4.098 ng/ul
RT: 8.187 min Scan# 884
Delta R.T. 0.001 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07

Tgt Ion: 45 Resp: 57410
Ion Ratio Lower Upper
45 100
77 15.1 10.6 16.0
79 10.3 7.4 11.2









#16
Acetophenone
Concen: 3.820 ng/ul
RT: 8.493 min Scan# 91
Delta R.T. 0.006 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07




Instrument :
BNA_M
ClientSampleId :
SSTD005041


Tgt Ion:105 Resp: 53218
Ion Ratio Lower Upper
105 100
77 73.3 59.7 89.5
51 28.0 20.6 31.0



#17
N-Nitroso-di-n-propylamine
Concen: 3.508 ng/ul
RT: 8.475 min Scan# 933
Delta R.T. 0.000 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07

Tgt Ion: 70 Resp: 25065
Ion Ratio Lower Upper
70 100
42 61.7 47.0 70.4
101 10.5 8.5 12.7
130 23.0 18.4 27.6









#18
4-Methylphenol
Concen: 3.635 ng/ul
RT: 8.457 min Scan# 910
Delta R.T. 0.012 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07




Instrument :
BNA_M
ClientSampleId :
SSTD005041


Tgt Ion:108 Resp: 34631
Ion Ratio Lower Upper
108 100
107 104.9 88.8 133.2



#19
Hexachloroethane
Concen: 4.332 ng/ul
RT: 8.740 min Scan# 978
Delta R.T. 0.000 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07

Tgt Ion:117 Resp: 16796
Ion Ratio Lower Upper
117 100
201 99.5 83.8 125.8
199 64.2 52.1 78.1









#22
Nitrobenzene
Concen: 3.766 ng/ul
RT: 8.869 min Scan# 1000
Delta R.T. 0.006 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07




Instrument :
BNA_M
ClientSampleId :
SSTD005041


Tgt Ion: 77 Resp: 36602
Ion Ratio Lower Upper
77 100
123 45.6 36.6 55.0
65 11.5 9.5 14.3



#23
Isophorone
Concen: 3.715 ng/ul
RT: 9.393 min Scan# 1089
Delta R.T. 0.001 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07

Tgt Ion: 82 Resp: 69961
Ion Ratio Lower Upper
82 100
95 6.8 5.0 7.6
138 17.9 14.2 21.4









#25
2-Nitrophenol
Concen: 3.502 ng/ul
RT: 9.581 min Scan# 1121
Delta R.T. 0.006 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07




Instrument :
BNA_M
ClientSampleId :
SSTD005041


Tgt Ion:139 Resp: 17393
Ion Ratio Lower Upper
139 100
65 52.6 37.7 56.5
109 29.9 22.3 33.5



#26
2,4-Dimethylphenol
Concen: 3.989 ng/ul
RT: 9.646 min Scan# 1132
Delta R.T. 0.006 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07

Tgt Ion:107 Resp: 36373
Ion Ratio Lower Upper
107 100
121 53.0 42.9 64.3
122 92.1 69.8 104.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 3.624 ng/ul
 RT: 9.875 min Scan# 1171
 Delta R.T. 0.001 min
 Lab File: BM048512.D
 Acq: 07 Nov 2024 15:07




Instrument :
 BNA_M
 ClientSampleId :
 SSTD005041


Tgt Ion:	93	Resp:	42337
Ion	Ratio	Lower	Upper
93	100		
95	30.3	25.3	37.9
123	11.5	9.5	14.3



#29
 2,4-Dichlorophenol
 Concen: 3.861 ng/ul
 RT: 10.122 min Scan# 1213
 Delta R.T. 0.012 min
 Lab File: BM048512.D
 Acq: 07 Nov 2024 15:07

Tgt Ion:	162	Resp:	32665
Ion	Ratio	Lower	Upper
162	100		
164	67.0	54.6	82.0
98	37.0	28.2	42.2









#30
Naphthalene
Concen: 4.433 ng/ul
RT: 10.504 min Scan# 11
Delta R.T. 0.001 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07




Instrument :
BNA_M
ClientSampleId :
SSTD005041


Tgt Ion:128 Resp: 126312
Ion Ratio Lower Upper
128 100
129 11.0 8.5 12.7
127 13.0 10.0 15.0



#32
4-Chloroaniline
Concen: 3.413 ng/ul
RT: 10.634 min Scan# 1300
Delta R.T. 0.012 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07

Tgt Ion:127 Resp: 37969
Ion Ratio Lower Upper
127 100
129 28.9 26.2 39.4









#33
Hexachlorobutadiene
Concen: 4.272 ng/ul
RT: 10.787 min Scan# 1326
Delta R.T. 0.001 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07




Instrument :
BNA_M
ClientSampleId :
SSTD005041


Tgt Ion:225 Resp: 24086
Ion Ratio Lower Upper
225 100
223 66.8 48.5 72.7
227 67.2 51.5 77.3



#34
Caprolactam
Concen: 2.930 ng/ul
RT: 11.410 min Scan# 1432
Delta R.T. 0.012 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07

Tgt Ion:113 Resp: 7906
Ion Ratio Lower Upper
113 100
55 159.8 138.5 207.7
56 109.1 104.1 156.1









#35
4-Chloro-3-methylphenol
Concen: 3.593 ng/ul
RT: 11.763 min Scan# 1492
Delta R.T. 0.006 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07




Instrument :
BNA_M
ClientSampleId :
SSTD005041


Tgt Ion:107 Resp: 30583
Ion Ratio Lower Upper
107 100
144 24.8 20.6 31.0
142 77.4 66.7 100.1



#36
2-Methylnaphthalene
Concen: 4.027 ng/ul
RT: 12.122 min Scan# 1553
Delta R.T. 0.000 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07

Tgt Ion:142 Resp: 77349
Ion Ratio Lower Upper
142 100
141 84.0 71.6 107.4





#37

1-Methylnaphthalene

Concen: 4.136 ng/ul

RT: 12.345 min Scan# 11

Delta R.T. 0.001 min

Lab File: BM048512.D

Acq: 07 Nov 2024 15:07

Instrument :

BNA_M

ClientSampleId :

SSTD005041





Tgt Ion:142 Resp: 80585

Ion Ratio Lower Upper

142 100

141 90.3 73.6 110.4

116 3.7 2.7 4.1



#39

1,2,4,5-Tetrachlorobenzene

Concen: 4.215 ng/ul

RT: 12.498 min Scan# 1617

Delta R.T. 0.001 min

Lab File: BM048512.D

Acq: 07 Nov 2024 15:07

Tgt Ion:216 Resp: 40807




Ion Ratio Lower Upper


216 100

214 82.6 62.6 93.8

179 21.3 16.5 24.7

108 18.7 12.9 19.3









#40
Hexachlorocyclopentadiene
Concen: 2.336 ng/ul
RT: 12.475 min Scan# 1613
Delta R.T. 0.000 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07




Instrument :
BNA_M
ClientSampleId :
SSTD005041


Tgt Ion:237 Resp: 13540
Ion Ratio Lower Upper
237 100
235 55.5 50.0 75.0
272 14.1 10.6 16.0



#41
2,4,6-Trichlorophenol
Concen: 3.875 ng/ul
RT: 12.751 min Scan# 1660
Delta R.T. 0.006 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07

Tgt Ion:196 Resp: 24217
Ion Ratio Lower Upper
196 100
198 96.0 76.6 114.8
200 29.7 25.8 38.6









#42
2,4,5-Trichlorophenol
Concen: 3.789 ng/ul
RT: 12.834 min Scan# 1
Delta R.T. 0.012 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07




Instrument :
BNA_M
ClientSampleId :
SSTD005041


Tgt Ion:196 Resp: 25005
Ion Ratio Lower Upper
196 100
198 92.4 78.7 118.1
200 33.9 24.6 37.0



#43
1,1'-Biphenyl
Concen: 4.178 ng/ul
RT: 13.145 min Scan# 1727
Delta R.T. 0.001 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07

Tgt Ion:154 Resp: 97984
Ion Ratio Lower Upper
154 100
153 41.0 31.4 47.2
76 14.4 9.9 14.9





#44

2-Chloronaphthalene

Concen: 4.282 ng/ul

RT: 13.187 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM048512.D

Acq: 07 Nov 2024 15:07

Instrument :

BNA_M

ClientSampleId :

SSTD005041





Tgt Ion:162 Resp: 77816

Ion Ratio Lower Upper

162 100

164 32.5 25.8 38.6

127 38.4 29.5 44.3



#45

2-Nitroaniline

Concen: 3.064 ng/ul

RT: 13.410 min Scan# 1772

Delta R.T. 0.012 min

Lab File: BM048512.D

Acq: 07 Nov 2024 15:07




Tgt Ion: 65 Resp: 15412


Ion Ratio Lower Upper

65 100

92 67.1 56.2 84.2

138 100.0 89.6 134.4









#47
Dimethylphthalate
Concen: 3.995 ng/ul
RT: 13.775 min Scan# 1834
Delta R.T. 0.000 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07




Instrument :
BNA_M
ClientSampleId :
SSTD005041


Tgt Ion:163	Resp:	91550
Ion Ratio	Lower	Upper
163	100	
194	3.5	3.0 4.4
164	9.2	7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 3.257 ng/ul
RT: 13.898 min Scan# 1855
Delta R.T. 0.001 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07

Tgt Ion:165	Resp:	14516
Ion Ratio	Lower	Upper
165	100	
89	64.6	45.4 68.0
121	24.3	17.1 25.7









#50
Acenaphthylene
Concen: 4.154 ng/ul
RT: 14.034 min Scan# 11
Delta R.T. 0.001 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07




Instrument :
BNA_M
ClientSampleId :
SSTD005041


Tgt Ion:152 Resp: 116920
Ion Ratio Lower Upper
152 100
151 19.3 15.8 23.6
153 12.6 10.5 15.7



#51
3-Nitroaniline
Concen: 2.701 ng/ul
RT: 14.245 min Scan# 1914
Delta R.T. 0.012 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07

Tgt Ion:138 Resp: 12668
Ion Ratio Lower Upper
138 100
108 12.5 7.8 11.8#
92 123.8 91.0 136.4





#52

Acenaphthene

Concen: 4.278 ng/ul

RT: 14.381 min Scan# 1937

Delta R.T. 0.000 min

Lab File: BM048512.D

Acq: 07 Nov 2024 15:07

Instrument :

BNA_M

ClientSampleId :

SSTD005041





Tgt Ion:153 Resp: 85444

Ion Ratio Lower Upper

153 100

152 50.5 38.3 57.5

154 88.1 71.1 106.7



#53

2,4-Dinitrophenol

Concen: 0.833 ng/ul

RT: 14.463 min Scan# 1951

Delta R.T. 0.018 min

Lab File: BM048512.D

Acq: 07 Nov 2024 15:07




Tgt Ion:184 Resp: 2524


Ion Ratio Lower Upper

184 100

63 73.6 52.5 78.7

154 67.2 51.4 77.0









#55
4-Nitrophenol
Concen: 1.745 ng/ul
RT: 14.575 min Scan# 1970
Delta R.T. 0.024 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07




Instrument :
BNA_M
ClientSampleId :
SSTD005041


Tgt Ion:109 Resp: 5323
Ion Ratio Lower Upper
109 100
139 153.1 115.9 173.9
65 152.7 107.0 160.4



#56
Dibenzofuran
Concen: 4.235 ng/ul
RT: 14.716 min Scan# 1994
Delta R.T. 0.001 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07

Tgt Ion:168 Resp: 116357
Ion Ratio Lower Upper
168 100
139 34.9 29.7 44.5









#57
2,4-Dinitrotoluene
Concen: 3.110 ng/ul
RT: 14.698 min Scan# 1991
Delta R.T. 0.006 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07




Instrument :
BNA_M
ClientSampleId :
SSTD005041


Tgt Ion:165 Resp: 20261
Ion Ratio Lower Upper
165 100
63 46.8 34.6 51.8
182 2.4 2.4 3.6#



#58
2,3,4,6-Tetrachlorophenol
Concen: 3.731 ng/ul
RT: 14.951 min Scan# 2034
Delta R.T. 0.006 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07

Tgt Ion:232 Resp: 21115
Ion Ratio Lower Upper
232 100
131 45.7 36.3 54.5
130 2.8 1.8 2.8#
166 35.6 24.2 36.2





#59

Diethylphthalate

Concen: 3.792 ng/ul

RT: 15.145 min Scan# 2067

Delta R.T. 0.000 min

Lab File: BM048512.D

Acq: 07 Nov 2024 15:07

Instrument :

BNA_M

ClientSampleId :

SSTD005041





Tgt Ion:149 Resp: 87424

Ion Ratio Lower Upper

149 100

177 22.5 17.8 26.8

150 11.8 9.6 14.4



#61

Fluorene

Concen: 4.279 ng/ul

RT: 15.363 min Scan# 2104

Delta R.T. 0.000 min

Lab File: BM048512.D

Acq: 07 Nov 2024 15:07




Tgt Ion:166 Resp: 93629


Ion Ratio Lower Upper

166 100

165 95.0 78.7 118.1

167 13.9 11.0 16.6









#62
4-Chlorophenyl-phenylether
Concen: 4.207 ng/ul
RT: 15.363 min Scan# 2104
Delta R.T. 0.000 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07




Instrument :
BNA_M
ClientSampleId :
SSTD005041


Tgt Ion:204 Resp: 46207
Ion Ratio Lower Upper
204 100
206 32.5 26.0 39.0
141 60.7 45.4 68.2



#63
4-Nitroaniline
Concen: 1.992 ng/ul
RT: 15.410 min Scan# 2112
Delta R.T. 0.012 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07

Tgt Ion:138 Resp: 8459
Ion Ratio Lower Upper
138 100
92 65.9 39.7 59.5#
108 80.9 58.5 87.7









#66
4,6-Dinitro-2-methylphenol
Concen: 1.966 ng/ul
RT: 15.463 min Scan# 2120
Delta R.T. 0.006 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07




Instrument :
BNA_M
ClientSampleId :
SSTD005041


Tgt Ion:198 Resp: 8398
Ion Ratio Lower Upper
198 100
182 9.6 5.0 7.6#
77 31.8 20.1 30.1#



#67
N-Nitrosodiphenylamine
Concen: 3.983 ng/ul
RT: 15.581 min Scan# 2141
Delta R.T. 0.006 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07

Tgt Ion:169 Resp: 71708
Ion Ratio Lower Upper
169 100
168 67.9 52.6 79.0
167 37.8 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 4.039 ng/ul
RT: 16.251 min Scan# 21
Delta R.T. -0.005 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07




Instrument :
BNA_M
ClientSampleId :
SSTD005041


Tgt Ion:248 Resp: 26571
Ion Ratio Lower Upper
248 100
250 91.4 79.4 119.2
141 73.8 53.8 80.6



#69
Hexachlorobenzene
Concen: 4.154 ng/ul
RT: 16.369 min Scan# 2275
Delta R.T. 0.000 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07

Tgt Ion:284 Resp: 32702
Ion Ratio Lower Upper
284 100
142 31.9 27.5 41.3
249 28.2 23.2 34.8





#70

Atrazine

Concen: 3.891 ng/ul

RT: 16.533 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM048512.D

Acq: 07 Nov 2024 15:07

Instrument :

BNA_M

ClientSampleId :

SSTD005041





Tgt Ion:200 Resp: 23988

Ion Ratio Lower Upper

200 100

173 26.7 20.2 30.2

215 48.6 38.2 57.2



#71

Pentachlorophenol

Concen: 3.362 ng/ul

RT: 16.722 min Scan# 2335

Delta R.T. 0.006 min

Lab File: BM048512.D

Acq: 07 Nov 2024 15:07




Tgt Ion:266 Resp: 16745


Ion Ratio Lower Upper

266 100

264 58.1 51.1 76.7

268 58.3 53.2 79.8





#72

Phenanthrene

Concen: 4.286 ng/ul

RT: 17.104 min Scan# 2400

Delta R.T. 0.000 min

Lab File: BM048512.D

Acq: 07 Nov 2024 15:07

Instrument :

BNA_M

ClientSampleId :

SSTD005041





Tgt Ion:178 Resp: 143116

Ion Ratio Lower Upper

178 100

179 15.4 12.0 18.0

176 19.0 15.1 22.7



#74

Anthracene

Concen: 4.177 ng/ul

RT: 17.192 min Scan# 2415

Delta R.T. 0.001 min

Lab File: BM048512.D

Acq: 07 Nov 2024 15:07




Tgt Ion:178 Resp: 139592


Ion Ratio Lower Upper

178 100

179 15.0 12.6 19.0

176 18.1 15.0 22.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 4.430 ng/uL
RT: 13.110 min Scan# 1
Delta R.T. 0.001 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07




Instrument :
BNA_M
ClientSampleId :
SSTD005041


Tgt Ion:216 Resp: 42279
Ion Ratio Lower Upper
216 100
214 79.4 61.9 92.9
218 49.6 38.6 58.0



#76
Pentachlorobenzene
Concen: 4.218 ng/uL
RT: 14.639 min Scan# 1981
Delta R.T. 0.006 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07

Tgt Ion:250 Resp: 40962
Ion Ratio Lower Upper
250 100
248 63.5 49.7 74.5
252 66.5 50.2 75.2









#77
 Carbazole
 Concen: 3.749 ng/ul
 RT: 17.475 min Scan# 2463
 Delta R.T. 0.006 min
 Lab File: BM048512.D
 Acq: 07 Nov 2024 15:07




Instrument :
 BNA_M
 ClientSampleId :
 SSTD005041


Tgt Ion:167 Resp: 110016
 Ion Ratio Lower Upper
 167 100
 166 20.8 16.4 24.6
 139 12.9 10.3 15.5



#78
 Di-n-butylphthalate
 Concen: 3.516 ng/ul
 RT: 18.033 min Scan# 2558
 Delta R.T. 0.006 min
 Lab File: BM048512.D
 Acq: 07 Nov 2024 15:07

Tgt Ion:149 Resp: 135610
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.4 11.2
 104 4.9 4.0 6.0









#80
Fluoranthene
Concen: 3.831 ng/ul
RT: 19.122 min Scan# 21
Delta R.T. 0.001 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07




Instrument :
BNA_M
ClientSampleId :
SSTD005041


Tgt Ion:202 Resp: 158235
Ion Ratio Lower Upper
202 100
101 11.7 8.5 12.7
100 8.6 7.0 10.4



#82
Pyrene
Concen: 3.971 ng/ul
RT: 19.486 min Scan# 2805
Delta R.T. 0.006 min
Lab File: BM048512.D
Acq: 07 Nov 2024 15:07

Tgt Ion:202 Resp: 175896
Ion Ratio Lower Upper
202 100
101 13.2 10.6 15.8
100 10.8 8.9 13.3





#83

Butylbenzylphthalate

Concen: 3.161 ng/ul

RT: 20.386 min Scan# 2958

Delta R.T. 0.000 min

Lab File: BM048512.D

Acq: 07 Nov 2024 15:07

Instrument :

BNA_M

ClientSampleId :

SSTD005041





Tgt Ion:149 Resp: 57855

Ion Ratio Lower Upper

149 100

91 68.5 53.1 79.7

206 21.0 16.6 25.0



#84

3,3'-Dichlorobenzidine

Concen: 3.807 ng/ul

RT: 21.204 min Scan# 3097

Delta R.T. 0.006 min

Lab File: BM048512.D

Acq: 07 Nov 2024 15:07




Tgt Ion:252 Resp: 48593


Ion Ratio Lower Upper

252 100

254 61.5 52.1 78.1

126 10.3 8.3 12.5





#85

Benzo(a)anthracene

Concen: 4.091 ng/ul

RT: 21.268 min Scan# 3107

Delta R.T. 0.006 min

Lab File: BM048512.D

Acq: 07 Nov 2024 15:07

Instrument :

BNA_M

ClientSampleId :

SSTD005041





Tgt Ion:228 Resp: 172600

Ion Ratio Lower Upper

228 100

229 19.3 15.7 23.5

226 26.6 21.7 32.5



#86

Bis(2-ethylhexyl)phthalate

Concen: 3.271 ng/ul

RT: 21.192 min Scan# 3095

Delta R.T. 0.000 min

Lab File: BM048512.D

Acq: 07 Nov 2024 15:07




Tgt Ion:149 Resp: 86282


Ion Ratio Lower Upper

149 100

167 26.6 22.0 33.0

279 3.6 3.1 4.7





#87

Chrysene

Concen: 4.309 ng/ul

RT: 21.327 min Scan# 3118

Delta R.T. 0.000 min

Lab File: BM048512.D

Acq: 07 Nov 2024 15:07

Instrument :

BNA_M

ClientSampleId :

SSTD005041





Tgt Ion:228 Resp: 170947

Ion Ratio Lower Upper

228 100

226 30.6 23.4 35.2

229 18.5 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 2.672 ng/ul




RT: 22.292 min Scan# 3282


Delta R.T. 0.000 min

Lab File: BM048512.D

Acq: 07 Nov 2024 15:07

Tgt Ion:149 Resp: 141816





#90

Benzo(b)fluoranthene

Concen: 3.987 ng/ul

RT: 23.280 min Scan# 3449

Delta R.T. 0.006 min

Lab File: BM048512.D

Acq: 07 Nov 2024 15:07

Instrument :

BNA_M

ClientSampleId :

SSTD005041





Tgt Ion:252 Resp: 180517

Ion Ratio Lower Upper

252 100

253 23.0 17.7 26.5

125 9.9 8.7 13.1



#91

Benzo(k)fluoranthene

Concen: 3.895 ng/ul

RT: 23.339 min Scan# 3460

Delta R.T. 0.000 min

Lab File: BM048512.D

Acq: 07 Nov 2024 15:07




Tgt Ion:252 Resp: 176559


Ion Ratio Lower Upper

252 100

253 22.4 17.4 26.0

125 10.8 8.0 12.0





#93

Benzo(a)pyrene

Concen: 4.024 ng/ul

RT: 24.062 min Scan# 3

Delta R.T. 0.006 min

Lab File: BM048512.D

Acq: 07 Nov 2024 15:07

Instrument :

BNA_M

ClientSampleId :

SSTD005041





Tgt Ion:252 Resp: 165743

Ion Ratio Lower Upper

252 100

253 20.5 17.8 26.8

125 11.1 9.1 13.7



#94

Indeno(1,2,3-cd)pyrene

Concen: 4.326 ng/ul

RT: 27.468 min Scan# 4162

Delta R.T. 0.001 min

Lab File: BM048512.D

Acq: 07 Nov 2024 15:07




Tgt Ion:276 Resp: 214610


Ion Ratio Lower Upper

276 100

138 21.6 16.2 24.2

277 22.6 19.5 29.3









#95
 Dibenzo(a,h)anthracene
 Concen: 4.205 ng/ul
 RT: 27.521 min Scan# 4170
 Delta R.T. 0.006 min
 Lab File: BM048512.D
 Acq: 07 Nov 2024 15:07

Instrument :
 BNA_M
 ClientSampleId :
 SSTD005041

Tgt Ion:278 Resp: 170612
 Ion Ratio Lower Upper
 278 100
 139 15.9 12.6 18.8
 279 23.1 19.0 28.6



#96
 Benzo(g,h,i)perylene
 Concen: 2.056 ng/ul
 RT: 28.486 min Scan# 4335
 Delta R.T. 0.000 min
 Lab File: BM048512.D
 Acq: 07 Nov 2024 15:07

Tgt Ion:276 Resp: 78251
 Ion Ratio Lower Upper
 276 100
 138 22.5 17.6 26.4
 277 25.2 18.6 28.0

