

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110724\
 Data File : BM048518.D
 Acq On : 07 Nov 2024 19:16
 Operator : RC/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV047

Quant Time: Nov 07 22:53:45 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM110724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 07 22:50:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.663	152	156934	20.000	ng/u1	0.00
20) Naphthalene-d8	10.451	136	628030	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.316	164	381374	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.057	188	765102	20.000	ng/u1	0.00
79) Chrysene-d12	21.286	240	751107	20.000	ng/u1	0.00
88) Perylene-d12	24.197	264	896826	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.128	96	28129	7.145	ng/uL	0.00
4) Pyridine-d5	3.546	84	201196	18.545	ng/u1	0.00
7) Phenol-d5	6.845	99	228650	19.140	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.004	67	146815	20.461	ng/u1	0.00
11) 2-Chlorophenol-d4	7.198	132	194694	19.674	ng/u1	0.00
15) 4-Methylphenol-d8	8.381	113	179884	19.150	ng/u1	0.00
21) Nitrobenzene-d5	8.822	128	92866	20.402	ng/u1	0.00
24) 2-Nitrophenol-d4	9.545	143	100400	20.737	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.080	165	185147	20.585	ng/u1	0.00
31) 4-Chloroaniline-d4	10.598	131	252654	19.776	ng/u1	0.00
46) Dimethylphthalate-d6	13.727	166	484004	19.881	ng/u1	0.00
49) Acenaphthylene-d8	14.004	160	586180	20.422	ng/u1	0.00
54) 4-Nitrophenol-d4	14.533	143	89015	19.686	ng/u1	0.00
60) Fluorene-d10	15.310	176	421431	19.831	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.439	200	73418	19.075	ng/u1	0.00
73) Anthracene-d10	17.156	188	647995	20.014	ng/u1	0.00
81) Pyrene-d10	19.450	212	763243	20.248	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.997	264	830875	19.829	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.163	88	29577	6.846	ng/uL	93
5) Pyridine	3.563	79	187323	16.820	ng/u1	97
6) Benzaldehyde	6.810	77	111294	18.852	ng/u1	97
8) Phenol	6.875	94	234406	18.701	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.092	93	185874	18.870	ng/u1	98
12) 2-Chlorophenol	7.234	128	197207	19.236	ng/u1	99
13) 2-Methylphenol	8.116	108	176147	18.984	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.186	45	267668	18.478	ng/u1	96
16) Acetophenone	8.492	105	290514	19.841	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.475	70	135695	19.720	ng/u1	98
18) 4-Methylphenol	8.445	108	195179	19.212	ng/u1	97
19) Hexachloroethane	8.739	117	81759	18.871	ng/u1	98
22) Nitrobenzene	8.863	77	203033	19.672	ng/u1	95
23) Isophorone	9.392	82	374140	19.598	ng/u1	97
25) 2-Nitrophenol	9.575	139	107882	20.446	ng/u1	95
26) 2,4-Dimethylphenol	9.639	107	170566	17.306	ng/u1	96
27) Bis(2-Chloroethoxy)met...	9.875	93	234250	19.658	ng/u1	99
29) 2,4-Dichlorophenol	10.110	162	177734	19.481	ng/u1	99
30) Naphthalene	10.498	128	614600	19.335	ng/u1	100
32) 4-Chloroaniline	10.622	127	242148	20.102	ng/u1	98
33) Hexachlorobutadiene	10.786	225	119646	19.379	ng/u1	97
34) Caprolactam	11.404	113	53630	18.651	ng/u1	96
35) 4-Chloro-3-methylphenol	11.757	107	176142	19.952	ng/u1	99
36) 2-Methylnaphthalene	12.122	142	409375	20.127	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110724\
 Data File : BM048518.D
 Acq On : 07 Nov 2024 19:16
 Operator : RC/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SICV047


Quant Time: Nov 07 22:53:45 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM110724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 07 22:50:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.339	142	382386	18.320	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.492	216	214570	19.634	ng/ul	98
40) Hexachlorocyclopentadiene	12.469	237	100058	17.912	ng/ul	98
41) 2,4,6-Trichlorophenol	12.745	196	136531	19.778	ng/ul	98
42) 2,4,5-Trichlorophenol	12.821	196	147082	20.007	ng/ul	98
43) 1,1'-Biphenyl	13.145	154	512013	19.635	ng/ul	99
44) 2-Chloronaphthalene	13.180	162	402286	19.550	ng/ul	98
45) 2-Nitroaniline	13.398	65	106925	20.692	ng/ul	98
47) Dimethylphthalate	13.774	163	480237	19.638	ng/ul	99
48) 2,6-Dinitrotoluene	13.898	165	103045	22.103	ng/ul	95
50) Acenaphthylene	14.033	152	639591	20.216	ng/ul	100
51) 3-Nitroaniline	14.227	138	111856	22.094	ng/ul	98
52) Acenaphthene	14.380	153	432575	19.421	ng/ul	99
53) 2,4-Dinitrophenol	14.445	184	44680	17.825	ng/ul	95
55) 4-Nitrophenol	14.551	109	62537	19.130	ng/ul	97
56) Dibenzofuran	14.715	168	606307	19.732	ng/ul	99
57) 2,4-Dinitrotoluene	14.692	165	140899	20.792	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	14.945	232	129831	21.231	ng/ul	98
59) Diethylphthalate	15.145	149	474988	19.714	ng/ul	99
61) Fluorene	15.362	166	475837	19.560	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.362	204	232192	19.373	ng/ul	98
63) 4-Nitroaniline	15.398	138	112276	22.522	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.457	198	81664	19.497	ng/ul	94
67) N-Nitrosodiphenylamine	15.574	169	401350	19.903	ng/ul	99
68) 4-Bromophenyl-phenylether	16.257	248	139691	19.359	ng/ul	98
69) Hexachlorobenzene	16.368	284	169725	19.489	ng/ul	98
70) Atrazine	16.533	200	140669	19.167	ng/ul	100
71) Pentachlorophenol	16.715	266	108011	19.515	ng/ul	97
72) Phenanthrene	17.098	178	744223	19.613	ng/ul	99
74) Anthracene	17.192	178	752804	19.768	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.110	216	198240	18.163	ng/uL	99
76) Pentachlorobenzene	14.633	250	203830	19.167	ng/uL	97
77) Carbazole	17.468	167	702002	20.466	ng/ul	99
78) Di-n-butylphthalate	18.027	149	808659	20.052	ng/ul	99
80) Fluoranthene	19.121	202	869217	19.874	ng/ul	99
82) Pyrene	19.480	202	948030	19.990	ng/ul	99
83) Butylbenzylphthalate	20.386	149	367513	20.174	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.197	252	328745	24.089	ng/ul	100
85) Benzo(a)anthracene	21.268	228	932421	19.693	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.192	149	532763	20.047	ng/ul	100
87) Chrysene	21.327	228	893378	19.635	ng/ul	100
89) Di-n-octyl phthalate	22.291	149	930901	18.413	ng/ul	100
90) Benzo(b)fluoranthene	23.274	252	945942	19.161	ng/ul	99
91) Benzo(k)fluoranthene	23.338	252	974297	19.793	ng/ul	100
93) Benzo(a)pyrene	24.056	252	926867	19.969	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	27.468	276	1180795	19.575	ng/ul	97
95) Dibenzo(a,h)anthracene	27.515	278	971948	19.727	ng/ul	99
96) Benzo(g,h,i)perylene	28.485	276	893866	18.899	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 07 22:53:45 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM110724.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Nov 07 22:50:20 2024
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 6.846 ng/uL
RT: 3.163 min Scan# 30
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16




Instrument :
BNA_M
ClientSampleId :
SICV047


Tgt Ion:	88	Resp:	29577
Ion	Ratio	Lower	Upper
88	100		
43	24.8	22.1	33.1
58	69.6	60.7	91.1



#5
Pyridine
Concen: 16.820 ng/uL
RT: 3.563 min Scan# 98
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16

Tgt Ion:	79	Resp:	187323
Ion	Ratio	Lower	Upper
79	100		
52	74.7	62.6	94.0
51	35.7	27.8	41.8









#6
Benzaldehyde
Concen: 18.852 ng/ul
RT: 6.810 min Scan# 650
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16




Instrument :
BNA_M
ClientSampleId :
SICV047


Tgt Ion: 77 Resp: 111294
Ion Ratio Lower Upper
77 100
105 100.0 78.0 117.0
106 99.3 76.6 115.0



#8
Phenol
Concen: 18.701 ng/ul
RT: 6.875 min Scan# 661
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16

Tgt Ion: 94 Resp: 234406
Ion Ratio Lower Upper
94 100
65 25.0 20.8 31.2
66 32.4 26.1 39.1









#10
Bis(2-Chloroethyl)ether
Concen: 18.870 ng/ul
RT: 7.092 min Scan# 698
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16

Instrument :
BNA_M
ClientSampleId :
SICV047

Tgt Ion: 93 Resp: 185874




Ion	Ratio	Lower	Upper
93	100		
63	72.1	59.6	89.4
95	32.0	25.0	37.6




#12
2-Chlorophenol
Concen: 19.236 ng/ul
RT: 7.234 min Scan# 722
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16

Tgt Ion: 128 Resp: 197207

Ion	Ratio	Lower	Upper
128	100		
64	42.6	35.2	52.8
130	32.1	25.8	38.6









#13
2-Methylphenol
Concen: 18.984 ng/ul
RT: 8.116 min Scan# 872
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16




Instrument :
BNA_M
ClientSampleId :
SICV047


Tgt Ion:108 Resp: 176147
Ion Ratio Lower Upper
108 100
107 89.9 71.5 107.3



#14
2,2'-oxybis(1-Chloropropane)
Concen: 18.478 ng/ul
RT: 8.186 min Scan# 884
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16

Tgt Ion: 45 Resp: 267668
Ion Ratio Lower Upper
45 100
77 14.5 10.6 16.0
79 11.0 7.4 11.2









#16
Acetophenone
Concen: 19.841 ng/ul
RT: 8.492 min Scan# 911
Delta R.T. 0.006 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16




Instrument :
BNA_M
ClientSampleId :
SICV047


Tgt Ion:105 Resp: 290514
Ion Ratio Lower Upper
105 100
77 72.9 59.7 89.5
51 25.3 20.6 31.0



#17
N-Nitroso-di-n-propylamine
Concen: 19.720 ng/ul
RT: 8.475 min Scan# 933
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16

Tgt Ion: 70 Resp: 135695
Ion Ratio Lower Upper
70 100
42 56.5 47.0 70.4
101 9.9 8.5 12.7
130 23.5 18.4 27.6





#18

4-Methylphenol

Concen: 19.212 ng/ul

RT: 8.445 min Scan# 91

Delta R.T. 0.000 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16

Instrument :

BNA_M

ClientSampleId :





SICV047

Tgt Ion:108 Resp: 195179

Ion Ratio Lower Upper

108 100

107 108.0 88.8 133.2



#19

Hexachloroethane

Concen: 18.871 ng/ul

RT: 8.739 min Scan# 978

Delta R.T. 0.000 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16




Tgt Ion:117 Resp: 81759


Ion Ratio Lower Upper

117 100

201 107.5 83.8 125.8

199 65.0 52.1 78.1









#22
Nitrobenzene
Concen: 19.672 ng/ul
RT: 8.863 min Scan# 999
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16




Instrument :
BNA_M
ClientSampleId :
SICV047


Tgt Ion: 77 Resp: 203033
Ion Ratio Lower Upper
77 100
123 49.1 36.6 55.0
65 13.3 9.5 14.3



#23
Isophorone
Concen: 19.598 ng/ul
RT: 9.392 min Scan# 1089
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16

Tgt Ion: 82 Resp: 374140
Ion Ratio Lower Upper
82 100
95 6.7 5.0 7.6
138 19.2 14.2 21.4









#25
2-Nitrophenol
Concen: 20.446 ng/ul
RT: 9.575 min Scan# 1120
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16




Instrument :
BNA_M
ClientSampleId :
SICV047


Tgt Ion:139 Resp: 107882
Ion Ratio Lower Upper
139 100
65 42.9 37.7 56.5
109 26.8 22.3 33.5



#26
2,4-Dimethylphenol
Concen: 17.306 ng/ul
RT: 9.639 min Scan# 1131
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16

Tgt Ion:107 Resp: 170566
Ion Ratio Lower Upper
107 100
121 51.6 42.9 64.3
122 91.1 69.8 104.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 19.658 ng/ul
 RT: 9.875 min Scan# 1171
 Delta R.T. 0.000 min
 Lab File: BM048518.D
 Acq: 07 Nov 2024 19:16




Instrument :
 BNA_M
 ClientSampleId :
 SICV047


Tgt Ion: 93 Resp: 234250
 Ion Ratio Lower Upper
 93 100
 95 32.6 25.3 37.9
 123 11.9 9.5 14.3



#29
 2,4-Dichlorophenol
 Concen: 19.481 ng/ul
 RT: 10.110 min Scan# 1211
 Delta R.T. 0.000 min
 Lab File: BM048518.D
 Acq: 07 Nov 2024 19:16

Tgt Ion: 162 Resp: 177734
 Ion Ratio Lower Upper
 162 100
 164 67.6 54.6 82.0
 98 34.0 28.2 42.2





#30

Naphthalene

Concen: 19.335 ng/ul

RT: 10.498 min Scan# 1277

Delta R.T. -0.006 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16

Instrument :

BNA_M

ClientSampleId :

SICV047





Tgt Ion:128 Resp: 614600

Ion Ratio Lower Upper

128 100

129 10.6 8.5 12.7

127 12.8 10.0 15.0



#32

4-Chloroaniline

Concen: 20.102 ng/ul

RT: 10.622 min Scan# 1298

Delta R.T. 0.000 min

Lab File: BM048518.D




Acq: 07 Nov 2024 19:16


Tgt Ion:127 Resp: 242148

Ion Ratio Lower Upper

127 100

129 31.9 26.2 39.4









#33
Hexachlorobutadiene
Concen: 19.379 ng/ul
RT: 10.786 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16




Instrument :
BNA_M
ClientSampleId :
SICV047


Tgt Ion:225 Resp: 119646
Ion Ratio Lower Upper
225 100
223 64.9 48.5 72.7
227 63.4 51.5 77.3



#34
Caprolactam
Concen: 18.651 ng/ul
RT: 11.404 min Scan# 1431
Delta R.T. 0.006 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16

Tgt Ion:113 Resp: 53630
Ion Ratio Lower Upper
113 100
55 168.1 138.5 207.7
56 125.0 104.1 156.1









#35
4-Chloro-3-methylphenol
Concen: 19.952 ng/ul
RT: 11.757 min Scan# 1491
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16




Instrument :
BNA_M
ClientSampleId :
SICV047


Tgt Ion:107 Resp: 176142
Ion Ratio Lower Upper
107 100
144 27.3 20.6 31.0
142 83.9 66.7 100.1



#36
2-Methylnaphthalene
Concen: 20.127 ng/ul
RT: 12.122 min Scan# 1553
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16

Tgt Ion:142 Resp: 409375
Ion Ratio Lower Upper
142 100
141 88.7 71.6 107.4





#37

1-Methylnaphthalene

Concen: 18.320 ng/ul

RT: 12.339 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16

Instrument :

BNA_M

ClientSampleId :

SICV047





Tgt Ion:142 Resp: 382386

Ion Ratio Lower Upper

142 100

141 92.3 73.6 110.4

116 3.7 2.7 4.1



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.634 ng/ul

RT: 12.492 min Scan# 1616

Delta R.T. -0.006 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16

Tgt Ion:216 Resp: 214570




Ion Ratio Lower Upper


216 100

214 79.5 62.6 93.8

179 21.1 16.5 24.7

108 18.1 12.9 19.3









#40
 Hexachlorocyclopentadiene
 Concen: 17.912 ng/ul
 RT: 12.469 min Scan# 1
 Delta R.T. -0.006 min
 Lab File: BM048518.D
 Acq: 07 Nov 2024 19:16




Instrument :
 BNA_M
 ClientSampleId :
 SICV047


Tgt Ion:237 Resp: 100058
 Ion Ratio Lower Upper
 237 100
 235 61.2 50.0 75.0
 272 12.1 10.6 16.0



#41
 2,4,6-Trichlorophenol
 Concen: 19.778 ng/ul
 RT: 12.745 min Scan# 1659
 Delta R.T. 0.000 min
 Lab File: BM048518.D
 Acq: 07 Nov 2024 19:16

Tgt Ion:196 Resp: 136531
 Ion Ratio Lower Upper
 196 100
 198 97.5 76.6 114.8
 200 31.5 25.8 38.6









#42
2,4,5-Trichlorophenol
Concen: 20.007 ng/ul
RT: 12.821 min Scan# 1
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16




Instrument :
BNA_M
ClientSampleId :
SICV047


Tgt Ion:196 Resp: 147082
Ion Ratio Lower Upper
196 100
198 96.0 78.7 118.1
200 32.0 24.6 37.0



#43
1,1'-Biphenyl
Concen: 19.635 ng/ul
RT: 13.145 min Scan# 1727
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16

Tgt Ion:154 Resp: 512013
Ion Ratio Lower Upper
154 100
153 39.1 31.4 47.2
76 11.7 9.9 14.9





#44

2-Chloronaphthalene

Concen: 19.550 ng/ul

RT: 13.180 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16

Instrument :

BNA_M

ClientSampleId :

SICV047





Tgt Ion:162 Resp: 402286

Ion Ratio Lower Upper

162 100

164 32.8 25.8 38.6

127 38.1 29.5 44.3



#45

2-Nitroaniline

Concen: 20.692 ng/ul

RT: 13.398 min Scan# 1770

Delta R.T. 0.000 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16




Tgt Ion: 65 Resp: 106925


Ion Ratio Lower Upper

65 100

92 70.0 56.2 84.2

138 114.9 89.6 134.4









#47
Dimethylphthalate
Concen: 19.638 ng/ul
RT: 13.774 min Scan# 1834
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16




Instrument :
BNA_M
ClientSampleId :
SICV047


Tgt Ion:163 Resp: 480237
Ion Ratio Lower Upper
163 100
194 3.8 3.0 4.4
164 10.3 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 22.103 ng/ul
RT: 13.898 min Scan# 1855
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16

Tgt Ion:165 Resp: 103045
Ion Ratio Lower Upper
165 100
89 51.4 45.4 68.0
121 21.3 17.1 25.7





#50

Acenaphthylene

Concen: 20.216 ng/ul

RT: 14.033 min Scan# 11

Delta R.T. 0.000 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16

Instrument :

BNA_M

ClientSampleId :

SICV047





Tgt Ion:152 Resp: 639591

Ion Ratio Lower Upper

152 100

151 19.9 15.8 23.6

153 13.1 10.5 15.7



#51

3-Nitroaniline

Concen: 22.094 ng/ul

RT: 14.227 min Scan# 1911

Delta R.T. -0.006 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16




Tgt Ion:138 Resp: 111856


Ion Ratio Lower Upper

138 100

108 10.3 7.8 11.8

92 111.8 91.0 136.4









#52
Acenaphthene
Concen: 19.421 ng/ul
RT: 14.380 min Scan# 1937
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16




Instrument :
BNA_M
ClientSampleId :
SICV047


Tgt Ion:153 Resp: 432575
Ion Ratio Lower Upper
153 100
152 48.0 38.3 57.5
154 87.2 71.1 106.7



#53
2,4-Dinitrophenol
Concen: 17.825 ng/ul
RT: 14.445 min Scan# 1948
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16

Tgt Ion:184 Resp: 44680
Ion Ratio Lower Upper
184 100
63 60.1 52.5 78.7
154 61.2 51.4 77.0









#55
4-Nitrophenol
Concen: 19.130 ng/ul
RT: 14.551 min Scan# 1966
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16




Instrument :
BNA_M
ClientSampleId :
SICV047


Tgt Ion:109 Resp: 62537
Ion Ratio Lower Upper
109 100
139 147.8 115.9 173.9
65 128.8 107.0 160.4



#56
Dibenzofuran
Concen: 19.732 ng/ul
RT: 14.715 min Scan# 1994
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16

Tgt Ion:168 Resp: 606307
Ion Ratio Lower Upper
168 100
139 36.3 29.7 44.5









#57
 2,4-Dinitrotoluene
 Concen: 20.792 ng/ul
 RT: 14.692 min Scan# 1990
 Delta R.T. 0.000 min
 Lab File: BM048518.D
 Acq: 07 Nov 2024 19:16




Instrument :
 BNA_M
 ClientSampleId :
 SICV047


Tgt Ion:165 Resp: 140899
 Ion Ratio Lower Upper
 165 100
 63 40.5 34.6 51.8
 182 2.6 2.4 3.6



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 21.231 ng/ul
 RT: 14.945 min Scan# 2033
 Delta R.T. 0.000 min
 Lab File: BM048518.D
 Acq: 07 Nov 2024 19:16

Tgt Ion:232 Resp: 129831
 Ion Ratio Lower Upper
 232 100
 131 46.7 36.3 54.5
 130 2.3 1.8 2.8
 166 31.0 24.2 36.2





#59

Diethylphthalate

Concen: 19.714 ng/ul

RT: 15.145 min Scan# 2067

Delta R.T. 0.000 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16

Instrument :

BNA_M

ClientSampleId :

SICV047





Tgt Ion:149 Resp: 474988

Ion Ratio Lower Upper

149 100

177 22.7 17.8 26.8

150 12.0 9.6 14.4



#61

Fluorene

Concen: 19.560 ng/ul

RT: 15.362 min Scan# 2104

Delta R.T. 0.000 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16




Tgt Ion:166 Resp: 475837


Ion Ratio Lower Upper

166 100

165 98.5 78.7 118.1

167 13.4 11.0 16.6









#62
4-Chlorophenyl-phenylether
Concen: 19.373 ng/ul
RT: 15.362 min Scan# 2104
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16




Instrument :
BNA_M
ClientSampleId :
SICV047


Tgt Ion:204 Resp: 232192
Ion Ratio Lower Upper
204 100
206 32.4 26.0 39.0
141 54.5 45.4 68.2



#63
4-Nitroaniline
Concen: 22.522 ng/ul
RT: 15.398 min Scan# 2110
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16

Tgt Ion:138 Resp: 112276
Ion Ratio Lower Upper
138 100
92 48.3 39.7 59.5
108 69.4 58.5 87.7









#66
 4,6-Dinitro-2-methylphenol
 Concen: 19.497 ng/ul
 RT: 15.457 min Scan# 2120
 Delta R.T. 0.000 min
 Lab File: BM048518.D
 Acq: 07 Nov 2024 19:16




Instrument :
 BNA_M
 ClientSampleId :
 SICV047


Tgt Ion:198 Resp: 81664
 Ion Ratio Lower Upper
 198 100
 182 5.4 5.0 7.6
 77 21.5 20.1 30.1



#67
 N-Nitrosodiphenylamine
 Concen: 19.903 ng/ul
 RT: 15.574 min Scan# 2140
 Delta R.T. 0.000 min
 Lab File: BM048518.D
 Acq: 07 Nov 2024 19:16

Tgt Ion:169 Resp: 401350
 Ion Ratio Lower Upper
 169 100
 168 65.4 52.6 79.0
 167 35.7 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 19.359 ng/ul
RT: 16.257 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16




Instrument :
BNA_M
ClientSampleId :
SICV047


Tgt Ion:248 Resp: 139691
Ion Ratio Lower Upper
248 100
250 96.1 79.4 119.2
141 68.2 53.8 80.6



#69
Hexachlorobenzene
Concen: 19.489 ng/ul
RT: 16.368 min Scan# 2275
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16

Tgt Ion:284 Resp: 169725
Ion Ratio Lower Upper
284 100
142 33.2 27.5 41.3
249 30.5 23.2 34.8





#70

Atrazine

Concen: 19.167 ng/ul

RT: 16.533 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16

Instrument :

BNA_M

ClientSampleId :

SICV047





Tgt Ion:200 Resp: 140669

Ion Ratio Lower Upper

200 100

173 24.8 20.2 30.2

215 47.7 38.2 57.2



#71

Pentachlorophenol

Concen: 19.515 ng/ul

RT: 16.715 min Scan# 2334

Delta R.T. 0.000 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16




Tgt Ion:266 Resp: 108011


Ion Ratio Lower Upper

266 100

264 62.4 51.1 76.7

268 63.2 53.2 79.8





#72

Phenanthrene

Concen: 19.613 ng/ul

RT: 17.098 min Scan# 2400

Delta R.T. -0.006 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16

Instrument :

BNA_M

ClientSampleId :

SICV047





Tgt Ion:178 Resp: 744223

Ion Ratio Lower Upper

178 100

179 15.4 12.0 18.0

176 19.2 15.1 22.7



#74

Anthracene

Concen: 19.768 ng/ul

RT: 17.192 min Scan# 2415

Delta R.T. 0.000 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16




Tgt Ion:178 Resp: 752804


Ion Ratio Lower Upper

178 100

179 15.5 12.6 19.0

176 18.6 15.0 22.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 18.163 ng/uL
RT: 13.110 min Scan# 1721
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16




Instrument :
BNA_M
ClientSampleId :
SICV047


Tgt Ion:216 Resp: 198240
Ion Ratio Lower Upper
216 100
214 77.2 61.9 92.9
218 49.0 38.6 58.0



#76
Pentachlorobenzene
Concen: 19.167 ng/uL
RT: 14.633 min Scan# 1980
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16

Tgt Ion:250 Resp: 203830
Ion Ratio Lower Upper
250 100
248 64.8 49.7 74.5
252 64.9 50.2 75.2









#77
 Carbazole
 Concen: 20.466 ng/ul
 RT: 17.468 min Scan# 2462
 Delta R.T. 0.000 min
 Lab File: BM048518.D
 Acq: 07 Nov 2024 19:16




Instrument :
 BNA_M
 ClientSampleId :
 SICV047


Tgt Ion:167 Resp: 702002
 Ion Ratio Lower Upper
 167 100
 166 21.4 16.4 24.6
 139 12.6 10.3 15.5



#78
 Di-n-butylphthalate
 Concen: 20.052 ng/ul
 RT: 18.027 min Scan# 2557
 Delta R.T. 0.000 min
 Lab File: BM048518.D
 Acq: 07 Nov 2024 19:16

Tgt Ion:149 Resp: 808659
 Ion Ratio Lower Upper
 149 100
 150 8.9 7.4 11.2
 104 4.7 4.0 6.0





#80

Fluoranthene

Concen: 19.874 ng/ul

RT: 19.121 min Scan# 21

Delta R.T. 0.000 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16

Instrument :

BNA_M

ClientSampleId :

SICV047





Tgt Ion:202 Resp: 869217

Ion Ratio Lower Upper

202 100

101 11.1 8.5 12.7

100 8.5 7.0 10.4



#82

Pyrene

Concen: 19.990 ng/ul

RT: 19.480 min Scan# 2804

Delta R.T. 0.000 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16




Tgt Ion:202 Resp: 948030


Ion Ratio Lower Upper

202 100

101 13.0 10.6 15.8

100 10.6 8.9 13.3









#83
Butylbenzylphthalate
Concen: 20.174 ng/ul
RT: 20.386 min Scan# 2958
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16




Instrument :
BNA_M
ClientSampleId :
SICV047


Tgt Ion:149 Resp: 367513
Ion Ratio Lower Upper
149 100
91 64.6 53.1 79.7
206 20.6 16.6 25.0



#84
3,3'-Dichlorobenzidine
Concen: 24.089 ng/ul
RT: 21.197 min Scan# 3096
Delta R.T. 0.000 min
Lab File: BM048518.D
Acq: 07 Nov 2024 19:16

Tgt Ion:252 Resp: 328745
Ion Ratio Lower Upper
252 100
254 64.7 52.1 78.1
126 10.4 8.3 12.5





#85

Benzo(a)anthracene

Concen: 19.693 ng/ul

RT: 21.268 min Scan# 3107

Delta R.T. 0.006 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16

Instrument :

BNA_M

ClientSampleId :

SICV047





Tgt Ion:228 Resp: 932421

Ion Ratio Lower Upper

228 100

229 19.6 15.7 23.5

226 26.4 21.7 32.5



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.047 ng/ul

RT: 21.192 min Scan# 3095

Delta R.T. 0.000 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16




Tgt Ion:149 Resp: 532763


Ion Ratio Lower Upper

149 100

167 27.6 22.0 33.0

279 4.2 3.1 4.7





#87

Chrysene

Concen: 19.635 ng/ul

RT: 21.327 min Scan# 3118

Delta R.T. 0.000 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16

Instrument :

BNA_M

ClientSampleId :

SICV047





Tgt Ion:228 Resp: 893378

Ion Ratio Lower Upper

228 100

226 29.2 23.4 35.2

229 19.5 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 18.413 ng/ul




RT: 22.291 min Scan# 3282


Delta R.T. 0.000 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16

Tgt Ion:149 Resp: 930901





#90

Benzo(b)fluoranthene

Concen: 19.161 ng/ul

RT: 23.274 min Scan# 3449

Delta R.T. 0.000 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16

Instrument :

BNA_M

ClientSampleId :

SICV047





Tgt Ion:252 Resp: 945942

Ion Ratio Lower Upper

252 100

253 22.4 17.7 26.5

125 10.4 8.7 13.1



#91

Benzo(k)fluoranthene

Concen: 19.793 ng/ul

RT: 23.338 min Scan# 3460

Delta R.T. 0.000 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16




Tgt Ion:252 Resp: 974297


Ion Ratio Lower Upper

252 100

253 21.7 17.4 26.0

125 10.3 8.0 12.0





#93

Benzo(a)pyrene

Concen: 19.969 ng/ul

RT: 24.056 min Scan# 3

Delta R.T. 0.000 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16

Instrument :

BNA_M

ClientSampleId :

SICV047





Tgt Ion:252 Resp: 926867

Ion Ratio Lower Upper

252 100

253 21.7 17.8 26.8

125 11.2 9.1 13.7



#94

Indeno(1,2,3-cd)pyrene

Concen: 19.575 ng/ul

RT: 27.468 min Scan# 4162

Delta R.T. 0.000 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16




Tgt Ion:276 Resp: 1180795


Ion Ratio Lower Upper

276 100

138 22.2 16.2 24.2

277 23.6 19.5 29.3





#95

Dibenzo(a,h)anthracene

Concen: 19.727 ng/ul

RT: 27.515 min Scan# 4170

Delta R.T. 0.000 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16

Instrument :

BNA_M

ClientSampleId :

SICV047





Tgt Ion:278 Resp: 971948

Ion Ratio Lower Upper

278 100

139 16.0 12.6 18.8

279 24.3 19.0 28.6



#96

Benzo(g,h,i)perylene

Concen: 18.899 ng/ul

RT: 28.485 min Scan# 4335

Delta R.T. 0.000 min

Lab File: BM048518.D

Acq: 07 Nov 2024 19:16

Tgt Ion:276 Resp: 893866

Ion Ratio Lower Upper

276 100

138 22.2 17.6 26.4

277 23.7 18.6 28.0

