


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM110817\
 Data File : BM012813.D
 Acq On : 08 Nov 2017 09:42
 Operator : SJ/JU
 Sample : DFTPP44
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP44

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM110417MA.M
 Title : SVOA CALIBRATION
 Last Update : Thu Nov 09 05:48:58 2017



AutoFind: Scans 2470, 2471, 2472; Background Corrected with Scan 2465

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	24.6	54354	PASS
68	69	0.00	2	2.0	1331	PASS
69	198	0.00	100	30.4	67210	PASS
70	69	0.00	2	0.6	381	PASS
127	198	10	80	38.9	86080	PASS
197	198	0.00	2	0.5	1120	PASS
198	198	100	100	100.0	221333	PASS
199	198	5	9	6.8	14998	PASS
275	198	10	60	32.9	72816	PASS
365	198	1	100	4.0	8958	PASS
441	443	0.01	100	78.8	28082	PASS
442	198	50	100	82.6	182877	PASS
443	442	15	24	19.5	35656	PASS

m/z	Abundance
38.00	845.0
39.00	2955.0
40.00	370.0
43.90	497.0
49.10	343.0
50.00	14719.0
51.00	55304.0
52.10	2743.0
56.00	2047.0
57.00	4163.0
61.00	624.0
62.00	917.0
63.10	2395.0
64.10	310.0
65.00	1031.0
68.10	1564.0
69.00	64696.0
70.00	318.0
73.10	874.0
74.00	7495.0
75.00	10576.0
76.00	3450.0
77.00	70304.0
78.10	5023.0
79.00	5502.0
80.00	3769.0
81.00	5362.0
82.00	1221.0
83.00	1346.0
84.00	332.0
85.00	1131.0
86.00	1770.0
86.90	681.0
91.00	1317.0
92.00	1647.0
93.00	9621.0
93.90	712.0
96.00	311.0
98.00	7159.0
99.00	5520.0
99.90	545.0
100.90	3092.0
103.00	1186.0
104.00	1947.0
105.00	1802.0
106.00	470.0
107.00	24592.0
108.00	3759.0
108.90	571.0
110.00	41872.0
111.00	7095.0
112.00	874.0
113.10	350.0
115.90	1322.0
117.00	19384.0
118.00	2014.0
121.90	1705.0
123.00	2156.0
124.00	1090.0
125.00	963.0
127.00	78232.0
128.00	6682.0
129.00	31168.0
130.00	3001.0
130.90	534.0
132.00	328.0
132.90	317.0
133.90	987.0
135.00	2619.0
136.00	1172.0
137.00	1419.0
140.00	398.0
141.00	4469.0
142.00	1412.0
142.90	692.0
146.00	797.0
147.00	2201.0
148.00	5078.0
148.90	936.0
149.90	387.0
150.90	383.0
151.80	402.0
153.00	1634.0
154.00	1227.0
155.10	2248.0
156.00	3396.0
157.10	704.0
157.80	826.0
158.90	736.0
160.00	1493.0
161.00	2327.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

162.00	699.0
163.00	376.0
164.90	1748.0
166.00	1241.0
167.00	9176.0
168.00	4790.0
169.10	734.0
170.10	378.0
170.80	327.0
171.90	833.0
173.00	878.0
173.90	1996.0
175.00	3677.0
175.90	1136.0
177.00	1691.0
177.90	700.0
179.00	6719.0
180.00	5232.0
181.00	2498.0
182.10	450.0
182.90	323.0
183.80	654.0
185.00	3513.0
186.00	26256.0
187.00	8208.0
188.00	879.0
189.00	1738.0
190.90	934.0
192.00	2157.0
193.00	2435.0
193.80	562.0
194.90	441.0
196.00	4982.0
197.10	1304.0
198.00	188864.0
198.90	13276.0
199.90	1023.0
201.40	689.0
201.70	722.0
203.00	1142.0
204.00	6489.0
205.00	11456.0
206.00	49472.0
207.00	7731.0
208.00	1689.0
209.10	788.0
210.10	690.0
210.90	1959.0
212.00	369.0
215.00	607.0
216.00	1183.0
217.00	13822.0
218.00	1985.0
221.00	8188.0
222.00	1689.0
223.00	3180.0
224.00	28056.0
225.00	7724.0
226.00	1114.0
227.00	12096.0
228.00	1604.0
229.00	2566.0
229.90	419.0
230.90	1177.0
233.90	675.0
235.00	836.0
235.90	687.0
236.90	1185.0
238.90	438.0
240.00	585.0
241.00	543.0
241.90	1244.0
243.10	1500.0
244.00	22576.0
245.10	3037.0
246.00	4790.0
246.90	924.0
248.90	937.0
253.00	557.0
255.00	120968.0
256.00	17928.0
257.00	1554.0
258.00	6590.0
259.00	1210.0
265.00	3312.0
266.00	341.0
272.00	307.0
273.00	3604.0
274.00	9653.0
275.00	59440.0
276.00	8511.0
277.00	4593.0
277.90	1203.0
278.90	303.0
281.00	410.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

282.90	560.0
283.90	600.0
285.00	823.0
292.90	785.0
293.90	356.0
296.00	15805.0
297.00	2231.0
302.00	471.0
303.00	1629.0
304.00	504.0
314.10	728.0
315.00	1536.0
316.00	1030.0
321.00	424.0
323.00	4668.0
324.10	951.0
327.00	1034.0
327.90	471.0
332.10	385.0
333.00	465.0
334.00	2942.0
335.00	1057.0
341.00	652.0
346.00	855.0
352.10	1691.0
353.10	976.0
354.00	1672.0
365.00	6099.0
365.80	591.0
372.00	2396.0
373.00	647.0
383.00	700.0
402.00	1055.0
402.90	1023.0
421.00	1045.0
422.00	1130.0
423.10	7404.0
424.00	2021.0
441.10	18712.0
442.10	120792.0
443.10	24360.0
444.00	2280.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

m/z	Abundance
36.90	316.0
38.00	931.0
39.00	4526.0
40.00	409.0
43.90	544.0
49.10	815.0
50.10	17240.0
51.10	63776.0
52.00	3516.0
55.00	662.0
56.00	1923.0
57.00	4712.0
61.00	808.0
61.90	959.0
63.00	3174.0
64.00	535.0
65.00	1235.0
68.10	1365.0
69.00	81624.0
70.00	481.0
73.00	1074.0
74.00	8991.0
75.00	12980.0
76.10	4340.0
77.10	87536.0
78.10	5737.0
79.00	6213.0
80.00	5103.0
81.00	6752.0
82.00	1637.0
83.00	1568.0
85.00	1166.0
86.00	2234.0
87.10	834.0
87.90	348.0
91.00	1655.0
92.00	1852.0
93.00	11415.0
94.00	818.0
96.10	575.0
98.00	8936.0
99.00	6827.0
99.90	506.0
101.00	3736.0
102.90	1633.0
104.00	2828.0
105.10	2338.0
106.00	854.0
107.00	28992.0
108.00	4373.0
109.10	940.0
110.00	54608.0
111.00	8443.0
112.10	1053.0
113.00	512.0
116.10	1875.0
117.00	27056.0
118.00	2043.0
119.00	307.0
120.00	490.0
122.00	2380.0
123.00	3434.0
124.00	1074.0
125.00	1422.0
127.00	105240.0
128.00	8572.0
129.00	42248.0
129.90	3627.0
130.90	775.0
131.90	339.0
134.00	1330.0
134.90	3615.0
136.00	1522.0
137.10	1925.0
137.70	488.0
138.10	422.0
139.90	854.0
141.00	5743.0
142.00	2080.0
142.90	1080.0
144.10	385.0
145.00	375.0
145.90	1062.0
147.00	3370.0
148.00	7118.0
149.00	1228.0
151.00	595.0
151.70	339.0
152.10	374.0
153.00	2027.0
154.00	1501.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

155.00	3415.0
156.00	5586.0
157.10	873.0
158.00	1283.0
158.90	721.0
160.00	1976.0
161.00	2854.0
162.00	841.0
164.00	440.0
165.00	2334.0
166.00	2165.0
167.00	11745.0
168.00	6002.0
168.90	1110.0
170.00	354.0
171.00	605.0
172.00	919.0
172.90	1270.0
174.00	2636.0
175.00	4860.0
176.00	1612.0
177.00	2928.0
177.90	1070.0
179.00	9600.0
180.00	6895.0
181.00	3150.0
181.90	535.0
182.90	336.0
183.90	548.0
185.00	5060.0
186.00	35608.0
187.00	10968.0
187.90	886.0
189.00	2529.0
191.00	1317.0
192.00	2759.0
193.00	3658.0
194.00	923.0
195.20	485.0
196.00	7376.0
196.70	1801.0
197.10	2056.0
198.00	266048.0
199.00	17936.0
199.90	1431.0
201.50	1673.0
203.00	1802.0
204.00	9361.0
205.00	16231.0
206.00	74840.0
207.00	10678.0
208.00	2516.0
209.00	1000.0
209.90	1337.0
211.00	2697.0
211.80	688.0
214.80	673.0
216.00	1491.0
217.00	18416.0
218.00	2485.0
221.00	11274.0
223.00	3821.0
224.00	42248.0
225.00	10397.0
226.10	1184.0
227.00	16896.0
228.00	2570.0
229.00	3905.0
229.90	590.0
230.90	1368.0
233.90	1121.0
235.00	1185.0
236.00	885.0
237.00	1284.0
239.00	873.0
240.00	384.0
241.00	962.0
242.10	2121.0
243.00	2365.0
244.10	32488.0
245.00	4306.0
246.00	6935.0
247.00	1271.0
249.00	1603.0
250.00	305.0
252.00	307.0
252.90	732.0
254.20	1686.0
255.00	182528.0
256.00	26944.0
257.00	2301.0
258.00	9786.0
259.10	1591.0
260.00	301.0
261.00	426.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

263.90	312.0
265.00	4435.0
266.10	575.0
272.00	538.0
273.00	5432.0
274.00	13693.0
275.00	86576.0
276.00	11990.0
277.00	7109.0
277.90	1204.0
279.00	416.0
281.00	511.0
283.10	863.0
284.00	500.0
285.00	1189.0
285.90	396.0
291.90	411.0
293.00	1591.0
294.00	400.0
295.10	570.0
296.00	23752.0
297.00	3095.0
301.00	445.0
302.10	375.0
303.10	2628.0
304.00	860.0
310.00	371.0
314.10	1321.0
315.00	2807.0
316.00	1447.0
321.00	862.0
322.00	709.0
323.10	7410.0
324.10	1233.0
327.00	1543.0
328.00	467.0
332.00	504.0
333.00	807.0
334.00	4604.0
335.00	1403.0
341.00	1092.0
342.00	348.0
346.00	1335.0
346.90	436.0
352.00	2549.0
353.00	1846.0
354.00	2733.0
355.00	444.0
365.00	10380.0
366.00	1468.0
370.90	561.0
372.00	3847.0
372.90	721.0
383.10	895.0
383.90	359.0
390.00	425.0
391.10	504.0
402.00	1310.0
403.00	2321.0
404.00	733.0
421.10	1918.0
422.00	1581.0
423.00	12091.0
424.10	3132.0
425.10	385.0
441.10	32192.0
442.10	211136.0
443.10	42264.0
444.10	3540.0

Instrument :

BNA_M

ClientSampleId :

DFTPP44

m/z	Abundance
38.10	451.0
39.00	2972.0
44.00	476.0
49.10	459.0
50.10	12024.0
51.00	43984.0
52.00	2082.0
54.90	399.0
56.00	1396.0
57.00	3212.0
61.00	624.0
61.90	925.0
63.00	2470.0
65.00	1160.0
68.10	1065.0
69.00	55312.0
70.00	346.0
73.00	897.0
74.00	6604.0
75.00	8693.0
76.10	3434.0
77.00	60704.0
78.00	4115.0
79.00	5123.0
80.00	3598.0
81.00	5021.0
82.00	1394.0
83.00	845.0
85.00	765.0
86.00	1618.0
86.90	745.0
91.00	1379.0
91.90	1236.0
93.00	8338.0
94.00	584.0
96.00	327.0
98.00	5807.0
99.00	4389.0
100.00	543.0
101.00	2838.0
102.90	866.0
104.00	1766.0
105.00	1871.0
106.00	671.0
107.00	21368.0
108.00	3591.0
109.00	537.0
110.00	39152.0
111.00	5809.0
112.10	558.0
116.10	1761.0
117.00	18920.0
118.10	1214.0
122.00	1595.0
123.00	2237.0
123.90	1023.0
125.00	902.0
127.00	74768.0
128.00	5860.0
129.00	29264.0
130.00	2154.0
131.10	486.0
132.00	302.0
133.90	786.0
135.00	2637.0
136.10	1169.0
137.00	1626.0
140.00	432.0
141.00	3984.0
142.00	1304.0
142.90	1005.0
144.00	383.0
146.00	997.0
147.00	2400.0
148.00	4600.0
149.00	973.0
150.00	343.0
151.10	489.0
152.90	1389.0
154.00	1083.0
155.00	2224.0
156.10	3751.0
157.00	692.0
157.90	924.0
159.00	581.0
159.90	1370.0
161.00	1912.0
161.90	788.0
165.00	1810.0
166.00	1484.0
167.00	8860.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

168.00	4203.0
169.00	680.0
170.00	362.0
171.00	405.0
172.00	696.0
173.00	1003.0
174.00	1995.0
175.00	4239.0
176.00	1046.0
176.90	1394.0
178.00	928.0
179.00	7281.0
180.00	5062.0
181.00	2478.0
182.10	346.0
182.90	333.0
184.00	644.0
185.00	3332.0
186.00	27576.0
187.00	8153.0
188.00	847.0
188.90	1855.0
190.00	320.0
190.90	1226.0
191.90	2461.0
193.00	2788.0
194.00	552.0
194.90	401.0
196.00	6066.0
198.00	209088.0
199.00	13784.0
200.00	1150.0
201.30	750.0
203.00	1408.0
204.00	7995.0
205.10	12278.0
206.00	55472.0
207.10	9151.0
208.00	2199.0
209.00	899.0
210.00	796.0
211.10	2447.0
214.90	845.0
216.00	1248.0
217.00	14247.0
217.90	1834.0
221.00	9081.0
223.00	3839.0
224.00	32424.0
225.00	8341.0
226.00	871.0
227.00	12907.0
228.00	1861.0
229.00	2728.0
231.00	1062.0
234.00	729.0
235.10	1062.0
235.90	639.0
236.90	1177.0
238.90	695.0
240.00	453.0
241.00	897.0
242.00	1875.0
243.10	2206.0
244.00	26320.0
245.00	3258.0
246.00	5406.0
247.00	1336.0
249.00	995.0
252.90	478.0
254.20	1404.0
255.00	151424.0
256.00	20752.0
257.00	1806.0
258.00	8937.0
259.00	1227.0
261.00	308.0
265.00	3082.0
265.80	500.0
272.00	470.0
273.00	4146.0
274.00	12700.0
275.00	72432.0
276.00	9812.0
277.00	5955.0
277.90	919.0
281.10	535.0
282.00	329.0
283.00	851.0
284.10	466.0
285.10	1088.0
292.90	1362.0
293.90	300.0
295.20	442.0
296.00	21744.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

297.00	3265.0
302.10	543.0
303.00	2567.0
304.10	595.0
310.00	339.0
314.00	1113.0
315.00	2175.0
316.00	1385.0
321.10	575.0
322.10	487.0
323.00	6939.0
324.00	1368.0
326.90	1187.0
327.90	617.0
332.00	504.0
333.00	725.0
334.00	4412.0
335.00	1605.0
341.10	852.0
346.00	1638.0
347.00	319.0
352.10	2594.0
353.00	1545.0
354.10	2363.0
355.00	475.0
365.00	10397.0
366.00	1218.0
371.00	624.0
372.00	3596.0
373.00	857.0
383.00	1143.0
390.10	403.0
391.00	451.0
402.00	1747.0
403.10	1953.0
404.00	602.0
421.10	1609.0
422.00	1364.0
423.10	11903.0
424.00	2440.0
425.10	443.0
441.10	33344.0
442.10	216704.0
443.10	40344.0
444.10	4026.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44

Scan 2465 (17.586 min): BM012813.D
DFTPP44

m/z	Abundance
73.00	342.0
207.10	1557.0
208.00	452.0
281.10	359.0

Instrument :
BNA_M
ClientSampleId :
DFTPP44