

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110824\
 Data File : BM048564.D
 Acq On : 09 Nov 2024 13:56
 Operator : RC/JU
 Sample : PB164721BL
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK721

Quant Time: Nov 09 21:50:51 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM110724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 07 22:50:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.663	152	217866	20.000	ng/ul	0.00
20) Naphthalene-d8	10.445	136	886699	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.310	164	537351	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.057	188	1082446	20.000	ng/ul	0.00
79) Chrysene-d12	21.280	240	1034179	20.000	ng/ul	0.00
88) Perylene-d12	24.191	264	1235667	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.128	96	27634	5.056	ng/uL	0.00
4) Pyridine-d5	3.540	84	377692	25.077	ng/ul	0.00
7) Phenol-d5	6.845	99	444707	26.814	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.998	67	259641	26.064	ng/ul	0.00
11) 2-Chlorophenol-d4	7.198	132	383953	27.948	ng/ul	0.00
15) 4-Methylphenol-d8	8.375	113	353034	27.072	ng/ul	0.00
21) Nitrobenzene-d5	8.816	128	178424	27.763	ng/ul	0.00
24) 2-Nitrophenol-d4	9.539	143	197019	28.822	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.081	165	331299	26.089	ng/ul	0.00
31) 4-Chloroaniline-d4	10.592	131	486156	26.953	ng/ul	0.00
46) Dimethylphthalate-d6	13.722	166	944384	27.532	ng/ul	0.00
49) Acenaphthylene-d8	14.004	160	1123363	27.777	ng/ul	0.00
54) 4-Nitrophenol-d4	14.539	143	139506	21.897	ng/ul	0.00
60) Fluorene-d10	15.304	176	830720	27.744	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.439	200	107254	19.696	ng/ul	0.00
73) Anthracene-d10	17.157	188	1280315	27.951	ng/ul	0.00
81) Pyrene-d10	19.451	212	1508761	29.070	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.991	264	1622850	28.109	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110824\
Data File : BM048564.D
Acq On : 09 Nov 2024 13:56
Operator : RC/JU
Sample : PB164721BL
Misc :
ALS Vial : 46 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK721

Quant Time: Nov 09 21:50:51 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM110724.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Nov 07 22:50:20 2024
Response via : Initial Calibration

