

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110824\
 Data File : BM048573.D
 Acq On : 09 Nov 2024 19:49
 Operator : RC/JU
 Sample : PB164612BS
 Misc :
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS612

Quant Time: Nov 09 21:52:28 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM110724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 07 22:50:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.657	152	197545	20.000	ng/u1	0.00
20) Naphthalene-d8	10.445	136	809618	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.310	164	495238	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.057	188	1014575	20.000	ng/u1	0.00
79) Chrysene-d12	21.280	240	975699	20.000	ng/u1	0.00
88) Perylene-d12	24.191	264	1113392	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.128	96	25840	5.215	ng/uL	0.00
4) Pyridine-d5	3.540	84	366464	26.835	ng/u1	0.00
7) Phenol-d5	6.845	99	450220	29.939	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.998	67	252678	27.975	ng/u1	0.00
11) 2-Chlorophenol-d4	7.198	132	383494	30.786	ng/u1	0.00
15) 4-Methylphenol-d8	8.381	113	349101	29.524	ng/u1	0.00
21) Nitrobenzene-d5	8.816	128	179307	30.557	ng/u1	0.00
24) 2-Nitrophenol-d4	9.539	143	194984	31.240	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.075	165	354557	30.579	ng/u1	0.00
31) 4-Chloroaniline-d4	10.592	131	403839	24.521	ng/u1	0.00
46) Dimethylphthalate-d6	13.727	166	933418	29.526	ng/u1	0.00
49) Acenaphthylene-d8	13.998	160	1112645	29.852	ng/u1	0.00
54) 4-Nitrophenol-d4	14.533	143	188471	32.098	ng/u1	0.00
60) Fluorene-d10	15.304	176	823489	29.841	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.439	200	154135	30.199	ng/u1	0.00
73) Anthracene-d10	17.151	188	1272305	29.634	ng/u1	0.00
81) Pyrene-d10	19.450	212	1530197	31.250	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.991	264	1599584	30.749	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.157	88	52849	9.718	ng/uL	97
5) Pyridine	3.557	79	375994	26.821	ng/u1	98
6) Benzaldehyde	6.810	77	30503	4.105	ng/u1	99
8) Phenol	6.869	94	461448	29.247	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.092	93	342813	27.648	ng/u1	98
12) 2-Chlorophenol	7.228	128	384671	29.808	ng/u1	99
13) 2-Methylphenol	8.110	108	341601	29.248	ng/u1	96
14) 2,2'-oxybis(1-Chloropr...	8.181	45	495647	27.183	ng/u1	97
16) Acetophenone	8.486	105	523927	28.426	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.469	70	251063	28.985	ng/u1	98
18) 4-Methylphenol	8.439	108	373909	29.238	ng/u1	98
19) Hexachloroethane	8.728	117	149723	27.454	ng/u1	97
22) Nitrobenzene	8.857	77	385354	28.963	ng/u1	96
23) Isophorone	9.386	82	711156	28.896	ng/u1	99
25) 2-Nitrophenol	9.569	139	207652	30.528	ng/u1	97
26) 2,4-Dimethylphenol	9.639	107	348134	27.401	ng/u1	97
27) Bis(2-Chloroethoxy)met...	9.869	93	446787	29.084	ng/u1	99
29) 2,4-Dichlorophenol	10.104	162	352959	30.010	ng/u1	96
30) Naphthalene	10.498	128	1149342	28.047	ng/u1	99
32) 4-Chloroaniline	10.616	127	250074	16.104	ng/u1	97
33) Hexachlorobutadiene	10.780	225	219732	27.607	ng/u1	96
34) Caprolactam	11.398	113	109756	29.609	ng/u1	88
35) 4-Chloro-3-methylphenol	11.751	107	348670	30.637	ng/u1	97
36) 2-Methylnaphthalene	12.116	142	750274	28.615	ng/u1	98


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110824\
 Data File : BM048573.D
 Acq On : 09 Nov 2024 19:49
 Operator : RC/JU
 Sample : PB164612BS
 Misc :
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS612

Quant Time: Nov 09 21:52:28 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM110724.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Nov 07 22:50:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.339	142	743633	27.636	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.492	216	402560	28.366	ng/ul	99
40) Hexachlorocyclopentadiene	12.469	237	161913	22.321	ng/ul	100
41) 2,4,6-Trichlorophenol	12.739	196	269935	30.113	ng/ul	100
42) 2,4,5-Trichlorophenol	12.816	196	291327	30.517	ng/ul	97
43) 1,1'-Biphenyl	13.139	154	965586	28.515	ng/ul	99
44) 2-Chloronaphthalene	13.180	162	766299	28.678	ng/ul	100
45) 2-Nitroaniline	13.392	65	211622	31.536	ng/ul	95
47) Dimethylphthalate	13.774	163	917494	28.892	ng/ul	100
48) 2,6-Dinitrotoluene	13.892	165	192272	31.760	ng/ul	97
50) Acenaphthylene	14.027	152	1180314	28.729	ng/ul	99
51) 3-Nitroaniline	14.227	138	195724	29.771	ng/ul	96
52) Acenaphthene	14.374	153	828499	28.644	ng/ul	99
53) 2,4-Dinitrophenol	14.439	184	94549	29.047	ng/ul	93
55) 4-Nitrophenol	14.545	109	131437	30.963	ng/ul	98
56) Dibenzofuran	14.710	168	1137776	28.515	ng/ul	99
57) 2,4-Dinitrotoluene	14.686	165	282439	32.097	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	14.945	232	243397	30.651	ng/ul	97
59) Diethylphthalate	15.139	149	924613	29.552	ng/ul	99
61) Fluorene	15.363	166	904272	28.625	ng/ul	98
62) 4-Chlorophenyl-phenyle...	15.357	204	445633	28.633	ng/ul	99
63) 4-Nitroaniline	15.392	138	229306	35.421	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.451	198	161860	29.142	ng/ul	96
67) N-Nitrosodiphenylamine	15.574	169	762027	28.497	ng/ul	99
68) 4-Bromophenyl-phenylether	16.251	248	277622	29.014	ng/ul	99
69) Hexachlorobenzene	16.362	284	329130	28.501	ng/ul	98
70) Atrazine	16.527	200	270935	27.839	ng/ul	98
71) Pentachlorophenol	16.715	266	224156	30.541	ng/ul	96
72) Phenanthrene	17.098	178	1437820	28.574	ng/ul	99
74) Anthracene	17.186	178	1448502	28.684	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.104	216	395604	27.333	ng/ul	99
76) Pentachlorobenzene	14.633	250	379324	26.899	ng/ul	97
77) Carbazole	17.462	167	1378390	30.303	ng/ul	99
78) Di-n-butylphthalate	18.027	149	1631280	30.504	ng/ul	99
80) Fluoranthene	19.115	202	1719174	30.259	ng/ul	99
82) Pyrene	19.480	202	1840579	29.877	ng/ul	98
83) Butylbenzylphthalate	20.380	149	773743	32.697	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.192	252	520258	29.347	ng/ul	99
85) Benzo(a)anthracene	21.262	228	1834598	29.828	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.186	149	1111856	32.208	ng/ul	99
87) Chrysene	21.321	228	1729880	29.268	ng/ul	99
89) Di-n-octyl phthalate	22.286	149	1988982	31.689	ng/ul	100
90) Benzo(b)fluoranthene	23.274	252	1880744	30.686	ng/ul	98
91) Benzo(k)fluoranthene	23.333	252	1835566	30.037	ng/ul	100
93) Benzo(a)pyrene	24.056	252	1726002	29.952	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	27.462	276	2232369	29.810	ng/ul	98
95) Dibenzo(a,h)anthracene	27.509	278	1820514	29.763	ng/ul	99
96) Benzo(g,h,i)perylene	28.479	276	1739996	29.633	ng/ul	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 9.718 ng/uL
 RT: 3.157 min Scan# 29
 Delta R.T. -0.006 min
 Lab File: BM048573.D
 Acq: 09 Nov 2024 19:49




Instrument :
 BNA_M
 ClientSampleId :
 SLCS612


Tgt Ion: 88 Resp: 52849
 Ion Ratio Lower Upper
 88 100
 43 28.5 22.1 33.1
 58 72.8 60.7 91.1



#5
 Pyridine
 Concen: 26.821 ng/uL
 RT: 3.557 min Scan# 97
 Delta R.T. -0.006 min
 Lab File: BM048573.D
 Acq: 09 Nov 2024 19:49

Tgt Ion: 79 Resp: 375994
 Ion Ratio Lower Upper
 79 100
 52 76.6 62.6 94.0
 51 34.1 27.8 41.8









#6
Benzaldehyde
Concen: 4.105 ng/ul
RT: 6.810 min Scan# 610
Delta R.T. 0.000 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49




Instrument :
BNA_M
ClientSampleId :
SLCS612


Tgt Ion: 77 Resp: 30503
Ion Ratio Lower Upper
77 100
105 95.3 78.0 117.0
106 95.7 76.6 115.0



#8
Phenol
Concen: 29.247 ng/ul
RT: 6.869 min Scan# 660
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49

Tgt Ion: 94 Resp: 461448
Ion Ratio Lower Upper
94 100
65 26.7 20.8 31.2
66 33.0 26.1 39.1









#10
Bis(2-Chloroethyl)ether
Concen: 27.648 ng/ul
RT: 7.092 min Scan# 698
Delta R.T. 0.000 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49




Instrument :
BNA_M
ClientSampleId :
SLCS612


Tgt Ion: 93 Resp: 342813
Ion Ratio Lower Upper
93 100
63 73.7 59.6 89.4
95 33.1 25.0 37.6



#12
2-Chlorophenol
Concen: 29.808 ng/ul
RT: 7.228 min Scan# 721
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49

Tgt Ion:128 Resp: 384671
Ion Ratio Lower Upper
128 100
64 43.0 35.2 52.8
130 32.1 25.8 38.6





#13

2-Methylphenol

Concen: 29.248 ng/ul

RT: 8.110 min Scan# 81

Delta R.T. -0.006 min

Lab File: BM048573.D

Acq: 09 Nov 2024 19:49

Instrument :

BNA_M

ClientSampleId :





SLCS612

Tgt Ion:108 Resp: 341601

Ion Ratio Lower Upper

108 100

107 86.0 71.5 107.3



#14

2,2'-oxybis(1-Chloropropane)

Concen: 27.183 ng/ul

RT: 8.181 min Scan# 883

Delta R.T. -0.006 min

Lab File: BM048573.D

Acq: 09 Nov 2024 19:49




Tgt Ion: 45 Resp: 495647


Ion Ratio Lower Upper

45 100

77 14.2 10.6 16.0

79 10.6 7.4 11.2









#16
Acetophenone
Concen: 28.426 ng/ul
RT: 8.486 min Scan# 91
Delta R.T. 0.000 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49




Instrument :
BNA_M
ClientSampleId :
SLCS612


Tgt Ion:105 Resp: 523927
Ion Ratio Lower Upper
105 100
77 73.1 59.7 89.5
51 24.3 20.6 31.0



#17
N-Nitroso-di-n-propylamine
Concen: 28.985 ng/ul
RT: 8.469 min Scan# 932
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49

Tgt Ion: 70 Resp: 251063
Ion Ratio Lower Upper
70 100
42 57.0 47.0 70.4
101 10.2 8.5 12.7
130 23.1 18.4 27.6









#18
4-Methylphenol
Concen: 29.238 ng/ul
RT: 8.439 min Scan# 911
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49




Instrument :
BNA_M
ClientSampleId :
SLCS612


Tgt Ion:108 Resp: 373909
Ion Ratio Lower Upper
108 100
107 113.5 88.8 133.2



#19
Hexachloroethane
Concen: 27.454 ng/ul
RT: 8.728 min Scan# 976
Delta R.T. -0.012 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49

Tgt Ion:117 Resp: 149723
Ion Ratio Lower Upper
117 100
201 100.4 83.8 125.8
199 64.7 52.1 78.1









#22
Nitrobenzene
Concen: 28.963 ng/ul
RT: 8.857 min Scan# 998
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49




Instrument :
BNA_M
ClientSampleId :
SLCS612


Tgt Ion: 77 Resp: 385354
Ion Ratio Lower Upper
77 100
123 48.6 36.6 55.0
65 12.5 9.5 14.3



#23
Isophorone
Concen: 28.896 ng/ul
RT: 9.386 min Scan# 1088
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49

Tgt Ion: 82 Resp: 711156
Ion Ratio Lower Upper
82 100
95 6.3 5.0 7.6
138 18.3 14.2 21.4









#25
2-Nitrophenol
Concen: 30.528 ng/ul
RT: 9.569 min Scan# 1119
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49




Instrument :
BNA_M
ClientSampleId :
SLCS612


Tgt Ion:139 Resp: 207652
Ion Ratio Lower Upper
139 100
65 44.8 37.7 56.5
109 26.4 22.3 33.5



#26
2,4-Dimethylphenol
Concen: 27.401 ng/ul
RT: 9.639 min Scan# 1131
Delta R.T. 0.000 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49

Tgt Ion:107 Resp: 348134
Ion Ratio Lower Upper
107 100
121 51.9 42.9 64.3
122 91.0 69.8 104.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 29.084 ng/ul
 RT: 9.869 min Scan# 1170
 Delta R.T. -0.006 min
 Lab File: BM048573.D
 Acq: 09 Nov 2024 19:49

Instrument :
 BNA_M
 ClientSampleId :
 SLCS612

Tgt Ion: 93 Resp: 446787




Ion	Ratio	Lower	Upper
93	100		
95	32.1	25.3	37.9
123	12.0	9.5	14.3




#29
 2,4-Dichlorophenol
 Concen: 30.010 ng/ul
 RT: 10.104 min Scan# 1210
 Delta R.T. -0.006 min
 Lab File: BM048573.D
 Acq: 09 Nov 2024 19:49

Tgt Ion: 162 Resp: 352959

Ion	Ratio	Lower	Upper
162	100		
164	64.7	54.6	82.0
98	33.4	28.2	42.2









#30
Naphthalene
Concen: 28.047 ng/ul
RT: 10.498 min Scan# 1277
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49




Instrument :
BNA_M
ClientSampleId :
SLCS612


Tgt Ion:128 Resp: 1149342
Ion Ratio Lower Upper
128 100
129 10.7 8.5 12.7
127 12.9 10.0 15.0



#32
4-Chloroaniline
Concen: 16.104 ng/ul
RT: 10.616 min Scan# 1297
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49

Tgt Ion:127 Resp: 250074
Ion Ratio Lower Upper
127 100
129 31.0 26.2 39.4





#33

Hexachlorobutadiene

Concen: 27.607 ng/ul

RT: 10.780 min Scan# 11

Delta R.T. -0.006 min

Lab File: BM048573.D

Acq: 09 Nov 2024 19:49

Instrument :

BNA_M

ClientSampleId :

SLCS612





Tgt Ion:225 Resp: 219732

Ion Ratio Lower Upper

225 100

223 65.4 48.5 72.7

227 66.0 51.5 77.3



#34

Caprolactam

Concen: 29.609 ng/ul

RT: 11.398 min Scan# 1430

Delta R.T. 0.000 min

Lab File: BM048573.D

Acq: 09 Nov 2024 19:49




Tgt Ion:113 Resp: 109756


Ion Ratio Lower Upper

113 100

55 159.2 138.5 207.7

56 114.0 104.1 156.1









#35
4-Chloro-3-methylphenol
Concen: 30.637 ng/ul
RT: 11.751 min Scan# 1490
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49




Instrument :
BNA_M
ClientSampleId :
SLCS612


Tgt Ion:107 Resp: 348670
Ion Ratio Lower Upper
107 100
144 26.3 20.6 31.0
142 80.6 66.7 100.1



#36
2-Methylnaphthalene
Concen: 28.615 ng/ul
RT: 12.116 min Scan# 1552
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49

Tgt Ion:142 Resp: 750274
Ion Ratio Lower Upper
142 100
141 87.5 71.6 107.4





#37

1-Methylnaphthalene

Concen: 27.636 ng/ul

RT: 12.339 min Scan# 1590

Delta R.T. -0.006 min

Lab File: BM048573.D

Acq: 09 Nov 2024 19:49

Instrument :

BNA_M

ClientSampleId :

SLCS612





Tgt Ion:142 Resp: 743633

Ion Ratio Lower Upper

142 100

141 91.6 73.6 110.4

116 3.5 2.7 4.1



#39

1,2,4,5-Tetrachlorobenzene

Concen: 28.366 ng/ul

RT: 12.492 min Scan# 1616

Delta R.T. -0.006 min

Lab File: BM048573.D

Acq: 09 Nov 2024 19:49

Tgt Ion:216 Resp: 402560




Ion Ratio Lower Upper


216 100

214 79.9 62.6 93.8

179 20.6 16.5 24.7

108 16.1 12.9 19.3









#40
Hexachlorocyclopentadiene
Concen: 22.321 ng/ul
RT: 12.469 min Scan# 1612
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49




Instrument :
BNA_M
ClientSampleId :
SLCS612


Tgt Ion:237 Resp: 161913
Ion Ratio Lower Upper
237 100
235 62.8 50.0 75.0
272 13.0 10.6 16.0



#41
2,4,6-Trichlorophenol
Concen: 30.113 ng/ul
RT: 12.739 min Scan# 1658
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49

Tgt Ion:196 Resp: 269935
Ion Ratio Lower Upper
196 100
198 95.7 76.6 114.8
200 31.4 25.8 38.6









#42
2,4,5-Trichlorophenol
Concen: 30.517 ng/ul
RT: 12.816 min Scan# 1671
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49




Instrument :
BNA_M
ClientSampleId :
SLCS612


Tgt Ion:196 Resp: 291327
Ion Ratio Lower Upper
196 100
198 95.0 78.7 118.1
200 31.1 24.6 37.0



#43
1,1'-Biphenyl
Concen: 28.515 ng/ul
RT: 13.139 min Scan# 1726
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49

Tgt Ion:154 Resp: 965586
Ion Ratio Lower Upper
154 100
153 39.5 31.4 47.2
76 12.0 9.9 14.9









#44
2-Chloronaphthalene
Concen: 28.678 ng/ul
RT: 13.180 min Scan# 1733
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49




Instrument :
BNA_M
ClientSampleId :
SLCS612


Tgt Ion:162 Resp: 766299
Ion Ratio Lower Upper
162 100
164 32.4 25.8 38.6
127 37.1 29.5 44.3



#45
2-Nitroaniline
Concen: 31.536 ng/ul
RT: 13.392 min Scan# 1769
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49

Tgt Ion: 65 Resp: 211622
Ion Ratio Lower Upper
65 100
92 72.0 56.2 84.2
138 119.7 89.6 134.4









#47
Dimethylphthalate
Concen: 28.892 ng/ul
RT: 13.774 min Scan# 1834
Delta R.T. 0.000 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49




Instrument :
BNA_M
ClientSampleId :
SLCS612


Tgt Ion:163 Resp: 917494
Ion Ratio Lower Upper
163 100
194 4.0 3.0 4.4
164 9.8 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 31.760 ng/ul
RT: 13.892 min Scan# 1854
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49

Tgt Ion:165 Resp: 192272
Ion Ratio Lower Upper
165 100
89 53.7 45.4 68.0
121 20.9 17.1 25.7









#50
Acenaphthylene
Concen: 28.729 ng/ul
RT: 14.027 min Scan# 1877
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49




Instrument :
BNA_M
ClientSampleId :
SLCS612


Tgt Ion:152 Resp: 1180314
Ion Ratio Lower Upper
152 100
151 20.1 15.8 23.6
153 13.1 10.5 15.7



#51
3-Nitroaniline
Concen: 29.771 ng/ul
RT: 14.227 min Scan# 1911
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49

Tgt Ion:138 Resp: 195724
Ion Ratio Lower Upper
138 100
108 10.1 7.8 11.8
92 109.5 91.0 136.4









#52
Acenaphthene
Concen: 28.644 ng/ul
RT: 14.374 min Scan# 1936
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49




Instrument :
BNA_M
ClientSampleId :
SLCS612


Tgt Ion:153 Resp: 828499
Ion Ratio Lower Upper
153 100
152 46.9 38.3 57.5
154 88.3 71.1 106.7



#53
2,4-Dinitrophenol
Concen: 29.047 ng/ul
RT: 14.439 min Scan# 1947
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49

Tgt Ion:184 Resp: 94549
Ion Ratio Lower Upper
184 100
63 59.2 52.5 78.7
154 59.6 51.4 77.0









#55
4-Nitrophenol
Concen: 30.963 ng/ul
RT: 14.545 min Scan# 1965
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49




Instrument :
BNA_M
ClientSampleId :
SLCS612


Tgt Ion:109 Resp: 131437
Ion Ratio Lower Upper
109 100
139 149.4 115.9 173.9
65 132.8 107.0 160.4



#56
Dibenzofuran
Concen: 28.515 ng/ul
RT: 14.710 min Scan# 1993
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49

Tgt Ion:168 Resp: 1137776
Ion Ratio Lower Upper
168 100
139 36.5 29.7 44.5









#57
2,4-Dinitrotoluene
Concen: 32.097 ng/ul
RT: 14.686 min Scan# 1989
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49




Instrument :
BNA_M
ClientSampleId :
SLCS612


Tgt Ion:165 Resp: 282439
Ion Ratio Lower Upper
165 100
63 40.6 34.6 51.8
182 2.9 2.4 3.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 30.651 ng/ul
RT: 14.945 min Scan# 2033
Delta R.T. 0.000 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49

Tgt Ion:232 Resp: 243397
Ion Ratio Lower Upper
232 100
131 42.9 36.3 54.5
130 2.1 1.8 2.8
166 29.5 24.2 36.2









#59
Diethylphthalate
Concen: 29.552 ng/ul
RT: 15.139 min Scan# 2066
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49




Instrument :
BNA_M
ClientSampleId :
SLCS612


Tgt Ion:149 Resp: 924613
Ion Ratio Lower Upper
149 100
177 22.1 17.8 26.8
150 12.2 9.6 14.4



#61
Fluorene
Concen: 28.625 ng/ul
RT: 15.363 min Scan# 2104
Delta R.T. 0.000 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49

Tgt Ion:166 Resp: 904272
Ion Ratio Lower Upper
166 100
165 96.7 78.7 118.1
167 13.3 11.0 16.6









#62
4-Chlorophenyl-phenylether
Concen: 28.633 ng/ul
RT: 15.357 min Scan# 2103
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49




Instrument :
BNA_M
ClientSampleId :
SLCS612


Tgt Ion:204 Resp: 445633
Ion Ratio Lower Upper
204 100
206 32.7 26.0 39.0
141 55.4 45.4 68.2



#63
4-Nitroaniline
Concen: 35.421 ng/ul
RT: 15.392 min Scan# 2109
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49

Tgt Ion:138 Resp: 229306
Ion Ratio Lower Upper
138 100
92 47.4 39.7 59.5
108 69.8 58.5 87.7









#66
 4,6-Dinitro-2-methylphenol
 Concen: 29.142 ng/ul
 RT: 15.451 min Scan# 2119
 Delta R.T. -0.006 min
 Lab File: BM048573.D
 Acq: 09 Nov 2024 19:49




Instrument :
 BNA_M
 ClientSampleId :
 SLCS612


Tgt Ion:198 Resp: 161860
 Ion Ratio Lower Upper
 198 100
 182 5.1 5.0 7.6
 77 23.4 20.1 30.1



#67
 N-Nitrosodiphenylamine
 Concen: 28.497 ng/ul
 RT: 15.574 min Scan# 2140
 Delta R.T. 0.000 min
 Lab File: BM048573.D
 Acq: 09 Nov 2024 19:49

Tgt Ion:169 Resp: 762027
 Ion Ratio Lower Upper
 169 100
 168 66.2 52.6 79.0
 167 35.9 29.6 44.4









#68
4-Bromophenyl-phenylether
Concen: 29.014 ng/ul
RT: 16.251 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49




Instrument :
BNA_M
ClientSampleId :
SLCS612


Tgt Ion:248 Resp: 277622
Ion Ratio Lower Upper
248 100
250 98.4 79.4 119.2
141 67.8 53.8 80.6



#69
Hexachlorobenzene
Concen: 28.501 ng/ul
RT: 16.362 min Scan# 2274
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49

Tgt Ion:284 Resp: 329130
Ion Ratio Lower Upper
284 100
142 35.3 27.5 41.3
249 30.2 23.2 34.8





#70

Atrazine

Concen: 27.839 ng/ul

RT: 16.527 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM048573.D

Acq: 09 Nov 2024 19:49

Instrument :

BNA_M

ClientSampleId :

SLCS612





Tgt Ion:200 Resp: 270935

Ion Ratio Lower Upper

200 100

173 27.3 20.2 30.2

215 47.8 38.2 57.2



#71

Pentachlorophenol

Concen: 30.541 ng/ul

RT: 16.715 min Scan# 2334

Delta R.T. 0.000 min

Lab File: BM048573.D

Acq: 09 Nov 2024 19:49




Tgt Ion:266 Resp: 224156


Ion Ratio Lower Upper

266 100

264 61.4 51.1 76.7

268 63.3 53.2 79.8





#72

Phenanthrene

Concen: 28.574 ng/ul

RT: 17.098 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM048573.D

Acq: 09 Nov 2024 19:49

Instrument :

BNA_M

ClientSampleId :

SLCS612





Tgt Ion:178 Resp: 1437820

Ion Ratio Lower Upper

178 100

179 15.6 12.0 18.0

176 19.1 15.1 22.7



#74

Anthracene

Concen: 28.684 ng/ul

RT: 17.186 min Scan# 2414

Delta R.T. -0.006 min

Lab File: BM048573.D

Acq: 09 Nov 2024 19:49




Tgt Ion:178 Resp: 1448502


Ion Ratio Lower Upper

178 100

179 15.5 12.6 19.0

176 19.0 15.0 22.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 27.333 ng/uL
RT: 13.104 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49




Instrument :
BNA_M
ClientSampleId :
SLCS612


Tgt Ion:216 Resp: 395604
Ion Ratio Lower Upper
216 100
214 78.8 61.9 92.9
218 48.1 38.6 58.0



#76
Pentachlorobenzene
Concen: 26.899 ng/uL
RT: 14.633 min Scan# 1980
Delta R.T. 0.000 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49

Tgt Ion:250 Resp: 379324
Ion Ratio Lower Upper
250 100
248 63.0 49.7 74.5
252 66.2 50.2 75.2









#77
 Carbazole
 Concen: 30.303 ng/ul
 RT: 17.462 min Scan# 2461
 Delta R.T. -0.006 min
 Lab File: BM048573.D
 Acq: 09 Nov 2024 19:49




Instrument :
 BNA_M
 ClientSampleId :
 SLCS612


Tgt Ion:167 Resp: 1378390
 Ion Ratio Lower Upper
 167 100
 166 21.5 16.4 24.6
 139 12.8 10.3 15.5



#78
 Di-n-butylphthalate
 Concen: 30.504 ng/ul
 RT: 18.027 min Scan# 2557
 Delta R.T. 0.000 min
 Lab File: BM048573.D
 Acq: 09 Nov 2024 19:49

Tgt Ion:149 Resp: 1631280
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.4 11.2
 104 4.5 4.0 6.0









#80
Fluoranthene
Concen: 30.259 ng/ul
RT: 19.115 min Scan# 2742
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49




Instrument :
BNA_M
ClientSampleId :
SLCS612


Tgt Ion:202 Resp: 1719174
Ion Ratio Lower Upper
202 100
101 10.9 8.5 12.7
100 8.3 7.0 10.4



#82
Pyrene
Concen: 29.877 ng/ul
RT: 19.480 min Scan# 2804
Delta R.T. 0.000 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49

Tgt Ion:202 Resp: 1840579
Ion Ratio Lower Upper
202 100
101 12.6 10.6 15.8
100 10.4 8.9 13.3





#83

Butylbenzylphthalate

Concen: 32.697 ng/ul

RT: 20.380 min Scan# 2957

Delta R.T. -0.006 min

Lab File: BM048573.D

Acq: 09 Nov 2024 19:49

Instrument :

BNA_M

ClientSampleId :

SLCS612





Tgt Ion:149 Resp: 773743

Ion Ratio Lower Upper

149 100

91 68.1 53.1 79.7

206 20.1 16.6 25.0



#84

3,3'-Dichlorobenzidine

Concen: 29.347 ng/ul

RT: 21.192 min Scan# 3095

Delta R.T. -0.006 min

Lab File: BM048573.D

Acq: 09 Nov 2024 19:49




Tgt Ion:252 Resp: 520258


Ion Ratio Lower Upper

252 100

254 64.1 52.1 78.1

126 10.9 8.3 12.5





#85

Benzo(a)anthracene

Concen: 29.828 ng/ul

RT: 21.262 min Scan# 3107

Delta R.T. 0.000 min

Lab File: BM048573.D

Acq: 09 Nov 2024 19:49

Instrument :

BNA_M

ClientSampleId :

SLCS612





Tgt Ion:228 Resp: 1834598

Ion Ratio Lower Upper

228 100

229 19.6 15.7 23.5

226 26.9 21.7 32.5



#86

Bis(2-ethylhexyl)phthalate

Concen: 32.208 ng/ul

RT: 21.186 min Scan# 3094

Delta R.T. -0.006 min

Lab File: BM048573.D

Acq: 09 Nov 2024 19:49




Tgt Ion:149 Resp: 1111856


Ion Ratio Lower Upper

149 100

167 28.3 22.0 33.0

279 3.9 3.1 4.7









#87
Chrysene
Concen: 29.268 ng/ul
RT: 21.321 min Scan# 3117
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49




Instrument :
BNA_M
ClientSampleId :
SLCS612


Tgt Ion:228 Resp: 1729880
Ion Ratio Lower Upper
228 100
226 29.7 23.4 35.2
229 19.3 15.6 23.4



#89
Di-n-octyl phthalate
Concen: 31.689 ng/ul
RT: 22.286 min Scan# 3281
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49

Tgt Ion:149 Resp: 1988982









#90
Benzo(b)fluoranthene
Concen: 30.686 ng/u1
RT: 23.274 min Scan# 3448
Delta R.T. 0.000 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49




Instrument :
BNA_M
ClientSampleId :
SLCS612


Tgt Ion:252 Resp: 1880744
Ion Ratio Lower Upper
252 100
253 21.6 17.7 26.5
125 9.9 8.7 13.1



#91
Benzo(k)fluoranthene
Concen: 30.037 ng/u1
RT: 23.333 min Scan# 3459
Delta R.T. -0.006 min
Lab File: BM048573.D
Acq: 09 Nov 2024 19:49

Tgt Ion:252 Resp: 1835566
Ion Ratio Lower Upper
252 100
253 21.6 17.4 26.0
125 10.2 8.0 12.0





#93

Benzo(a)pyrene

Concen: 29.952 ng/ul

RT: 24.056 min Scan# 3581

Delta R.T. 0.000 min

Lab File: BM048573.D

Acq: 09 Nov 2024 19:49

Instrument :

BNA_M

ClientSampleId :

SLCS612





Tgt Ion:252 Resp: 1726002

Ion Ratio Lower Upper

252 100

253 21.9 17.8 26.8

125 11.0 9.1 13.7



#94

Indeno(1,2,3-cd)pyrene

Concen: 29.810 ng/ul

RT: 27.462 min Scan# 4161

Delta R.T. -0.006 min

Lab File: BM048573.D

Acq: 09 Nov 2024 19:49




Tgt Ion:276 Resp: 2232369


Ion Ratio Lower Upper

276 100

138 21.5 16.2 24.2

277 24.0 19.5 29.3





#95

Dibenzo(a,h)anthracene

Concen: 29.763 ng/ul

RT: 27.509 min Scan# 4169

Delta R.T. -0.006 min

Lab File: BM048573.D

Acq: 09 Nov 2024 19:49

Instrument :

BNA_M

ClientSampleId :

SLCS612





Tgt Ion:278 Resp: 1820514

Ion Ratio Lower Upper

278 100

139 16.3 12.6 18.8

279 24.3 19.0 28.6



#96

Benzo(g,h,i)perylene

Concen: 29.633 ng/ul

RT: 28.479 min Scan# 4334

Delta R.T. -0.006 min

Lab File: BM048573.D

Acq: 09 Nov 2024 19:49

Tgt Ion:276 Resp: 1739996

Ion Ratio Lower Upper

276 100

138 21.3 17.6 26.4

277 23.8 18.6 28.0

