

Method Path : Z:\HPCHEM1\BNA_M\METHODS\
 Method File : SOM02.2-EPA-1,4-DIO SIM-BM110915.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Nov 10 02:32:24 2015
 Response Via : Initial Calibration

Calibration Files

0.5 =BM003070.D 1 =BM003071.D 2 =BM003072.D
 4 =BM003073.D 8 =BM003074.D

Compound		0.5	1	2	4	8	Avg	%RSD

1) I	1,4-Dichlorobenzene-d	-----ISTD-----						
2) SURR	1,4-Dioxane-d8	0.415	0.414	0.375	0.369	0.380	0.391	5.64
3)	1,4-Dioxane	0.506	0.499	0.446	0.434	0.444	0.466	7.30
4) I	Naphthalene-d8	-----ISTD-----						
5) I	Acenaphthene-d10	-----ISTD-----						
6) I	Phenanthrene-d10	-----ISTD-----						
7) I	Chrysene-d12	-----ISTD-----						
8) I	Perylene-d12	-----ISTD-----						

(#) = Out of Range