

Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM110920\
 Data File : BM027758.D
 Acq On : 09 Nov 2020 15:31
 Operator : JU/CG
 Sample : L4661-05
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 F4R04

Manual Integrations
 APPROVED

mohammad
 11/10/2020 12:01:23 PM

Quant Time: Nov 10 06:59:13 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM02.4-EPA-SIM-BM110920.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 10 06:30:49 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.39	152	819	0.40	ng/ul	0.00
2) Naphthalene-d8	11.23	136	3097	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.00	164	2065	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.72	188	3888m	0.40	ng/ul	0.00
16) Chrysene-d12	21.82	240	3484	0.40	ng/ul	0.00
20) Perylene-d12	24.26	264	3207	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.79	152	1914	0.34	ng/ul	0.00
14) Fluoranthene-d10	19.70	212	4527	0.39	ng/ul	0.00
Target Compounds						
8) Acenaphthene	15.07	153	846	0.109	ng/ul#	Ovalue 24

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA M\DATA\BM110920\
Data File : BM027758.D
Acq On : 09 Nov 2020 15:31
Operator : JU/CG
Sample : L4661-05
Misc :
ALS Vial : 11 Sample Multiplier: 1




Instrument :
BNA_M
ClientSampleId :
F4R04

Quant Time: Nov 10 06:59:13 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SOM02.4-EPA-SIM-BM110920.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 10 06:30:49 2020
Response via : Initial Calibration

Manual Integrations
APPROVED

mohammad
11/10/2020 12:01:23 PM





#8
 Acenaphthene
 Concen: 0.109 ng/ul
 RT: 15.07 min Scan# 3184
 Delta R.T. -0.00 min
 Lab File: BM027758.D
 Acq: 09 Nov 2020 15:31

Instrument :
 BNA_M
 ClientSampleId :
 F4R04

Tot Ion:	153	Resp:	846
Ion Ratio	Lower	Upper	
153	100		
152	8.1	46.7	70.1#
154	16.5	76.6	114.8#

Manual Integrations
 APPROVED

mohammad
 11/10/2020 12:01:23 PM

