

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110921\
 Data File : BM032976.D
 Acq On : 11 Nov 2021 05:57
 Operator : CG/JU
 Sample : PB140354BL
 Misc :
 ALS Vial : 72 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK354

Quant Time: Nov 11 09:36:39 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM110921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 09 13:31:56 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.462	152	2412	0.400	ng/ul	0.00
4) Naphthalene-d8	10.235	136	8788	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.117	164	5364	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.858	188	10881	0.400	ng/ul	0.00
17) Chrysene-d12	21.049	240	8149	0.400	ng/ul	0.00
23) Perylene-d12	23.154	264	8039	0.400	ng/ul	#-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.857	96	15696	5.096	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	11.836	152	3755	0.301	ng/ul	0.00
18) Fluoranthene-d10	18.888	212	9020	0.365	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110921\
Data File : BM032976.D
Acq On : 11 Nov 2021 05:57
Operator : CG/JU
Sample : PB140354BL
Misc :
ALS Vial : 72 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK354

Quant Time: Nov 11 09:36:39 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM110921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 09 13:31:56 2021
Response via : Initial Calibration

