

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110921\
 Data File : BM032992.D
 Acq On : 11 Nov 2021 16:35
 Operator : CG/JU
 Sample : M4615-02DL 5X
 Misc :
 ALS Vial : 88 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 COV01DL

Quant Time: Nov 11 17:09:50 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM110921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 11 13:40:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.459	152	3011	0.400	ng/ul	# 0.00
4) Naphthalene-d8	10.230	136	9933	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.114	164	5661	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.858	188	10420	0.400	ng/ul	0.00
17) Chrysene-d12	21.062	240	6010	0.400	ng/ul	# 0.01
23) Perylene-d12	23.163	264	6274	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.864	96	1466	0.381	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.831	152	560	0.040	ng/ul	0.00
18) Fluoranthene-d10	18.892	212	905	0.050	ng/ul	0.00
Target Compounds						
						Qvalue
5) Naphthalene	10.280	128	191645	6.644	ng/ul	97
7) 2-Methylnaphthalene	11.903	142	10533	0.527	ng/ul	98
8) 1-Methylnaphthalene	12.128	142	5667	0.290	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110921\
Data File : BM032992.D
Acq On : 11 Nov 2021 16:35
Operator : CG/JU
Sample : M4615-02DL 5X
Misc :
ALS Vial : 88 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0V01DL

Quant Time: Nov 11 17:09:50 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM110921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 11 13:40:18 2021
Response via : Initial Calibration

