

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110921\
 Data File : BM033012.D
 Acq On : 12 Nov 2021 05:16
 Operator : CG/JU
 Sample : PB140387BS
 Misc :
 ALS Vial : 108 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS387

Quant Time: Nov 12 05:50:44 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM110921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 11 13:40:18 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.459	152	1501	0.400	ng/ul	0.00
4) Naphthalene-d8	10.230	136	6223	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.114	164	3884	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.855	188	7993	0.400	ng/ul	0.00
17) Chrysene-d12	21.045	240	6465	0.400	ng/ul	0.00
23) Perylene-d12	23.149	264	5412	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.860	96	1417	0.739	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.831	152	3153	0.357	ng/ul	0.00
18) Fluoranthene-d10	18.889	212	7116	0.363	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	2.892	88	3797	1.929	ng/ul#	89
5) Naphthalene	10.280	128	6414	0.355	ng/ul	99
7) 2-Methylnaphthalene	11.903	142	4248	0.339	ng/ul	99
8) 1-Methylnaphthalene	12.128	142	4052	0.331	ng/ul	99
10) Acenaphthylene	13.833	152	5498	0.326	ng/ul	100
11) Acenaphthene	14.174	153	4930	0.334	ng/ul	99
12) Fluorene	15.168	166	5556	0.320	ng/ul	97
14) Pentachlorophenol	16.529	266	1156	0.555	ng/ul	97
15) Phenanthrene	16.897	178	8634	0.325	ng/ul	99
16) Anthracene	16.990	178	7215	0.309	ng/ul	99
19) Fluoranthene	18.919	202	10115	0.334	ng/ul	99
20) Pyrene	19.281	202	10506	0.333	ng/ul	99
21) Benzo(a)anthracene	21.028	228	7680	0.322	ng/ul	99
22) Chrysene	21.081	228	9193	0.342	ng/ul	99
24) Benzo(b)fluoranthene	22.526	252	8339	0.336	ng/ul	96
25) Benzo(k)fluoranthene	22.568	252	9207	0.346	ng/ul	94
26) Benzo(a)pyrene	23.059	252	6746	0.330	ng/ul	95
27) Indeno(1,2,3-cd)pyrene	25.208	276	8267	0.351	ng/ul#	99
28) Dibenzo(a,h)anthracene	25.216	278	6608	0.354	ng/ul	96
29) Benzo(g,h,i)perylene	25.836	276	7271	0.356	ng/ul	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM110921\
Data File : BM033012.D
Acq On : 12 Nov 2021 05:16
Operator : CG/JU
Sample : PB140387BS
Misc :
ALS Vial : 108 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SLCS387

Quant Time: Nov 12 05:50:44 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM110921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 11 13:40:18 2021
Response via : Initial Calibration

