

Data Path : Z:\HPCHEM1\BNA_M\Data\BM111015\
Data File : BM003083.D
Acq On : 10 Nov 2015 18:50
Operator : UM/IZ
Sample : SSTDCCC002EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD00240

Quant Time: Nov 13 10:45:37 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-1,4-DIO SIM-BM110915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 10 02:32:24 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
2 SURR	1,4-Dioxane-d8	0.391	0.375	4.1	110	-0.02
3	1,4-Dioxane	0.466	0.431	7.5	106	-0.02
4 I	Naphthalene-d8	1.000	1.000	0.0	121	0.00
5 I	Acenaphthene-d10	1.000	1.000	0.0	158#	-0.02
6 I	Phenanthrene-d10	1.000	1.000	0.0	179#	-0.02
7 I	Chrysene-d12	1.000	1.000	0.0	172#	0.00
8 I	Perylene-d12	1.000	1.000	0.0	158#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0