


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM111015\
 Data File : BM003075.D
 Acq On : 10 Nov 2015 13:59
 Operator : UM/IZ
 Sample : DFTPP41
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP41

Integration File: RTEINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-1,4-DIO SIM-BM110915.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Nov 10 02:32:24 2015



AutoFind: Scans 2464, 2465, 2466; Background Corrected with Scan 2458

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	28.9	27626	PASS
68	69	0.00	2	1.3	421	PASS
69	198	0.00	100	33.5	32000	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	48.1	46000	PASS
197	198	0.00	1	0.3	312	PASS
198	198	100	100	100.0	95584	PASS
199	198	5	9	6.8	6479	PASS
275	198	10	60	25.3	24154	PASS
365	198	1	100	3.4	3213	PASS
441	443	0.01	100	79.5	14222	PASS
442	198	50	100	97.1	92821	PASS
443	442	15	24	19.3	17886	PASS

m/z	Abundance
38.00	370.0
39.10	2044.0
40.00	401.0
41.00	424.0
43.00	897.0
44.00	878.0
45.00	489.0
46.90	583.0
48.00	303.0
49.00	3212.0
50.10	7086.0
51.00	27272.0
52.00	1481.0
55.00	474.0
56.00	1019.0
57.00	1866.0
60.10	433.0
61.00	359.0
62.00	413.0
63.00	1194.0
65.00	599.0
67.90	666.0
69.00	29792.0
73.00	445.0
74.00	3069.0
75.00	5123.0
76.00	1726.0
77.10	35448.0
78.00	2784.0
79.00	2268.0
80.00	1681.0
81.00	2579.0
82.00	822.0
83.00	728.0
84.00	2784.0
85.00	710.0
86.00	2602.0
87.00	493.0
87.90	455.0
91.00	726.0
91.90	643.0
93.00	4158.0
94.00	380.0
98.00	3229.0
99.00	3056.0
101.00	1564.0
103.00	602.0
104.00	732.0
105.00	980.0
107.00	11335.0
108.00	1760.0
109.00	361.0
110.00	20440.0
111.00	3302.0
111.90	407.0
116.00	672.0
117.00	9167.0
118.00	740.0
122.00	751.0
123.00	1421.0
124.00	612.0
124.90	487.0
127.00	39880.0
128.00	3303.0
129.00	14875.0
130.00	1325.0
131.00	304.0
134.00	358.0
134.90	1313.0
136.10	476.0
137.00	970.0
141.00	1825.0
142.00	581.0
143.00	489.0
145.90	317.0
147.00	1099.0
148.00	2607.0
149.00	706.0
150.90	342.0
153.00	606.0
154.00	463.0
155.00	1168.0
156.00	1584.0
158.10	323.0
160.00	628.0
161.00	895.0
162.00	305.0
165.00	612.0
166.00	750.0
167.10	3959.0
168.00	3255.0

Instrument :
BNA_M
ClientSampleId :
DFTPP41

169.00	418.0
171.90	339.0
173.00	419.0
174.00	843.0
175.00	1583.0
176.00	456.0
177.00	795.0
179.00	2868.0
180.00	2064.0
181.00	999.0
185.00	1339.0
186.00	11185.0
187.00	3201.0
188.00	334.0
189.00	650.0
191.00	372.0
192.00	881.0
193.00	1027.0
196.00	2326.0
196.80	452.0
198.00	77544.0
199.00	5355.0
200.00	534.0
201.40	429.0
203.00	614.0
204.00	2779.0
205.00	4711.0
206.10	19928.0
207.10	3052.0
208.00	726.0
210.20	374.0
211.00	867.0
216.00	308.0
217.00	5079.0
218.00	591.0
221.00	2897.0
223.00	1148.0
224.10	10613.0
225.10	2834.0
226.00	399.0
227.00	4569.0
228.00	672.0
229.00	1162.0
231.10	424.0
235.10	312.0
236.90	523.0
240.90	311.0
242.00	537.0
243.00	700.0
244.10	7814.0
245.10	1206.0
246.00	1632.0
247.00	397.0
248.90	316.0
255.10	40424.0
256.10	6071.0
257.10	488.0
258.00	2320.0
259.00	434.0
265.00	925.0
273.00	1180.0
274.00	2874.0
275.10	17904.0
276.00	2485.0
277.00	1626.0
278.00	339.0
293.00	349.0
296.00	5079.0
297.10	700.0
303.00	571.0
315.00	455.0
323.00	1500.0
324.10	328.0
327.00	305.0
334.10	1051.0
346.00	409.0
352.00	506.0
353.00	343.0
354.10	518.0
365.00	1990.0
372.00	836.0
402.00	340.0
403.00	440.0
421.00	366.0
422.10	386.0
423.10	2740.0
424.10	951.0
441.10	8404.0
442.10	55760.0
443.10	11203.0
444.10	1133.0

Instrument :
BNA_M
ClientSampleId :
DFTPP41

m/z	Abundance
38.00	439.0
39.00	2477.0
40.00	488.0
41.00	548.0
43.10	834.0
44.00	898.0
45.00	627.0
46.90	551.0
49.00	3029.0
50.00	8912.0
51.10	34184.0
52.10	1729.0
55.00	495.0
56.00	1247.0
57.00	2453.0
60.00	356.0
61.00	548.0
62.00	647.0
63.00	1571.0
65.00	754.0
68.10	717.0
69.00	37528.0
73.00	607.0
74.00	4095.0
75.00	6185.0
76.10	2202.0
77.10	46552.0
78.10	3532.0
79.00	3084.0
80.00	2329.0
81.00	3328.0
82.00	886.0
83.00	897.0
84.00	2860.0
84.90	810.0
86.00	2588.0
87.00	470.0
88.00	445.0
90.90	861.0
91.90	986.0
93.00	5340.0
94.00	439.0
96.00	365.0
98.00	4239.0
99.00	3636.0
101.00	1819.0
102.90	703.0
104.00	1199.0
105.00	1179.0
106.00	385.0
107.00	14464.0
108.00	2278.0
109.00	429.0
110.00	27368.0
111.00	4148.0
112.00	621.0
116.00	942.0
117.00	12760.0
118.00	921.0
121.90	1071.0
123.00	1829.0
124.00	859.0
125.00	610.0
127.00	54760.0
128.00	3935.0
129.00	20800.0
130.00	1732.0
130.90	395.0
134.00	703.0
135.00	1740.0
136.00	771.0
137.10	1060.0
141.00	2714.0
142.00	872.0
143.10	728.0
146.00	575.0
147.00	1491.0
148.00	3200.0
149.00	769.0
151.10	331.0
152.90	887.0
154.00	678.0
155.00	1580.0
156.10	2214.0
157.00	475.0
158.00	530.0
159.00	417.0
160.00	811.0
161.00	1295.0
161.90	405.0
165.00	989.0

Instrument :
BNA_M
ClientSampleId :
DFTPP41

166.00	886.0
167.00	5883.0
168.00	3912.0
169.00	564.0
172.00	490.0
173.00	551.0
174.00	1091.0
175.10	2103.0
176.00	536.0
177.00	1023.0
178.00	326.0
179.00	4365.0
180.00	3360.0
181.00	1275.0
182.00	304.0
184.00	325.0
185.00	2008.0
186.00	16261.0
187.00	4580.0
188.00	561.0
189.00	984.0
191.10	604.0
192.00	1234.0
193.00	1460.0
194.00	376.0
196.00	3023.0
197.00	538.0
198.00	113192.0
199.00	7909.0
200.00	507.0
201.70	600.0
203.00	746.0
204.00	4098.0
205.10	6936.0
206.10	30008.0
207.10	4348.0
208.00	1056.0
209.00	481.0
210.30	447.0
211.10	1235.0
216.00	629.0
217.00	7823.0
218.00	986.0
221.00	4312.0
221.90	1189.0
223.00	1554.0
224.10	16294.0
225.10	4408.0
226.00	486.0
227.00	6367.0
228.00	961.0
229.00	1390.0
231.00	588.0
233.90	432.0
235.00	469.0
236.00	368.0
237.00	649.0
241.00	445.0
242.10	834.0
243.00	810.0
244.10	12628.0
245.10	1626.0
246.00	2532.0
247.00	638.0
248.90	498.0
254.10	408.0
255.10	64328.0
256.10	9558.0
257.10	806.0
258.00	3640.0
258.90	579.0
264.90	1599.0
273.00	1815.0
274.00	5615.0
275.10	28352.0
276.00	3700.0
277.00	2414.0
278.00	566.0
285.00	446.0
293.10	540.0
296.10	8103.0
297.10	1340.0
303.00	1003.0
304.10	330.0
314.00	329.0
315.10	783.0
316.10	591.0
323.10	2555.0
324.10	420.0
327.00	480.0
334.10	1770.0
335.00	485.0
341.10	307.0
346.00	689.0
352.00	904.0

Instrument :
BNA_M
ClientSampleId :
DFTPP41

353.10	639.0
354.10	822.0
365.00	3998.0
366.00	506.0
372.10	1453.0
373.10	324.0
383.00	403.0
402.10	531.0
403.00	912.0
404.10	332.0
420.90	585.0
422.00	811.0
423.10	5143.0
424.10	1645.0
441.10	16672.0
442.10	107224.0
443.10	20768.0
444.10	2108.0

Instrument :
BNA_M
ClientSampleId :
DFTPP41

m/z	Abundance
38.00	361.0
39.00	1903.0
40.00	452.0
41.00	481.0
43.00	991.0
44.00	930.0
45.00	456.0
47.00	515.0
47.90	312.0
49.00	3162.0
50.10	6244.0
51.10	25224.0
52.00	1234.0
55.00	453.0
56.00	916.0
57.00	2045.0
61.00	474.0
62.00	391.0
63.00	1072.0
65.00	657.0
68.00	547.0
69.00	28680.0
73.00	482.0
74.00	3005.0
75.00	4328.0
76.10	1836.0
77.10	35760.0
78.10	2745.0
79.00	2250.0
80.00	1805.0
81.00	2551.0
82.00	725.0
83.00	753.0
84.00	2843.0
85.00	602.0
86.00	2571.0
87.00	431.0
87.90	458.0
91.00	742.0
91.90	735.0
93.00	4143.0
93.90	329.0
98.00	3285.0
99.00	2856.0
101.00	1800.0
103.00	602.0
104.00	852.0
105.00	1240.0
107.00	11947.0
108.00	1801.0
109.00	307.0
110.00	21520.0
111.00	3471.0
112.00	379.0
116.00	706.0
117.00	10691.0
118.00	746.0
122.00	878.0
123.00	1502.0
124.00	586.0
125.00	567.0
127.00	43360.0
128.00	3370.0
129.00	16536.0
130.00	1326.0
133.90	493.0
135.00	1358.0
135.90	671.0
137.00	896.0
141.00	2154.0
142.00	783.0
142.90	509.0
146.00	435.0
147.10	1198.0
148.00	2562.0
149.00	657.0
151.00	335.0
153.10	636.0
154.00	579.0
155.00	1297.0
156.00	1878.0
157.00	430.0
158.00	369.0
158.90	329.0
160.00	757.0
161.00	1097.0
165.00	977.0
166.00	628.0
167.00	4533.0
168.00	3049.0
169.00	482.0

Instrument :
BNA_M
ClientSampleId :
DFTPP41

170.90	301.0
172.00	458.0
173.00	584.0
174.00	1067.0
175.00	1919.0
176.10	525.0
177.00	894.0
179.00	3427.0
180.10	2454.0
181.00	1070.0
185.00	1728.0
186.10	12658.0
187.00	3891.0
188.10	369.0
189.00	840.0
190.90	418.0
192.00	1357.0
193.00	1359.0
196.00	2692.0
196.80	485.0
198.00	96016.0
199.00	6173.0
200.00	487.0
201.40	430.0
203.00	654.0
204.00	3461.0
205.10	5758.0
206.10	25672.0
207.00	3579.0
208.00	965.0
209.00	393.0
210.10	433.0
211.10	835.0
216.00	655.0
217.00	6939.0
218.00	720.0
221.10	4084.0
221.90	887.0
223.00	1323.0
224.10	13925.0
225.10	3716.0
226.10	371.0
227.00	5877.0
228.00	848.0
229.00	1308.0
231.00	377.0
233.90	397.0
234.90	384.0
237.10	542.0
241.00	345.0
242.00	869.0
243.00	864.0
244.10	11345.0
245.00	1489.0
246.10	2154.0
247.00	513.0
249.00	518.0
254.10	421.0
255.00	57320.0
256.00	8535.0
257.10	563.0
258.00	3465.0
258.90	457.0
265.10	1304.0
273.00	1775.0
274.00	4655.0
275.10	26208.0
276.10	3527.0
277.00	2331.0
278.00	410.0
285.00	479.0
293.00	447.0
296.00	7364.0
297.00	1072.0
303.10	878.0
314.00	414.0
315.00	875.0
316.10	471.0
323.10	2629.0
324.10	416.0
327.00	550.0
334.00	1679.0
335.00	454.0
346.10	502.0
352.10	746.0
353.10	472.0
354.10	936.0
365.10	3651.0
366.00	543.0
372.10	1430.0
383.10	445.0
402.00	588.0
403.10	775.0
404.00	325.0
421.10	716.0

Instrument :
BNA_M
ClientSampleId :
DFTPP41

422.10	571.0
423.10	5449.0
424.10	1484.0
441.10	17592.0
442.10	115480.0
443.10	21688.0
444.10	1937.0

Instrument :
BNA_M
ClientSampleId :
DFTPP41

m/z	Abundance
40.00	451.0
41.00	387.0
42.00	306.0
43.00	941.0
44.00	814.0
45.00	549.0
46.90	608.0
48.00	307.0
49.00	3140.0
51.00	1267.0
57.00	307.0
60.00	384.0
73.00	325.0
84.00	2867.0
85.90	1791.0
87.90	338.0
207.00	388.0

Instrument :
BNA_M
ClientSampleId :
DFTPP41