

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA M\DATA\BM111017\
 Data File : BM012914.D
 Acq On : 11 Nov 2017 22:22
 Operator : SJ/JU
 Sample : I6317-08MS
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DAHJ3MS

Manual Integrations
 APPROVED

SOHIL
 11/13/2017 4:49:46 PM

Quant Time: Nov 13 03:02:59 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM111017MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Nov 13 02:06:47 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.78	152	79746	20.00	ng/ul	0.00
2) Naphthalene-d8	10.57	136	301288	20.00	ng/ul	0.00
6) Acenaphthene-d10	14.42	164	172635	20.00	ng/ul	0.00
12) Phenanthrene-d10	17.17	188	404468	20.00	ng/ul	0.00
17) Chrysene-d12	21.35	240	490424	20.00	ng/ul	0.00
22) Perylene-d12	23.62	264	551049	20.00	ng/ul	0.00
System Monitoring Compounds						
7) Acenaphthylene-d8	14.12	160	283891	15.81	ng/ul	0.00
10) Fluorene-d10	15.42	176	211685	17.24	ng/ul	0.00
14) Anthracene-d10	17.27	188	314832	16.10	ng/ul	0.00
18) Pyrene-d10	19.56	212	363875	16.58	ng/ul	0.00
25) Benzo(a)pyrene-d12	23.48	264	397753	15.29	ng/ul	0.00
Target Compounds						
9) Acenaphthene	14.49	153	185050	15.815	ng/ul	96
13) Phenanthrene	17.21	178	54387	2.434	ng/ul	99
16) Fluoranthene	19.23	202	133350	4.852	ng/ul#	92
19) Pyrene	19.59	202	535458	19.020	ng/ul#	92
20) Benzo(a)anthracene	21.33	228	96726	3.205	ng/ul	98
21) Chrysene	21.39	228	110742	4.107	ng/ul	98
23) Benzo(b)fluoranthene	22.94	252	238241m	7.270	ng/ul	
24) Benzo(k)fluoranthene	22.99	252	83163m	2.655	ng/ul	
26) Benzo(a)pyrene	23.53	252	118886m	3.741	ng/ul	
27) Indeno(1,2,3-cd)pyrene	25.93	276	122900	2.998	ng/ul	96
29) Benzo(g,h,i)perylene	26.63	276	110416	3.227	ng/ul	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA M\DATA\BM111017\
Data File : BM012914.D
Acq On : 11 Nov 2017 22:22
Operator : SJ/JU
Sample : I6317-08MS
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DAHJ3MS

Manual Integrations
APPROVED

SOHIL
11/13/2017 4:49:46 PM

Quant Time: Nov 13 03:02:59 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM111017MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Nov 13 02:06:47 2017
Response via : Initial Calibration

