


Data Path : Z:\HPCHEM1\BNA M\DATA\BM111115\
 Data File : BM003084.D
 Acq On : 11 Nov 2015 14:02
 Operator : UM/IZ
 Sample : DFTPP42
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP42

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM102815.M
 Title : SVOA CALIBRATION
 Last Update : Wed Oct 28 15:26:49 2015



AutoFind: Scans 2464, 2465, 2466; Background Corrected with Scan 2459

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.3	29954	PASS
68	69	0.00	2	1.7	570	PASS
69	198	0.00	100	33.3	34072	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	48.0	49184	PASS
197	198	0.00	2	0.3	308	PASS
198	198	100	100	100.0	102362	PASS
199	198	5	9	6.7	6877	PASS
275	198	10	60	25.7	26352	PASS
365	198	1	100	3.2	3305	PASS
441	443	0.01	100	78.3	15268	PASS
442	198	50	100	97.6	99909	PASS
443	442	15	24	19.5	19494	PASS

Scan 2459 (17.551 min): BM003084.D
DFTPP42

m/z	Abundance
44.00	381.0
207.00	368.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42

m/z	Abundance
38.00	304.0
39.00	1783.0
40.00	342.0
44.00	530.0
50.10	6834.0
51.10	25648.0
52.10	1425.0
56.00	863.0
57.00	1748.0
61.00	363.0
62.00	341.0
63.10	1034.0
65.00	679.0
68.10	510.0
69.00	28752.0
73.10	386.0
74.00	3009.0
75.00	4820.0
76.10	1615.0
77.10	35536.0
78.00	2511.0
79.00	2412.0
80.00	1833.0
81.00	2485.0
82.00	603.0
83.00	521.0
85.00	444.0
86.00	824.0
87.00	412.0
91.00	583.0
92.00	646.0
93.00	4239.0
98.00	3148.0
99.00	2969.0
99.90	335.0
101.00	1424.0
102.90	507.0
104.00	837.0
105.00	872.0
106.00	355.0
107.00	11331.0
108.00	2009.0
110.00	19712.0
111.00	3253.0
112.10	474.0
116.00	638.0
117.00	9549.0
118.00	819.0
122.00	734.0
123.00	1308.0
124.10	464.0
125.00	473.0
127.00	39016.0
128.00	2957.0
129.00	14629.0
130.00	1334.0
134.00	462.0
135.00	1133.0
136.00	473.0
137.00	957.0
141.00	1851.0
141.90	690.0
143.00	416.0
146.00	404.0
147.10	1012.0
148.00	2285.0
149.00	638.0
153.00	533.0
154.00	385.0
155.00	955.0
156.00	1637.0
157.00	376.0
157.90	337.0
160.00	516.0
161.00	893.0
165.00	683.0
166.10	519.0
167.00	4026.0
168.00	3866.0
169.00	537.0
172.00	366.0
173.00	349.0
174.00	821.0
175.00	1676.0
175.90	467.0
177.10	639.0
179.00	3003.0
180.00	2264.0
181.00	1029.0
185.00	1284.0
186.10	10527.0
187.00	3106.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42

188.00	422.0
189.00	636.0
191.00	301.0
192.00	866.0
193.00	1057.0
196.10	1925.0
197.00	394.0
198.00	75856.0
199.00	5052.0
200.00	379.0
201.50	338.0
203.00	465.0
204.00	2649.0
205.10	4651.0
206.10	20168.0
207.10	3036.0
208.10	813.0
210.90	795.0
216.00	450.0
217.00	5053.0
217.90	723.0
221.10	2684.0
221.80	878.0
223.00	1038.0
224.10	10300.0
225.10	2818.0
226.00	313.0
227.00	4583.0
228.00	650.0
229.00	821.0
234.90	387.0
237.00	495.0
242.10	498.0
243.00	615.0
244.10	7943.0
245.10	1222.0
246.00	1544.0
254.10	358.0
255.00	41152.0
256.10	6194.0
257.00	518.0
258.00	2368.0
259.10	466.0
265.00	906.0
273.00	1171.0
274.10	3070.0
275.10	17424.0
276.00	2404.0
277.00	1502.0
278.00	353.0
293.00	396.0
296.00	5336.0
297.10	640.0
303.10	665.0
315.10	524.0
316.00	303.0
323.10	1551.0
327.00	330.0
334.00	903.0
335.00	340.0
352.10	501.0
353.00	307.0
354.00	546.0
365.00	1886.0
372.10	741.0
403.00	469.0
421.10	390.0
422.10	352.0
423.10	2716.0
424.10	887.0
441.10	8204.0
442.10	53632.0
443.10	10836.0
444.10	979.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42

m/z	Abundance
38.10	340.0
39.10	2495.0
40.00	304.0
44.00	399.0
50.10	8836.0
51.10	35792.0
52.00	1961.0
56.00	1143.0
57.00	2646.0
61.00	524.0
62.00	560.0
63.00	1659.0
65.00	827.0
68.00	659.0
69.00	40848.0
73.10	401.0
74.00	4270.0
75.00	6647.0
76.00	2345.0
77.10	49128.0
78.00	3603.0
79.00	2998.0
80.00	2420.0
81.00	3454.0
82.00	926.0
83.00	825.0
85.00	600.0
86.00	1015.0
87.10	441.0
91.00	805.0
92.00	887.0
93.00	5689.0
94.00	491.0
98.00	4848.0
99.00	4125.0
100.00	364.0
101.00	2099.0
102.90	677.0
104.00	1158.0
105.00	1151.0
106.00	368.0
107.00	16408.0
108.00	2346.0
109.00	479.0
110.00	29040.0
111.00	4756.0
112.00	548.0
116.00	922.0
117.00	13781.0
118.00	1146.0
122.00	1019.0
123.00	1497.0
124.00	830.0
125.00	730.0
127.00	58928.0
128.00	4244.0
129.00	22144.0
130.00	1896.0
131.00	337.0
133.90	521.0
135.00	1876.0
136.00	829.0
137.10	1238.0
139.90	305.0
141.00	2906.0
142.00	1046.0
143.00	732.0
146.00	513.0
147.00	1477.0
148.00	3757.0
149.00	800.0
151.00	353.0
153.00	973.0
154.00	704.0
155.00	1620.0
156.00	2519.0
157.00	383.0
158.00	589.0
159.00	459.0
160.00	999.0
161.00	1269.0
162.00	456.0
165.00	1009.0
166.10	869.0
167.00	5808.0
168.00	4288.0
169.00	610.0
172.00	550.0
173.00	743.0
174.00	1380.0
175.00	2656.0
175.90	611.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42

177.00	1151.0
178.10	385.0
179.00	4405.0
180.00	3401.0
181.10	1452.0
184.00	343.0
185.00	2188.0
186.00	16784.0
187.00	4822.0
188.00	565.0
189.00	1093.0
190.90	548.0
192.00	1494.0
193.00	1710.0
194.00	309.0
196.00	3282.0
198.00	119384.0
199.00	8183.0
200.00	646.0
201.60	602.0
203.00	865.0
204.10	4532.0
205.00	7358.0
206.10	31552.0
207.10	4485.0
208.00	1115.0
209.00	395.0
210.10	545.0
211.00	1273.0
215.00	430.0
216.00	723.0
217.00	8641.0
218.00	1089.0
221.10	4845.0
221.80	1203.0
223.00	1817.0
224.00	17592.0
225.10	4506.0
226.00	525.0
227.00	7618.0
228.10	969.0
229.00	1510.0
231.10	671.0
233.90	401.0
235.00	518.0
235.90	394.0
237.10	790.0
238.90	306.0
240.90	386.0
242.00	977.0
243.00	910.0
244.10	13319.0
245.10	1970.0
246.00	2505.0
247.00	516.0
248.90	501.0
253.00	322.0
254.10	438.0
255.10	69104.0
256.00	10180.0
257.00	946.0
258.00	4083.0
259.00	704.0
265.00	1649.0
273.00	2152.0
274.10	5692.0
275.10	31336.0
276.10	4217.0
277.00	2466.0
278.00	492.0
285.10	517.0
293.10	556.0
296.00	9049.0
297.00	1338.0
303.10	1047.0
304.10	302.0
314.10	363.0
315.00	820.0
316.10	545.0
323.10	2986.0
324.10	510.0
326.90	579.0
333.10	325.0
334.10	1938.0
335.10	458.0
341.10	333.0
345.90	481.0
352.10	895.0
353.10	585.0
354.10	936.0
365.00	3829.0
366.00	538.0
372.10	1690.0
373.00	374.0
383.10	399.0
402.10	624.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42

403.00	927.0
404.10	329.0
421.10	801.0
422.10	788.0
423.10	5199.0
424.10	1915.0
441.10	16896.0
442.10	113872.0
443.10	22456.0
444.10	2174.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42

m/z	Abundance
38.10	387.0
39.00	1786.0
39.90	317.0
44.00	453.0
50.00	7225.0
51.10	28424.0
52.00	1340.0
56.00	844.0
57.00	2025.0
60.90	382.0
62.00	415.0
63.00	1404.0
65.00	735.0
68.00	541.0
69.00	32616.0
74.00	3336.0
75.00	5689.0
76.10	1837.0
77.10	40720.0
78.00	2824.0
79.00	2689.0
80.00	1786.0
81.00	2752.0
82.00	764.0
83.00	698.0
85.00	571.0
85.90	799.0
86.90	372.0
90.90	630.0
92.00	824.0
93.00	4626.0
98.00	3619.0
99.00	3061.0
101.00	1691.0
102.90	575.0
104.00	1069.0
105.00	1051.0
105.90	378.0
107.00	13485.0
108.00	2017.0
109.10	320.0
110.00	24352.0
111.00	3728.0
111.90	499.0
115.90	696.0
117.00	11736.0
118.00	859.0
122.00	784.0
123.00	1449.0
123.90	667.0
125.00	597.0
127.00	49608.0
128.10	3756.0
129.00	19016.0
130.00	1471.0
131.00	358.0
133.90	589.0
135.00	1673.0
136.00	666.0
137.00	943.0
141.00	2603.0
142.00	944.0
142.90	553.0
146.00	483.0
147.10	1516.0
148.00	3186.0
149.00	740.0
153.00	801.0
154.00	577.0
155.00	1273.0
156.00	2285.0
157.00	389.0
158.00	496.0
159.00	461.0
160.00	849.0
161.00	1115.0
165.00	896.0
165.90	790.0
167.10	5774.0
168.00	3668.0
169.00	502.0
172.00	573.0
173.00	562.0
174.00	1307.0
175.00	2237.0
176.00	495.0
177.00	1029.0
179.00	4027.0
180.00	3078.0
181.00	1316.0
183.90	308.0
185.00	1865.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42

186.10	15523.0
187.00	4364.0
188.00	527.0
189.00	1054.0
191.00	400.0
192.00	1246.0
193.10	1602.0
196.10	3074.0
197.10	530.0
198.00	111848.0
199.00	7398.0
199.90	519.0
201.60	457.0
203.00	805.0
204.10	4362.0
205.10	6811.0
206.10	29520.0
207.10	4228.0
208.00	1082.0
209.10	368.0
211.00	1238.0
216.00	662.0
217.00	7675.0
218.00	960.0
221.10	4221.0
221.80	1166.0
223.00	1683.0
224.10	16848.0
225.10	4073.0
226.00	514.0
227.00	7061.0
228.10	997.0
229.00	1549.0
231.10	573.0
233.90	474.0
235.00	452.0
236.00	327.0
237.10	745.0
241.00	365.0
242.00	992.0
243.10	786.0
244.10	12368.0
245.10	1611.0
246.10	2792.0
247.00	506.0
249.10	470.0
253.00	359.0
254.00	423.0
255.00	68176.0
256.10	10196.0
257.00	753.0
258.00	3883.0
259.10	683.0
265.00	1706.0
273.10	2000.0
274.00	5352.0
275.10	30296.0
276.10	3815.0
277.00	2816.0
278.00	561.0
285.10	437.0
293.00	732.0
296.10	9070.0
297.00	1251.0
303.10	985.0
314.00	439.0
315.00	1081.0
316.00	534.0
323.10	3019.0
324.10	532.0
327.10	548.0
328.00	345.0
334.10	1860.0
335.00	566.0
346.00	612.0
352.10	925.0
353.10	648.0
354.10	1024.0
365.10	4201.0
366.10	587.0
372.10	1738.0
373.00	413.0
383.10	407.0
402.00	701.0
403.00	883.0
421.10	929.0
422.10	873.0
423.10	6234.0
424.10	1764.0
441.10	20704.0
442.10	132224.0
443.10	25192.0
444.10	2466.0

Instrument :
BNA_M
ClientSampleId :
DFTPP42