

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM111116\
 Data File : BM007817.D
 Acq On : 11 Nov 2016 10:05
 Operator : UM/SJ
 Sample : PB94641BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK41

Quant Time: Nov 11 11:44:54 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM102016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Nov 11 11:39:59 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.74	152	168283	20.00	ng/ul	0.00
18) Naphthalene-d8	10.52	136	632376	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.37	164	389597	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.12	188	764543	20.00	ng/ul	0.00
75) Chrysene-d12	21.31	240	954882	20.00	ng/ul	0.00
83) Perylene-d12	23.56	264	994437	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.29	96	30852	7.86	ng/uL	0.00
5) Phenol-d5	6.92	99	498310	38.37	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.08	67	298375	38.71	ng/ul	0.00
9) 2-Chlorophenol-d4	7.28	132	398659	40.02	ng/ul	0.00
13) 4-Methylphenol-d8	8.45	113	389130	38.29	ng/ul	0.00
19) Nitrobenzene-d5	8.90	128	188231	41.36	ng/ul	0.00
22) 2-Nitrophenol-d4	9.62	143	205063	42.50	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.15	165	369562	38.96	ng/ul	0.00
29) 4-Chloroaniline-d4	10.66	131	484583	44.79	ng/ul	0.00
43) Dimethylphthalate-d6	13.79	166	1113645	41.98	ng/ul	0.00
46) Acenaphthylene-d8	14.07	160	1328748	41.62	ng/ul	0.00
51) 4-Nitrophenol-d4	14.59	143	138345	32.98	ng/ul	0.00
57) Fluorene-d10	15.37	176	948614	40.80	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.51	200	148219	32.63	ng/ul	0.00
70) Anthracene-d10	17.22	188	1484304	43.08	ng/ul	0.00
76) Pyrene-d10	19.52	212	1696954	43.25	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.42	264	1867105	43.01	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM111116\
Data File : BM007817.D
Acq On : 11 Nov 2016 10:05
Operator : UM/SJ
Sample : PB94641BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK41

Quant Time: Nov 11 11:44:54 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM102016.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Nov 11 11:39:59 2016
Response via : Initial Calibration

