

Data Path : Z:\HPCHEM1\BNA_M\Data\BM111116\
 Data File : BM007829.D
 Acq On : 11 Nov 2016 17:10
 Operator : UM/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02055

Quant Time: Nov 11 17:48:45 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM102016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 09 23:34:58 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.510	0.509	0.2	101	0.00
3 S	1,4-Dioxane-d8	0.467	0.466	0.2	99	0.00
4	Benzaldehyde	0.976	1.133	-16.1	105	0.00
5 S	Phenol-d5	1.544	1.587	-2.8	108	0.00
6	Phenol	1.584	1.622	-2.4	106	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	0.916	0.944	-3.1	106	0.00
8	Bis(2-Chloroethyl)ether	1.221	1.261	-3.3	105	0.00
9 S	2-Chlorophenol-d4	1.184	1.251	-5.7	107	0.00
10	2-Chlorophenol	1.215	1.265	-4.1	105	0.00
11	2-Methylphenol	1.153	1.171	-1.6	107	0.00
12	2,2'-oxybis(1-Chloropropane	1.353	1.441	-6.5	108	0.00
13 S	4-Methylphenol-d8	1.208	1.204	0.3	104	0.00
14	Acetophenone	1.894	1.963	-3.6	107	0.00
15 P	N-Nitroso-di-n-propylamine	0.940	1.000	-6.4	109	0.00
16	4-Methylphenol	1.243	1.257	-1.1	106	0.00
17	Hexachloroethane	0.537	0.559	-4.1	103	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
19 S	Nitrobenzene-d5	0.144	0.153	-6.3	106	0.00
20	Nitrobenzene	0.370	0.386	-4.3	106	0.00
21	Isophorone	0.633	0.687	-8.5	110	0.00
22 S	2-Nitrophenol-d4	0.153	0.167	-9.2	112	0.00
23 C	2-Nitrophenol	0.159	0.172	-8.2	107	0.00
24	2,4-Dimethylphenol	0.364	0.378	-3.8	104	0.00
25	Bis(2-Chloroethoxy)methane	0.387	0.409	-5.7	107	0.00
26 S	2,4-Dichlorophenol-d3	0.300	0.322	-7.3	108	0.00
27 C	2,4-Dichlorophenol	0.295	0.313	-6.1	106	0.00
28	Naphthalene	0.973	1.009	-3.7	106	0.00
29 S	4-Chloroaniline-d4	0.342	0.387	-13.2	111	0.00
30	4-Chloroaniline	0.347	0.387	-11.5	109	0.00
31 C	Hexachlorobutadiene	0.241	0.247	-2.5	104	0.00
32	Caprolactam	0.082	0.084	-2.4	115	0.00
33 C	4-Chloro-3-methylphenol	0.310	0.332	-7.1	108	0.00
34	2-Methylnaphthalene	0.692	0.713	-3.0	104	0.00
35 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
36	1,2,4,5-Tetrachlorobenzene	0.646	0.671	-3.9	106	0.00
37	Hexachlorocyclopentadiene	0.380	0.360	5.3	106	0.00
38 C	2,4,6-Trichlorophenol	0.359	0.388	-8.1	110	0.00
39	2,4,5-Trichlorophenol	0.393	0.419	-6.6	110	0.00
40	1,1'-Biphenyl	1.410	1.462	-3.7	105	0.00
41	2-Chloronaphthalene	1.079	1.114	-3.2	105	0.00
42	2-Nitroaniline	0.286	0.316	-10.5	112	0.00
43 S	Dimethylphthalate-d6	1.362	1.415	-3.9	107	0.00
44	Dimethylphthalate	1.322	1.363	-3.1	106	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM111116\
 Data File : BM007829.D
 Acq On : 11 Nov 2016 17:10
 Operator : UM/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02055

Quant Time: Nov 11 17:48:45 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM102016.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 09 23:34:58 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	2,6-Dinitrotoluene	0.249	0.267	-7.2	111	0.00
46 S	Acenaphthylene-d8	1.639	1.730	-5.6	107	0.00
47	Acenaphthylene	1.662	1.756	-5.7	107	0.00
48	3-Nitroaniline	0.246	0.272	-10.6	115	0.00
49 C	Acenaphthene	1.147	1.184	-3.2	107	0.00
50	2,4-Dinitrophenol	0.153	0.124	19.0	100	0.00
51 S	4-Nitrophenol-d4	0.215	0.216	-0.5	114	0.00
52	4-Nitrophenol	0.214	0.204	4.7	104	0.00
53	Dibenzofuran	1.647	1.706	-3.6	106	0.00
54	2,4-Dinitrotoluene	0.359	0.394	-9.7	112	0.00
55	2,3,4,6-Tetrachlorophenol	0.351	0.374	-6.6	110	0.00
56	Diethylphthalate	1.309	1.374	-5.0	109	0.00
57 S	Fluorene-d10	1.194	1.232	-3.2	106	0.00
58	Fluorene	1.315	1.377	-4.7	107	0.00
59	4-Chlorophenyl-phenylether	0.701	0.723	-3.1	106	0.00
60	4-Nitroaniline	0.248	0.259	-4.4	123	0.00
61 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
62 S	4,6-Dinitro-2-methylphenol-	0.119	0.107	10.1	106	0.00
63	4,6-Dinitro-2-methylphenol	0.123	0.111	9.8	103	0.00
64	N-Nitrosodiphenylamine	0.572	0.596	-4.2	108	0.00
65	4-Bromophenyl-phenylether	0.228	0.234	-2.6	107	0.00
66	Hexachlorobenzene	0.253	0.259	-2.4	106	0.00
67	Atrazine	0.220	0.226	-2.7	111	0.00
68 C	Pentachlorophenol	0.142	0.133	6.3	111	0.00
69	Phenanthrene	1.063	1.100	-3.5	109	0.00
70 S	Anthracene-d10	0.901	0.953	-5.8	109	0.00
71	Anthracene	1.083	1.118	-3.2	107	0.00
72	Carbazole	0.937	0.979	-4.5	110	0.00
73	Di-n-butylphthalate	1.044	1.140	-9.2	112	0.00
74 C	Fluoranthene	1.243	1.316	-5.9	111	0.00
75 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
76 S	Pyrene-d10	0.822	0.843	-2.6	110	0.00
77	Pyrene	1.033	1.050	-1.6	109	0.00
78	Butylbenzylphthalate	0.368	0.405	-10.1	118	0.00
79	3,3'-Dichlorobenzidine	0.355	0.371	-4.5	112	0.00
80	Benzo(a)anthracene	1.069	1.111	-3.9	112	0.00
81	Bis(2-ethylhexyl)phthalate	0.541	0.604	-11.6	119	0.00
82	Chrysene	1.027	1.060	-3.2	110	0.00
83 I	Perylene-d12	1.000	1.000	0.0	110	0.00
84	Di-n-octyl phthalate	1.014	1.048	-3.4	119	0.00
85	Benzo(b)fluoranthene	1.143	1.135	0.7	109	0.00
86	Benzo(k)fluoranthene	1.051	1.131	-7.6	117	0.00
87 S	Benzo(a)pyrene-d12	0.873	0.921	-5.5	116	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM111116\
Data File : BM007829.D
Acq On : 11 Nov 2016 17:10
Operator : UM/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02055

Quant Time: Nov 11 17:48:45 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM102016.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Nov 09 23:34:58 2016
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88 C	Benzo(a)pyrene	1.076	1.127	-4.7	114	0.00
89	Indeno(1,2,3-cd)pyrene	1.289	1.361	-5.6	115	0.00
90	Dibenzo(a,h)anthracene	1.088	1.145	-5.2	115	0.00
91	Benzo(g,h,i)perylene	1.083	1.130	-4.3	114	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0