


Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM111120\
 Data File : BM027782.D
 Acq On : 11 Nov 2020 08:58
 Operator : JU/CG
 Sample : DFTPP89
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP89

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM111020.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Nov 10 15:45:47 2020



AutoFind: Scans 2570, 2571, 2572; Background Corrected with Scan 2564

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	58.3	50509	PASS
68	69	0.00	2	1.8	872	PASS
69	198	0.00	100	55.0	47690	PASS
70	69	0.00	2	0.5	235	PASS
127	198	10	80	51.5	44677	PASS
197	198	0.00	2	0.8	660	PASS
198	198	100	100	100.0	86706	PASS
199	198	5	9	6.6	5757	PASS
275	198	10	60	27.0	23453	PASS
365	198	1	100	4.7	4117	PASS
441	198	0.01	100	13.4	11622	PASS
442	198	50	100	81.9	71002	PASS
443	442	15	24	19.8	14084	PASS

Scan 2564 (18.068 min): BM027782.D
DFTPP89

m/z	Abundance
73.00	155.0
207.00	406.0

Instrument :
BNA_M
ClientSampleId :
DFTPP89

m/z	Abundance
37.10	241.0
38.00	752.0
39.10	3634.0
40.00	220.0
41.00	169.0
43.90	259.0
49.00	415.0
50.00	14871.0
51.10	49200.0
52.00	2758.0
55.00	263.0
56.00	1957.0
57.00	3096.0
58.00	174.0
61.00	758.0
62.00	694.0
63.00	1861.0
64.00	334.0
65.00	744.0
68.00	857.0
69.00	46600.0
69.90	282.0
73.10	556.0
74.10	5142.0
75.00	7446.0
76.00	2310.0
77.10	43856.0
78.00	2759.0
79.00	3883.0
80.00	2918.0
81.00	3675.0
81.90	1020.0
83.00	633.0
83.80	189.0
85.00	607.0
86.00	1293.0
86.90	422.0
91.00	840.0
92.00	952.0
93.00	7722.0
94.00	419.0
96.00	266.0
98.00	5465.0
99.00	4197.0
99.90	287.0
101.00	2228.0
103.00	747.0
103.90	1288.0
105.00	1283.0
106.10	415.0
107.00	14957.0
108.00	2080.0
109.00	364.0
110.00	20640.0
111.00	4160.0
112.00	527.0
113.00	348.0
115.90	782.0
117.00	17128.0
118.00	1182.0
122.00	1020.0
123.00	1653.0
123.90	750.0
124.90	717.0
125.80	222.0
126.20	227.0
127.00	39920.0
128.00	3054.0
129.00	16936.0
130.00	1828.0
131.00	252.0
132.00	156.0
134.00	630.0
135.00	1350.0
136.00	773.0
137.00	697.0
138.00	163.0
139.90	233.0
141.00	2699.0
142.00	592.0
142.90	524.0
143.90	218.0
146.00	525.0
147.00	1519.0
148.00	3339.0
149.00	725.0
151.00	379.0
151.80	206.0
153.00	707.0
154.00	594.0
155.00	1391.0
156.00	1757.0

Instrument :
BNA_M
ClientSampleId :
DFTPP89

157.00	433.0
158.00	486.0
159.00	275.0
159.90	846.0
160.90	1237.0
161.90	405.0
164.90	1256.0
166.00	589.0
167.00	5113.0
168.00	2730.0
168.90	416.0
170.00	165.0
170.90	154.0
171.90	466.0
172.90	540.0
174.00	779.0
175.00	1573.0
176.00	415.0
177.00	1037.0
178.00	303.0
178.90	3415.0
180.00	2345.0
180.90	1044.0
181.80	259.0
184.00	220.0
185.00	1463.0
186.00	11720.0
187.00	3826.0
188.00	270.0
189.00	1054.0
189.90	195.0
191.00	321.0
192.00	1182.0
193.00	1359.0
194.10	219.0
196.00	2027.0
196.90	532.0
197.10	530.0
198.00	73376.0
198.90	4983.0
199.90	385.0
201.50	399.0
202.90	744.0
204.00	3465.0
205.00	5437.0
206.00	21224.0
207.00	3279.0
207.90	769.0
209.00	288.0
209.90	264.0
211.00	1077.0
214.90	285.0
216.00	596.0
216.90	6758.0
217.90	926.0
220.90	1978.0
221.90	616.0
223.00	1311.0
224.00	11808.0
225.00	3636.0
226.10	317.0
227.00	7198.0
228.00	1096.0
229.00	1095.0
229.90	238.0
231.00	435.0
233.90	345.0
235.00	413.0
236.00	577.0
237.00	532.0
238.80	150.0
241.00	405.0
242.00	602.0
243.00	727.0
244.00	7951.0
245.00	1121.0
246.00	2301.0
246.80	314.0
248.90	339.0
250.90	176.0
253.00	337.0
254.20	735.0
255.00	44688.0
256.00	7271.0
256.90	438.0
258.00	3168.0
258.90	497.0
265.00	1719.0
265.90	163.0
272.00	203.0
273.00	1489.0
274.00	3883.0
275.00	18352.0
276.00	2555.0
276.90	2110.0
277.90	514.0

Instrument :
BNA_M
ClientSampleId :
DFTPP89

283.00	296.0
283.90	155.0
284.90	329.0
293.00	361.0
295.10	151.0
296.00	6224.0
297.00	960.0
303.00	701.0
304.00	232.0
314.00	266.0
315.00	682.0
316.00	379.0
320.90	150.0
323.00	1585.0
324.00	261.0
326.90	467.0
328.00	211.0
333.00	183.0
334.00	1006.0
335.00	225.0
341.10	165.0
346.00	366.0
351.90	376.0
352.90	357.0
354.00	696.0
365.00	2472.0
365.90	608.0
372.00	707.0
373.00	230.0
382.90	255.0
402.00	370.0
402.90	479.0
403.90	230.0
421.00	563.0
422.00	385.0
423.00	2422.0
424.00	780.0
441.00	7654.0
442.00	45064.0
443.00	9422.0
444.00	736.0

Instrument :
BNA_M
ClientSampleId :
DFTPP89

m/z	Abundance
37.00	245.0
38.00	725.0
39.10	4642.0
40.00	348.0
43.90	242.0
45.00	150.0
49.00	458.0
50.00	17440.0
51.10	61352.0
52.00	3315.0
53.00	196.0
55.00	408.0
56.00	1788.0
57.00	4273.0
58.00	201.0
61.00	778.0
62.00	957.0
63.00	2349.0
64.00	359.0
65.00	1060.0
68.10	1041.0
69.00	58032.0
70.00	230.0
73.00	748.0
74.00	6300.0
75.00	9979.0
76.10	3205.0
77.10	56016.0
78.10	3383.0
79.00	4362.0
80.00	3420.0
81.00	4928.0
82.00	1132.0
83.00	992.0
83.90	160.0
84.90	723.0
86.00	1597.0
86.90	497.0
88.00	216.0
91.00	1123.0
92.00	1554.0
93.00	9556.0
94.00	606.0
95.90	339.0
98.00	6501.0
99.00	5264.0
99.90	381.0
101.00	2576.0
102.90	665.0
104.00	1889.0
105.00	1793.0
106.00	452.0
107.00	21224.0
108.00	2723.0
108.90	450.0
110.00	29312.0
111.00	4967.0
111.90	583.0
116.00	1298.0
117.00	21872.0
118.00	1420.0
118.90	170.0
119.90	175.0
122.00	1428.0
122.90	1927.0
124.00	906.0
124.90	920.0
126.20	340.0
127.00	53184.0
128.00	4208.0
129.00	24048.0
130.00	2330.0
130.90	380.0
132.00	203.0
133.90	866.0
135.00	1865.0
135.90	883.0
137.00	1258.0
137.90	167.0
139.90	433.0
141.00	3757.0
142.00	1063.0
143.00	656.0
144.10	200.0
146.00	538.0
146.90	1942.0
148.00	4510.0
149.00	840.0
149.90	217.0
151.00	386.0
151.80	271.0
153.00	1092.0

Instrument :
BNA_M
ClientSampleId :
DFTPP89

153.90	765.0
155.00	1864.0
156.00	2656.0
156.90	389.0
158.00	562.0
159.00	611.0
160.00	1185.0
161.00	1368.0
162.00	652.0
164.90	1427.0
166.00	930.0
167.00	7225.0
168.00	3750.0
169.00	610.0
169.80	222.0
170.90	447.0
172.00	649.0
172.90	931.0
174.00	1341.0
175.00	2629.0
176.10	534.0
177.00	1338.0
178.00	485.0
178.90	4770.0
180.00	2772.0
181.00	1616.0
182.20	243.0
184.00	460.0
185.00	2692.0
186.00	17448.0
187.00	4925.0
187.90	441.0
188.90	1644.0
189.90	254.0
191.00	802.0
192.00	1532.0
193.00	2109.0
193.90	408.0
194.90	343.0
196.00	2776.0
197.10	739.0
198.00	105520.0
198.90	6908.0
199.90	598.0
201.50	415.0
202.90	1296.0
204.00	4404.0
205.00	8217.0
206.00	31608.0
207.00	4471.0
207.90	1565.0
208.90	425.0
210.20	579.0
211.00	1417.0
212.00	190.0
214.80	583.0
215.90	743.0
216.90	9556.0
217.90	1522.0
220.30	334.0
221.00	3332.0
222.00	1186.0
223.00	2017.0
224.00	18144.0
225.00	4344.0
225.90	485.0
227.00	10040.0
227.90	1396.0
228.90	1530.0
229.90	288.0
231.00	801.0
231.90	179.0
233.90	686.0
234.90	701.0
236.00	471.0
237.00	659.0
239.00	348.0
240.00	304.0
241.00	558.0
242.00	797.0
242.90	1204.0
244.00	11997.0
245.00	1533.0
246.00	2760.0
246.90	574.0
248.10	171.0
249.00	394.0
251.90	181.0
252.80	276.0
254.10	848.0
255.00	68448.0
256.00	9701.0
257.00	745.0
257.90	4817.0
259.00	768.0
265.00	2578.0

Instrument :
BNA_M
ClientSampleId :
DFTPP89

265.90	301.0
270.90	182.0
272.00	222.0
273.00	1751.0
274.00	4457.0
275.00	28344.0
276.00	3407.0
277.00	2953.0
278.00	652.0
279.00	157.0
283.00	457.0
284.00	179.0
285.00	353.0
292.90	607.0
293.90	203.0
295.10	332.0
296.00	10090.0
297.00	1390.0
303.10	1117.0
304.00	220.0
314.00	395.0
315.00	1026.0
315.90	557.0
320.90	276.0
323.00	2500.0
324.00	371.0
326.90	655.0
327.90	307.0
332.00	201.0
332.90	249.0
334.00	1747.0
335.00	315.0
341.00	416.0
345.90	661.0
351.90	863.0
352.90	479.0
354.00	816.0
355.00	229.0
365.00	4955.0
365.90	554.0
370.80	170.0
372.00	1352.0
372.90	360.0
383.00	423.0
402.00	610.0
403.00	699.0
403.80	293.0
421.00	762.0
422.00	544.0
423.00	4504.0
424.00	1099.0
441.00	13379.0
442.00	81472.0
443.00	16512.0
444.00	1515.0

Instrument :
BNA_M
ClientSampleId :
DFTPP89

m/z	Abundance
37.00	165.0
38.10	544.0
39.10	3408.0
39.90	277.0
48.90	273.0
50.10	11961.0
51.10	40976.0
52.00	2213.0
55.00	268.0
56.00	1353.0
57.00	2740.0
61.00	649.0
62.00	598.0
63.00	1735.0
64.10	223.0
65.10	720.0
68.10	718.0
69.00	38440.0
70.00	194.0
72.90	589.0
74.00	4875.0
75.00	6494.0
76.10	1804.0
77.10	37536.0
78.10	2665.0
79.00	3433.0
80.00	2485.0
81.00	3409.0
82.00	882.0
83.00	640.0
84.90	433.0
86.00	1122.0
87.00	457.0
87.90	172.0
91.00	1042.0
92.10	911.0
93.00	6455.0
93.90	437.0
95.90	246.0
98.00	5139.0
99.00	3279.0
100.00	211.0
101.00	1752.0
103.00	540.0
104.00	1486.0
104.90	1146.0
106.00	440.0
107.00	15099.0
108.00	2084.0
109.20	418.0
110.00	20936.0
111.00	3361.0
112.10	462.0
116.00	1147.0
117.00	16102.0
118.00	971.0
122.00	1042.0
122.90	1768.0
124.00	737.0
124.90	650.0
125.70	239.0
127.00	40928.0
128.00	3043.0
129.00	17120.0
130.00	1436.0
131.00	381.0
132.90	227.0
133.90	743.0
135.10	1104.0
135.90	481.0
137.10	725.0
139.80	173.0
141.00	2599.0
141.90	793.0
142.90	503.0
144.00	164.0
146.00	540.0
147.00	981.0
148.00	3695.0
149.00	648.0
151.00	210.0
151.90	336.0
152.90	861.0
153.90	562.0
155.00	1328.0
156.00	2139.0
156.90	257.0
157.90	568.0
159.00	393.0
160.00	1013.0
161.00	1207.0
162.00	413.0

Instrument :
BNA_M
ClientSampleId :
DFTPP89

165.00	1152.0
166.00	673.0
167.00	4737.0
168.00	2457.0
169.00	420.0
170.80	184.0
172.00	527.0
173.00	630.0
174.00	1039.0
175.00	2177.0
176.00	380.0
177.00	997.0
177.90	410.0
179.00	4260.0
179.90	2679.0
180.90	1065.0
181.80	156.0
182.20	155.0
183.90	345.0
185.00	1811.0
186.00	14223.0
187.00	4000.0
188.00	410.0
189.00	1136.0
189.90	240.0
190.90	633.0
192.00	1186.0
193.00	1443.0
194.00	308.0
195.90	1816.0
197.10	711.0
198.00	81224.0
199.00	5380.0
199.90	477.0
201.40	275.0
202.00	164.0
202.90	935.0
204.00	3921.0
205.00	6890.0
206.00	25112.0
207.00	3385.0
208.00	998.0
208.90	347.0
210.00	505.0
211.00	1323.0
211.80	188.0
214.90	404.0
215.90	799.0
217.00	8071.0
217.90	1080.0
221.00	2794.0
221.90	880.0
223.00	1907.0
224.00	14305.0
225.00	3906.0
226.10	466.0
227.00	7526.0
228.00	1056.0
228.90	1454.0
230.00	169.0
230.90	408.0
234.00	567.0
235.00	516.0
236.00	536.0
236.90	817.0
238.90	233.0
240.00	249.0
240.90	580.0
241.90	825.0
243.00	757.0
244.00	10331.0
245.00	1504.0
246.00	3023.0
246.90	610.0
248.90	471.0
252.90	353.0
254.10	728.0
255.00	59288.0
256.00	8095.0
256.90	996.0
258.00	4389.0
259.00	614.0
259.90	158.0
264.10	151.0
264.90	2068.0
265.80	223.0
271.90	243.0
273.00	1677.0
274.00	5021.0
275.00	23664.0
276.00	3251.0
277.00	2542.0
277.90	257.0
282.90	322.0
284.10	276.0
285.10	403.0

Instrument :
BNA_M
ClientSampleId :
DFTPP89

288.90	153.0
292.00	221.0
292.90	741.0
294.90	209.0
296.00	9022.0
296.90	1399.0
303.00	912.0
303.90	214.0
309.00	182.0
314.00	422.0
315.00	992.0
316.00	375.0
320.90	282.0
323.00	2712.0
324.00	274.0
327.00	485.0
327.90	231.0
331.90	157.0
333.00	176.0
334.00	1897.0
335.00	390.0
340.90	283.0
345.90	453.0
352.00	860.0
353.00	555.0
353.90	674.0
355.00	181.0
365.00	4926.0
365.90	482.0
371.10	216.0
372.00	1084.0
372.80	401.0
383.00	448.0
384.00	169.0
390.00	175.0
391.00	174.0
401.90	472.0
403.00	715.0
404.00	245.0
421.00	745.0
422.00	554.0
423.00	4685.0
424.00	926.0
441.00	13833.0
442.00	86472.0
443.00	16318.0
443.90	1307.0
444.90	208.0

Instrument :
BNA_M
ClientSampleId :
DFTPP89