




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111122\
 Data File : BM037459.D
 Acq On : 11 Nov 2022 09:53
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 12 05:04:18 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 11 22:49:50 2022
 Response via : Initial Calibration



TIC: BM037459.D\data.ms

(70) Pentachlorophenol (C)




16.851min (-0.000) 25006.56 ng m

response 669508

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.40	64.45
264.00	61.40	63.63
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111122\
Data File : BM037459.D
Acq On : 11 Nov 2022 09:53
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 12 05:04:18 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 11 22:49:50 2022
Response via : Initial Calibration



TIC: BM037459.D\data.ms

(77) Benzidine

19.456min (-0.000) 726208.32 ng m

response 3042885

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.10	13.92
183.00	11.20	11.39
0.00	0.00	0.00