

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111122\
 Data File : BM037480.D
 Acq On : 11 Nov 2022 23:07
 Operator : CG/JU
 Sample : N5530-05
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SB-11

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 11/12/2022
 Supervised By :Jagrut Upadhyay 11/12/2022

Quant Time: Nov 12 00:42:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 11 22:49:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.804	152	223142	20.000	ng	0.00
21) Naphthalene-d8	10.604	136	852669	20.000	ng	0.00
39) Acenaphthene-d10	14.445	164	485357	20.000	ng	0.00
64) Phenanthrene-d10	17.192	188	910422	20.000	ng	0.00
76) Chrysene-d12	21.380	240	870453	20.000	ng	0.00
86) Perylene-d12	23.721	264	973770	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.375	112	1826349	141.402	ng	0.00
7) Phenol-d6	6.987	99	2297491	131.062	ng	0.00
23) Nitrobenzene-d5	8.975	82	1408656	83.350	ng	0.00
42) 2,4,6-Tribromophenol	15.939	330	907356	158.635	ng	0.00
45) 2-Fluorobiphenyl	13.069	172	3153980	86.342	ng	0.00
79) Terphenyl-d14	19.821	244	4464586	91.978	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.252	88	65	0.012	ng	# 1
3) Pyridine	3.675	79	153	0.010	ng	# 1
4) n-Nitrosodimethylamine	3.575	42	252	0.037	ng	# 1
6) Aniline	7.146	93	376	0.018	ng	# 1
8) 2-Chlorophenol	7.387	128	122	0.008	ng	# 23
9) Benzaldehyde	6.981	77	4314	0.486	ng	# 6
10) Phenol	7.010	94	13599	0.692	ng	# 17
11) bis(2-Chloroethyl)ether	7.246	93	173	0.011	ng	# 63
15) Benzyl Alcohol	8.075	79	360	0.029	ng	# 49
16) 2,2'-oxybis(1-Chloropr...	8.340	45	174	0.008	ng	# 1
17) 2-Methylphenol	8.193	107	57	0.004	ng	# 1
18) Hexachloroethane	8.869	117	114	0.018	ng	# 3
19) n-Nitroso-di-n-propyla...	8.622	70	531	0.117	ng	# 1
20) 3+4-Methylphenols	8.687	107	62	0.004	ng	# 1
22) Acetophenone	8.645	105	1493	0.067	ng	# 1
24) Nitrobenzene	9.045	77	190	0.011	ng	# 29
25) Isophorone	9.528	82	199	0.007	ng	# 65
27) 2,4-Dimethylphenol	9.957	122	6762	0.594	ng	# 4
28) bis(2-Chloroethoxy)met...	10.039	93	120	0.007	ng	# 50
31) Naphthalene	10.657	128	646	0.013	ng	# 69
32) Benzoic acid	9.957	122	18716m	8.521	ng	
33) 4-Chloroaniline	10.928	127	60	0.003	ng	# 18
35) Caprolactam	11.651	113	74	0.019	ng	# 32
36) 4-Chloro-3-methylphenol	11.963	107	240	0.018	ng	# 17
46) 1,1'-Biphenyl	13.069	154	6168	0.157	ng	# 1
48) 2-Nitroaniline	13.551	65	74	0.008	ng	# 4
49) Acenaphthylene	14.169	152	788	0.016	ng	# 61
50) Dimethylphthalate	13.927	163	2307	0.062	ng	# 77
51) 2,6-Dinitrotoluene	14.039	165	53	0.007	ng	# 22
52) Acenaphthene	14.527	154	78	0.003	ng	# 4
55) Dibenzofuran	14.857	168	644	0.014	ng	# 76
56) 4-Nitrophenol	14.857	139	119	0.017	ng	# 61
57) 2,4-Dinitrotoluene	14.751	165	59	0.006	ng	# 22
58) Fluorene	15.498	166	445	0.012	ng	# 81
60) Diethylphthalate	15.274	149	11061	0.307	ng	97
61) 4-Chlorophenyl-phenyle...	15.498	204	79	0.004	ng	# 41

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111122\
 Data File : BM037480.D
 Acq On : 11 Nov 2022 23:07
 Operator : CG/JU
 Sample : N5530-05
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SB-11


Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 11/12/2022
 Supervised By :Jagrut Upadhyay 11/12/2022

Quant Time: Nov 12 00:42:22 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102822.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 11 22:49:50 2022
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
63) Azobenzene	15.780	77	345	0.010	ng	# 64
66) n-Nitrosodiphenylamine	15.939	169	25904	0.842	ng	# 47
67) 4-Bromophenyl-phenylether	16.392	248	54	0.005	ng	# 1
68) Hexachlorobenzene	16.492	284	122	0.010	ng	# 42
70) Pentachlorophenol	16.874	266	323	0.045	ng	# 24
71) Phenanthrene	17.239	178	4492	0.080	ng	# 86
72) Anthracene	17.333	178	1704	0.032	ng	# 82
73) Carbazole	17.639	167	1211	0.025	ng	# 52
74) Di-n-butylphthalate	18.168	149	12316	0.227	ng	# 98
75) Fluoranthene	19.262	202	6976	0.117	ng	100
77) Benzidine	19.498	184	126	0.010	ng	# 1
78) Pyrene	19.621	202	7634	0.114	ng	93
80) Butylbenzylphthalate	20.439	149	351	0.015	ng	86
81) Benzo(a)anthracene	21.368	228	7006	0.113	ng	# 91
82) 3,3'-Dichlorobenzidine	21.321	252	674	0.037	ng	# 57
83) Chrysene	21.415	228	6353	0.106	ng	# 88
84) Bis(2-ethylhexyl)phtha...	21.298	149	4448	0.139	ng	# 89
85) Di-n-octyl phthalate	22.198	149	2144	0.042	ng	# 67
87) Indeno(1,2,3-cd)pyrene	26.138	276	14472	0.182	ng	# 68
88) Benzo(b)fluoranthene	23.015	252	6851	0.101	ng	# 30
89) Benzo(k)fluoranthene	23.062	252	4024	0.058	ng	# 12
90) Benzo(a)pyrene	23.609	252	6229	0.113	ng	# 15
91) Dibenzo(a,h)anthracene	26.150	278	11469	0.173	ng	# 82
92) Benzo(g,h,i)perylene	26.885	276	16234	0.252	ng	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 7.804 min Scan# 80
Delta R.T. 0.000 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Instrument :
BNA_M
ClientSampleId :
SB-11



Tgt Ion:152 Resp: 223142
Ion Ratio Lower Upper
152 100
150 155.1 124.8 187.2
115 56.1 47.0 70.4




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022



#2
1,4-Dioxane
Concen: 0.012 ng
RT: 3.252 min Scan# 28
Delta R.T. 0.000 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07


Tgt Ion: 88 Resp: 65
Ion Ratio Lower Upper
88 100
58 587.7 63.5 95.3#
43 0.0 24.2 36.4#





#3
Pyridine
Concen: 0.010 ng
RT: 3.675 min Scan# 10
Delta R.T. 0.012 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Instrument :
BNA_M
ClientSampleId :
SB-11




Tgt Ion: 79 Resp: 153
Ion Ratio Lower Upper
79 100
52 33.3 57.8 86.6
51 339.2 29.0 43.6

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022




#4
n-Nitrosodimethylamine
Concen: 0.037 ng
RT: 3.575 min Scan# 83
Delta R.T. -0.006 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Tgt Ion: 42 Resp: 252
Ion Ratio Lower Upper
42 100
74 16.6 106.2 159.2#
44 134.9 5.4 8.0#





#5
2-Fluorophenol
Concen: 141.402 ng
RT: 5.375 min Scan# 389
Delta R.T. 0.000 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Instrument :
BNA_M
ClientSampleId :
SB-11




Tgt Ion: 112 Resp: 1826349
Ion Ratio Lower Upper
112 100
64 58.8 46.8 70.2
63 31.2 24.4 36.6

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022




#6
Aniline
Concen: 0.018 ng
RT: 7.146 min Scan# 690
Delta R.T. 0.000 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Tgt Ion: 93 Resp: 376
Ion Ratio Lower Upper
93 100
66 115.2 30.4 45.6#
65 95.7 16.2 24.2#





#7
Phenol-d6
Concen: 131.062 ng
RT: 6.987 min Scan# 60
Delta R.T. -0.006 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Instrument :
BNA_M
ClientSampleId :
SB-11




Tgt Ion: 99 Resp: 2297491
Ion Ratio Lower Upper
99 100
42 18.6 14.8 22.2
71 32.9 25.6 38.4

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022




#8
2-Chlorophenol
Concen: 0.008 ng
RT: 7.387 min Scan# 731
Delta R.T. 0.012 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Tgt Ion:128 Resp: 122
Ion Ratio Lower Upper
128 100
130 0.0 12.7 52.7#
64 106.6 26.4 66.4#





#9
Benzaldehyde
Concen: 0.486 ng
RT: 6.981 min Scan# 60
Delta R.T. 0.024 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Instrument :
BNA_M
ClientSampleId :
SB-11




Tgt Ion: 77 Resp: 4314
Ion Ratio Lower Upper
77 100
105 9.2 78.6 118.6
106 0.0 75.8 115.8

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022




#10
Phenol
Concen: 0.692 ng
RT: 7.010 min Scan# 667
Delta R.T. -0.006 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Tgt Ion: 94 Resp: 13599
Ion Ratio Lower Upper
94 100
65 50.1 8.1 48.1#
66 103.2 15.5 55.5#





#11
bis(2-Chloroethyl)ether
Concen: 0.011 ng
RT: 7.246 min Scan# 707
Delta R.T. 0.006 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Instrument :
BNA_M
ClientSampleId :
SB-11



Tgt Ion: 93 Resp: 173
Ion Ratio Lower Upper
93 100
63 54.2 59.3 99.3
95 0.0 12.0 52.0




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022



#15
Benzyl Alcohol
Concen: 0.029 ng
RT: 8.075 min Scan# 848
Delta R.T. 0.012 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07

Tgt Ion: 79 Resp: 360
Ion Ratio Lower Upper
79 100
108 0.0 63.1 94.7
77 67.8 51.7 77.5









#16
2,2'-oxybis(1-Chloropropane)
Concen: 0.008 ng
RT: 8.340 min Scan# 893
Delta R.T. -0.006 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07

Instrument :
BNA_M
ClientSampleId :
SB-11

Tgt Ion: 45 Resp: 174
Ion Ratio Lower Upper
45 100
77 129.5 0.0 34.0#
79 115.8 0.0 30.6#




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022



#17
2-Methylphenol
Concen: 0.004 ng
RT: 8.193 min Scan# 868
Delta R.T. -0.076 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07


Tgt Ion:107 Resp: 57
Ion Ratio Lower Upper
107 100
108 0.0 89.5 134.3#
77 214.8 36.2 54.2#
79 184.0 35.0 52.6#





#18
Hexachloroethane
Concen: 0.018 ng
RT: 8.869 min Scan# 983
Delta R.T. -0.006 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Instrument :
BNA_M
ClientSampleId :
SB-11



Tgt Ion: 117 Resp: 114
Ion Ratio Lower Upper
117 100
119 0.0 75.9 113.9#
201 0.0 75.0 112.4#




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022



#19
n-Nitroso-di-n-propylamine
Concen: 0.117 ng
RT: 8.622 min Scan# 941
Delta R.T. -0.006 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07


Tgt Ion: 70 Resp: 531
Ion Ratio Lower Upper
70 100
42 171.3 46.4 69.6#
101 0.0 7.8 11.8#
130 0.0 17.0 25.4#





#20
3+4-Methylphenols
Concen: 0.004 ng
RT: 8.687 min Scan# 91
Delta R.T. 0.088 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Instrument :
BNA_M
ClientSampleId :
SB-11



Tgt Ion:107 Resp: 62
Ion Ratio Lower Upper
107 100
108 0.0 69.4 109.4
77 264.4 12.8 52.8
79 129.4 8.4 48.4




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022



#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.604 min Scan# 1278
Delta R.T. -0.006 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07


Tgt Ion:136 Resp: 852669
Ion Ratio Lower Upper
136 100
137 10.9 9.0 13.4
54 7.8 7.0 10.4
68 6.0 5.4 8.0





#22
Acetophenone
Concen: 0.067 ng
RT: 8.645 min Scan# 94
Delta R.T. 0.000 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Instrument :
BNA_M
ClientSampleId :
SB-11



Tgt Ion	105	Resp	149
Ion Ratio	100	Lower	Upper
71	25.0	3.0	4.4
51	143.5	22.1	33.1
120	21.5	18.2	27.2




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022



#23
Nitrobenzene-d5
Concen: 83.350 ng
RT: 8.975 min Scan# 1001
Delta R.T. -0.006 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07

Tgt Ion	82	Resp	1408656
Ion Ratio	100	Lower	Upper
128	43.8	35.2	52.8
54	50.3	40.4	60.6









#24
Nitrobenzene
Concen: 0.011 ng
RT: 9.045 min Scan# 1095
Delta R.T. 0.024 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07

Instrument :
BNA_M
ClientSampleId :
SB-11

Tgt Ion: 77 Resp: 190
Ion Ratio Lower Upper
77 100
123 0.0 36.6 55.0#
65 43.5 10.6 16.0#




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022



#25
Isophorone
Concen: 0.007 ng
RT: 9.528 min Scan# 1095
Delta R.T. -0.023 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07

Tgt Ion: 82 Resp: 199
Ion Ratio Lower Upper
82 100
95 0.0 5.4 8.0#
138 0.0 14.2 21.4#





#27

2,4-Dimethylphenol

Concen: 0.594 ng

RT: 9.957 min Scan# 1151

Delta R.T. 0.159 min

Lab File: BM037480.D

Acq: 11 Nov 2022 23:07

Instrument :

BNA_M

ClientSampleId :

SB-11

Tgt Ion:122 Resp: 6762

Ion Ratio Lower Upper

122 100

107 0.0 89.7 134.5





121 0.0 47.1 70.7

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 11/12/2022

Supervised By :Jagrut Upadhyay 11/12/2022



#28

bis(2-Chloroethoxy)methane

Concen: 0.007 ng

RT: 10.039 min Scan# 1182

Delta R.T. 0.006 min

Lab File: BM037480.D

Acq: 11 Nov 2022 23:07




Tgt Ion: 93 Resp: 120


Ion Ratio Lower Upper

93 100

95 0.0 25.8 38.8#


123 0.0 9.3 13.9#





#31
Naphthalene
Concen: 0.013 ng
RT: 10.657 min Scan# 1168
Delta R.T. 0.000 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Instrument :
BNA_M
ClientSampleId :
SB-11



Tgt Ion:128 Resp: 640
Ion Ratio Lower Upper
128 100
129 0.0 8.7 13.1
127 0.0 10.2 15.2




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022



#32
Benzoic acid
Concen: 8.521 ng m
RT: 9.957 min Scan# 1168
Delta R.T. -0.023 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07


Tgt Ion:122 Resp: 18716
Ion Ratio Lower Upper
122 100
105 122.5 100.1 140.1
77 104.4 71.7 111.7





#33
4-Chloroaniline
Concen: 0.003 ng
RT: 10.928 min Scan# 1319
Delta R.T. 0.141 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Instrument :
BNA_M
ClientSampleId :
SB-11




Tgt Ion:127 Resp: 60
Ion Ratio Lower Upper
127 100
129 0.0 25.5 38.3#
65 101.8 24.0 36.0#
92 0.0 14.0 21.0#

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022




#35
Caprolactam
Concen: 0.019 ng
RT: 11.651 min Scan# 1456
Delta R.T. 0.035 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Tgt Ion:113 Resp: 74
Ion Ratio Lower Upper
113 100
55 296.2 156.9 196.9#
56 175.8 106.4 146.4#





#36
4-Chloro-3-methylphenol
Concen: 0.018 ng
RT: 11.963 min Scan# 1193
Delta R.T. 0.041 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Instrument :
BNA_M
ClientSampleId :
SB-11



Tgt Ion:107 Resp: 240
Ion Ratio Lower Upper
107 100
144 0.0 21.8 32.6
142 0.0 68.1 102.1




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022



#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.445 min Scan# 1931
Delta R.T. -0.006 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07


Tgt Ion:164 Resp: 485357
Ion Ratio Lower Upper
164 100
162 102.5 80.6 120.8
160 44.5 36.2 54.4





#42
2,4,6-Tribromophenol
Concen: 158.635 ng
RT: 15.939 min Scan# 2195
Delta R.T. -0.006 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Instrument :
BNA_M
ClientSampleId :
SB-11




Tgt Ion:330 Resp: 907356
Ion Ratio Lower Upper
330 100
332 96.7 78.0 117.0
141 32.2 27.5 41.3

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022




#45
2-Fluorobiphenyl
Concen: 86.342 ng
RT: 13.069 min Scan# 1697
Delta R.T. -0.006 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Tgt Ion:172 Resp: 3153980
Ion Ratio Lower Upper
172 100
171 35.2 28.1 42.1
170 23.6 19.0 28.4





#46
1,1'-Biphenyl
Concen: 0.157 ng
RT: 13.069 min Scan# 10
Delta R.T. -0.217 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Instrument :
BNA_M
ClientSampleId :
SB-11



Tgt Ion:154 Resp: 6168
Ion Ratio Lower Upper
154 100
153 723.1 20.3 60.3#
76 1480.0 0.0 33.7




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022



#48
2-Nitroaniline
Concen: 0.008 ng
RT: 13.551 min Scan# 1779
Delta R.T. 0.000 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07


Tgt Ion: 65 Resp: 74
Ion Ratio Lower Upper
65 100
92 0.0 53.7 80.5#
138 0.0 87.4 131.2#





#49
Acenaphthylene
Concen: 0.016 ng
RT: 14.169 min Scan# 1884
Delta R.T. -0.006 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Instrument :
BNA_M
ClientSampleId :
SB-11




Tgt Ion:152 Resp: 788
Ion Ratio Lower Upper
152 100
151 0.0 16.0 24.0#
153 0.0 10.6 15.8#

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022




#50
Dimethylphthalate
Concen: 0.062 ng
RT: 13.927 min Scan# 1843
Delta R.T. 0.000 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Tgt Ion:163 Resp: 2307
Ion Ratio Lower Upper
163 100
194 0.0 3.0 4.6#
164 0.0 8.2 12.2#





#51
2,6-Dinitrotoluene
Concen: 0.007 ng
RT: 14.039 min Scan# 1872
Delta R.T. -0.012 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Instrument :
BNA_M
ClientSampleId :
SB-11



Tgt Ion:165 Resp: 53
Ion Ratio Lower Upper
165 100
63 0.0 46.2 69.2#
89 0.0 46.6 69.8#




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022



#52
Acenaphthene
Concen: 0.003 ng
RT: 14.527 min Scan# 1945
Delta R.T. 0.012 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07


Tgt Ion:154 Resp: 78
Ion Ratio Lower Upper
154 100
153 0.0 91.5 137.3#
152 0.0 43.0 64.6#





#55
Dibenzenofuran
Concen: 0.014 ng
RT: 14.857 min Scan# 2001
Delta R.T. 0.006 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Instrument :
BNA_M
ClientSampleId :
SB-11



Tgt Ion:168 Resp: 644
Ion Ratio Lower Upper
168 100
139 48.5 29.3 43.9#
169 0.0 10.6 16.0#




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022



#56
4-Nitrophenol
Concen: 0.017 ng
RT: 14.857 min Scan# 2001
Delta R.T. 0.153 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07


Tgt Ion:139 Resp: 119
Ion Ratio Lower Upper
139 100
109 0.0 47.4 87.4#
65 87.9 73.8 113.8





#57
2,4-Dinitrotoluene
Concen: 0.006 ng
RT: 14.751 min Scan# 1983
Delta R.T. -0.088 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Instrument :
BNA_M
ClientSampleId :
SB-11



Tgt Ion:165 Resp: 59
Ion Ratio Lower Upper
165 100
63 0.0 36.3 54.5
89 0.0 56.3 84.5
182 0.0 2.1 3.1#




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022



#58
Fluorene
Concen: 0.012 ng
RT: 15.498 min Scan# 2110
Delta R.T. -0.006 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07

Tgt Ion:166 Resp: 445
Ion Ratio Lower Upper
166 100
165 81.0 78.2 117.4
167 0.0 10.6 15.8#









#60
Diethylphthalate
Concen: 0.307 ng
RT: 15.274 min Scan# 2082
Delta R.T. -0.006 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07

Instrument :
BNA_M
ClientSampleId :
SB-11

Tgt Ion:149 Resp: 11061
Ion Ratio Lower Upper
149 100
177 23.3 17.4 26.0
150 13.2 10.0 15.0




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022



#61
4-Chlorophenyl-phenylether
Concen: 0.004 ng
RT: 15.498 min Scan# 2110
Delta R.T. 0.000 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07


Tgt Ion:204 Resp: 79
Ion Ratio Lower Upper
204 100
206 0.0 26.6 39.8#
141 99.6 45.2 67.8#





#63
Azobenzene
Concen: 0.010 ng
RT: 15.780 min Scan# 2169
Delta R.T. -0.012 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Instrument :
BNA_M
ClientSampleId :
SB-11



Tgt Ion: 77 Resp: 345
Ion Ratio Lower Upper
77 100
182 0.0 5.1 45.1
105 23.9 0.5 40.5
51 52.7 9.0 49.0




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022



#64
Phenanthrene-d10
Concen: 20.000 ng
RT: 17.192 min Scan# 2398
Delta R.T. -0.006 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07

Tgt Ion:188 Resp: 910422
Ion Ratio Lower Upper
188 100
94 9.7 7.9 11.9
80 10.8 8.5 12.7





#66

n-Nitrosodiphenylamine

Concen: 0.842 ng

RT: 15.939 min Scan# 21

Delta R.T. 0.224 min

Lab File: BM037480.D

Acq: 11 Nov 2022 23:07

Instrument :

BNA_M

ClientSampleId :

SB-11





Tgt Ion:169	Resp:	25904
Ion Ratio	Lower	Upper
169	100	
168	7.7	54.8 82.2
167	43.9	29.5 44.3

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 11/12/2022

Supervised By :Jagrut Upadhyay 11/12/2022



#67

4-Bromophenyl-phenylether

Concen: 0.005 ng




RT: 16.392 min Scan# 2262


Delta R.T. 0.000 min

Lab File: BM037480.D

Acq: 11 Nov 2022 23:07


Tgt Ion:248	Resp:	54
Ion Ratio	Lower	Upper
248	100	
250	134.0	77.8 116.6
141	302.6	54.2 81.2





#68
Hexachlorobenzene
Concen: 0.010 ng
RT: 16.492 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Instrument :
BNA_M
ClientSampleId :
SB-11




Tgt Ion:284 Resp: 122
Ion Ratio Lower Upper
284 100
142 0.0 28.2 42.4#
249 0.0 23.1 34.7#

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022




#70
Pentachlorophenol
Concen: 0.045 ng
RT: 16.874 min Scan# 2344
Delta R.T. 0.024 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07



Tgt Ion:266 Resp: 323
Ion Ratio Lower Upper
266 100
268 120.1 50.7 76.1#
264 0.0 49.1 73.7#









#71
Phenanthrene
Concen: 0.080 ng
RT: 17.239 min Scan# 2415
Delta R.T. -0.006 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07

Instrument :
BNA_M
ClientSampleId :
SB-11

Tgt Ion:178 Resp: 4492
Ion Ratio Lower Upper
178 100
176 25.2 15.6 23.4#
179 21.7 12.4 18.6#




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022



#72
Anthracene
Concen: 0.032 ng
RT: 17.333 min Scan# 2422
Delta R.T. 0.000 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07


Tgt Ion:178 Resp: 1704
Ion Ratio Lower Upper
178 100
176 25.0 15.4 23.0#
179 26.1 12.6 19.0#





#73
 Carbazole
 Concen: 0.025 ng
 RT: 17.639 min Scan# 2477
 Delta R.T. 0.030 min
 Lab File: BM037480.D
 Acq: 11 Nov 2022 23:07




Instrument :
 BNA_M
 ClientSampleId :
 SB-11



Tgt Ion:167 Resp: 1211
 Ion Ratio Lower Upper
 167 100
 166 43.1 17.4 26.2
 139 34.6 10.4 15.6




Manual Integrations
 APPROVED


Reviewed By :Christian Giraldo 11/12/2022
 Supervised By :Jagrut Upadhyay 11/12/2022



#74
 Di-n-butylphthalate
 Concen: 0.227 ng
 RT: 18.168 min Scan# 2564
 Delta R.T. 0.000 min
 Lab File: BM037480.D
 Acq: 11 Nov 2022 23:07


Tgt Ion:149 Resp: 12316
 Ion Ratio Lower Upper
 149 100
 150 9.6 7.4 11.0
 104 6.2 4.0 6.0





#75
Fluoranthene
Concen: 0.117 ng
RT: 19.262 min Scan# 21
Delta R.T. 0.006 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Instrument :
BNA_M
ClientSampleId :
SB-11



Tgt Ion:202 Resp: 6976
Ion Ratio Lower Upper
202 100
101 12.5 0.0 32.4
203 18.0 0.0 37.8




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022



#76
Chrysene-d12
Concen: 20.000 ng
RT: 21.380 min Scan# 3110
Delta R.T. -0.006 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07


Tgt Ion:240 Resp: 870453
Ion Ratio Lower Upper
240 100
120 11.0 9.9 14.9
236 26.1 20.3 30.5





#77
Benzidine
Concen: 0.010 ng
RT: 19.498 min Scan# 2790
Delta R.T. 0.041 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Instrument :
BNA_M
ClientSampleId :
SB-11




Tgt Ion:184 Resp: 126
Ion Ratio Lower Upper
184 100
185 112.2 11.3 16.9
183 0.0 9.0 13.4

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022




#78
Pyrene
Concen: 0.114 ng
RT: 19.621 min Scan# 2811
Delta R.T. 0.000 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Tgt Ion:202 Resp: 7634
Ion Ratio Lower Upper
202 100
200 18.3 17.0 25.6
203 14.1 14.1 21.1





#79
 Terphenyl-d14
 Concen: 91.978 ng
 RT: 19.821 min Scan# 2853
 Delta R.T. -0.006 min
 Lab File: BM037480.D
 Acq: 11 Nov 2022 23:07




Instrument :
 BNA_M
 ClientSampleId :
 SB-11




Tgt Ion:244 Resp: 4464586
 Ion Ratio Lower Upper
 244 100
 212 6.8 5.4 8.2
 122 11.2 9.4 14.2

Manual Integrations
 APPROVED



Reviewed By :Christian Giraldo 11/12/2022
 Supervised By :Jagrut Upadhyay 11/12/2022




#80
 Butylbenzylphthalate
 Concen: 0.015 ng
 RT: 20.439 min Scan# 2950
 Delta R.T. -0.082 min
 Lab File: BM037480.D
 Acq: 11 Nov 2022 23:07



Tgt Ion:149 Resp: 351
 Ion Ratio Lower Upper
 149 100
 91 57.0 56.2 84.4
 206 22.8 15.4 23.0









#81
Benzo(a)anthracene
Concen: 0.113 ng
RT: 21.368 min Scan# 3115
Delta R.T. 0.000 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07

Instrument :
BNA_M
ClientSampleId :
SB-11

Tgt Ion:228 Resp: 7000
Ion Ratio Lower Upper
228 100
226 24.8 22.3 33.5
229 26.4 16.2 24.2




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022



#82
3,3'-Dichlorobenzidine
Concen: 0.037 ng
RT: 21.321 min Scan# 3100
Delta R.T. 0.012 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07

Tgt Ion:252 Resp: 674
Ion Ratio Lower Upper
252 100
254 44.5 52.7 79.1#
126 62.5 9.8 14.6#





#83

Chrysene

Concen: 0.106 ng

RT: 21.415 min Scan# 31

Delta R.T. -0.006 min

Lab File: BM037480.D

Acq: 11 Nov 2022 23:07

Instrument :

BNA_M

ClientSampleId :

SB-11





Tgt Ion:	228	Resp:	635
Ion Ratio	Lower	Upper	
228	100		
226	34.9	24.4	36.6
229	28.3	16.0	24.0

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 11/12/2022

Supervised By :Jagrut Upadhyay 11/12/2022



#84

Bis(2-ethylhexyl)phthalate

Concen: 0.139 ng




RT: 21.298 min Scan# 3096


Delta R.T. 0.000 min

Lab File: BM037480.D

Acq: 11 Nov 2022 23:07


Tgt Ion:	149	Resp:	4448
Ion Ratio	Lower	Upper	
149	100		
167	21.5	21.2	31.8
279	9.2	3.1	4.7





#85
Di-n-octyl phthalate
Concen: 0.042 ng
RT: 22.198 min Scan# 31
Delta R.T. 0.000 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07




Instrument :
BNA_M
ClientSampleId :
SB-11



Tgt Ion:149 Resp: 2144
Ion Ratio Lower Upper
149 100
167 26.8 1.3 1.9
43 0.0 8.0 12.0




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022



#86
Perylene-d12
Concen: 20.000 ng
RT: 23.721 min Scan# 3508
Delta R.T. 0.000 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07

Tgt Ion:264 Resp: 973770
Ion Ratio Lower Upper
264 100
260 23.7 18.9 28.3
265 21.9 17.7 26.5





#87

Indeno(1,2,3-cd)pyrene

Concen: 0.182 ng

RT: 26.138 min Scan# 3935

Delta R.T. -0.006 min

Lab File: BM037480.D

Acq: 11 Nov 2022 23:07

Instrument :

BNA_M

ClientSampleId :

SB-11





Tgt Ion:	276	Resp:	14472
Ion Ratio	Lower	Upper	
276	100		
138	0.0	21.5	32.3#
277	29.9	20.1	30.1#

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 11/12/2022

Supervised By :Jagrut Upadhyay 11/12/2022



#88

Benzo(b)fluoranthene

Concen: 0.101 ng




RT: 23.015 min Scan# 3388


Delta R.T. 0.000 min

Lab File: BM037480.D

Acq: 11 Nov 2022 23:07

Tgt Ion:	252	Resp:	6851
Ion Ratio	Lower	Upper	
252	100		
253	51.0	17.8	26.6#
125	45.6	9.1	13.7#









#89
Benzo(k)fluoranthene
Concen: 0.058 ng
RT: 23.062 min Scan# 3396
Delta R.T. 0.000 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07

Instrument :
BNA_M
ClientSampleId :
SB-11

Tgt Ion:252 Resp: 4024
Ion Ratio Lower Upper
252 100
253 52.6 17.8 26.6
125 63.9 9.1 13.7




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 11/12/2022
Supervised By :Jagrut Upadhyay 11/12/2022



#90
Benzo(a)pyrene
Concen: 0.113 ng
RT: 23.609 min Scan# 3489
Delta R.T. -0.012 min
Lab File: BM037480.D
Acq: 11 Nov 2022 23:07

Tgt Ion:252 Resp: 6229
Ion Ratio Lower Upper
252 100
253 62.0 17.4 26.2#
125 45.8 9.9 14.9#





#91

Dibenzo(a,h)anthracene

Concen: 0.173 ng

RT: 26.150 min Scan# 3937

Delta R.T. -0.012 min

Lab File: BM037480.D

Acq: 11 Nov 2022 23:07

Instrument :

BNA_M

ClientSampleId :

SB-11





Tgt Ion:	278	Resp:	11469
Ion Ratio	Lower	Upper	
278	100		
139	27.4	13.8	20.8
279	30.9	19.0	28.6

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 11/12/2022

Supervised By :Jagrut Upadhyay 11/12/2022



#92

Benzo(g,h,i)perylene

Concen: 0.252 ng

RT: 26.885 min Scan# 4046

Delta R.T. 0.000 min

Lab File: BM037480.D

Acq: 11 Nov 2022 23:07

Tgt Ion:	276	Resp:	16234
Ion Ratio	Lower	Upper	
276	100		
277	28.5	19.4	29.2
138	35.0	18.9	28.3

