


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111122\
Data File : BM037468.D
Acq On : 11 Nov 2022 15:48
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102822.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Fri Nov 11 22:49:50 2022






AutoFind: Scans 2471, 2472, 2473; Background Corrected with Scan 2464

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.5	142640	PASS
68	69	0.00	2	1.4	2034	PASS
69	198	0.00	100	39.8	147410	PASS
70	69	0.00	2	0.5	755	PASS
127	198	10	80	49.6	183595	PASS
197	198	0.00	2	0.3	1031	PASS
198	198	100	100	100.0	370027	PASS
199	198	5	9	6.6	24600	PASS
275	198	10	60	28.3	104845	PASS
365	198	1	100	3.8	13963	PASS
441	198	0.01	100	16.7	61781	PASS
442	442	50	100	100.0	402880	PASS
443	442	15	24	19.5	78691	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111122\
Data File : BM037468.D
Acq On : 11 Nov 2022 15:48
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Nov 12 05:05:49 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 11 22:49:50 2022
Response via : Initial Calibration



TIC: BM037468.D\data.ms

(70) Pentachlorophenol (C)

16.845min (-0.006) 155923.50 ng




response 475617

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.40	63.04
264.00	61.40	61.18
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111122\
Data File : BM037468.D
Acq On : 11 Nov 2022 15:48
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP

Quant Time: Nov 12 05:05:49 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM102822.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 11 22:49:50 2022
Response via : Initial Calibration



TIC: BM037468.D\data.ms

(77) Benzidine

19.456min (0.000) 73978.81 ng

response 2292297

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.10	13.85
183.00	11.20	11.54
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
11/11/2022	BNA_M	<u>BM037468.D</u>
Compound Name	Response	Retention Time
DDT	1556750	20.674
DDD	33315	20.239
DDE	1206	19.733
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
34521	1591271	2.17

Instrument :
BNA_M
ClientSampleId :
DFTPP