

Data Path : Z:\HPCHEM1\BNA_M\Data\BM111217\
 Data File : BM012930.D
 Acq On : 13 Nov 2017 11:26
 Operator : SJ/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02021

Quant Time: Nov 13 16:06:14 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM111017MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Nov 13 16:06:37 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
3	Naphthalene	1.013	0.966	4.6	99	0.00
4	2-Methylnaphthalene	0.746	0.702	5.9	97	0.00
5	1-Methylnaphthalene	0.731	0.692	5.3	98	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
7 S	Acenaphthylene-d8	2.080	2.008	3.5	99	0.00
8	Acenaphthylene	2.024	1.967	2.8	100	0.00
9 C	Acenaphthene	1.356	1.309	3.5	99	0.00
10 S	Fluorene-d10	1.423	1.375	3.4	100	0.00
11	Fluorene	1.617	1.578	2.4	100	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
13	Phenanthrene	1.105	1.055	4.5	105	0.00
14 S	Anthracene-d10	0.967	0.923	4.6	105	0.00
15	Anthracene	1.148	1.103	3.9	106	0.00
16 C	Fluoranthene	1.359	1.318	3.0	108	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	111	0.00
18 S	Pyrene-d10	0.895	0.855	4.5	108	0.00
19	Pyrene	1.148	1.112	3.1	109	0.00
20	Benzo(a)anthracene	1.231	1.155	6.2	109	0.00
21	Chrysene	1.100	1.038	5.6	110	0.00
22 I	Perylene-d12	1.000	1.000	0.0	111	0.00
23	Benzo(b)fluoranthene	1.189	1.208	-1.6	117	0.00
24	Benzo(k)fluoranthene	1.137	1.084	4.7	111	0.00
25 S	Benzo(a)pyrene-d12	0.944	0.900	4.7	112	0.00
26 C	Benzo(a)pyrene	1.153	1.098	4.8	110	0.00
27	Indeno(1,2,3-cd)pyrene	1.488	1.395	6.2	110	0.00
28	Dibenzo(a,h)anthracene	1.251	1.180	5.7	110	0.00
29	Benzo(g,h,i)perylene	1.242	1.167	6.0	110	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0