

Data Path : Z:\HPCHEM1\BNA_M\Data\BM111217\
 Data File : BM012932.D
 Acq On : 13 Nov 2017 12:47
 Operator : SJ/JU
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02022

Quant Time: Nov 13 16:15:14 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM111017MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Nov 13 16:06:37 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	88	0.00
3	Naphthalene	1.013	0.966	4.6	87	0.00
4	2-Methylnaphthalene	0.746	0.713	4.4	87	0.00
5	1-Methylnaphthalene	0.731	0.704	3.7	88	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	88	0.00
7 S	Acenaphthylene-d8	2.080	2.025	2.6	88	0.00
8	Acenaphthylene	2.024	1.954	3.5	88	0.00
9 C	Acenaphthene	1.356	1.319	2.7	88	0.00
10 S	Fluorene-d10	1.423	1.393	2.1	90	0.00
11	Fluorene	1.617	1.618	-0.1	91	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
13	Phenanthrene	1.105	1.069	3.3	95	0.00
14 S	Anthracene-d10	0.967	0.919	5.0	93	0.00
15	Anthracene	1.148	1.107	3.6	95	0.00
16 C	Fluoranthene	1.359	1.311	3.5	96	0.00
17 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
18 S	Pyrene-d10	0.895	0.863	3.6	98	0.00
19	Pyrene	1.148	1.119	2.5	98	0.00
20	Benzo(a)anthracene	1.231	1.152	6.4	98	0.00
21	Chrysene	1.100	1.048	4.7	99	0.00
22 I	Perylene-d12	1.000	1.000	0.0	101	0.00
23	Benzo(b)fluoranthene	1.189	1.087	8.6	96	0.00
24	Benzo(k)fluoranthene	1.137	1.078	5.2	101	0.00
25 S	Benzo(a)pyrene-d12	0.944	0.896	5.1	101	0.00
26 C	Benzo(a)pyrene	1.153	1.091	5.4	99	0.00
27	Indeno(1,2,3-cd)pyrene	1.488	1.382	7.1	99	0.00
28	Dibenzo(a,h)anthracene	1.251	1.163	7.0	99	0.00
29	Benzo(g,h,i)perylene	1.242	1.153	7.2	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0