

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM111219\
 Data File : BM023688.D
 Acq On : 13 Nov 2019 05:49
 Operator : JU
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD02014

Quant Time: Nov 13 07:58:48 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM11119MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 13 07:53:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.54	152	472592	20.00	ng/ul	0.00
18) Naphthalene-d8	10.32	136	2218092	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.20	164	1602475	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.95	188	3762297	20.00	ng/ul	0.00
78) Chrysene-d12	21.15	240	3347294	20.00	ng/ul	0.00
86) Perylene-d12	23.32	264	2962824	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.07	96	73397	6.40	ng/uL	0.00
5) Phenol-d5	6.73	99	846997	19.66	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.89	67	456550	18.44	ng/ul	0.00
9) 2-Chlorophenol-d4	7.08	132	624547	20.01	ng/ul	0.00
13) 4-Methylphenol-d8	8.26	113	698579	20.33	ng/ul	0.00
19) Nitrobenzene-d5	8.69	128	322866	20.34	ng/ul	0.00
22) 2-Nitrophenol-d4	9.41	143	374044	22.41	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.95	165	780421	20.51	ng/ul	0.00
29) 4-Chloroaniline-d4	10.46	131	843872	20.20	ng/ul	0.00
44) Dimethylphthalate-d6	13.62	166	2428276	19.43	ng/ul	0.00
47) Acenaphthylene-d8	13.89	160	2786265	19.54	ng/ul	0.00
52) 4-Nitrophenol-d4	14.42	143	428144	20.07	ng/ul	0.00
58) Fluorene-d10	15.20	176	2106789	19.09	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.33	200	395699	18.18	ng/ul	0.00
71) Anthracene-d10	17.05	188	3393809	19.02	ng/ul	0.00
79) Pyrene-d10	19.35	212	3732426	21.39	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.17	264	3020607	19.37	ng/ul	0.00

Target Compounds

					Qvalue
2) 1,4-Dioxane	3.10	88	80094	6.520	ng/uL 98
4) Benzaldehyde	6.69	77	527340	20.866	ng/ul 98
6) Phenol	6.76	94	855268	19.272	ng/ul 98
8) Bis(2-Chloroethyl)ether	6.98	93	634321	18.663	ng/ul 99
10) 2-Chlorophenol	7.11	128	640451	19.900	ng/ul 98
11) 2-Methylphenol	7.99	108	652844	19.709	ng/ul 99
12) 2,2'-oxybis(1-Chloropropan	8.08	45	628667	18.959	ng/ul 98
14) Acetophenone	8.36	105	1052514	19.638	ng/ul 99
15) N-Nitroso-di-n-propylamine	8.36	70	565931	20.099	ng/ul 98
16) 4-Methylphenol	8.32	108	731320	19.939	ng/ul 99
17) Hexachloroethane	8.61	117	244122	19.120	ng/ul 98
20) Nitrobenzene	8.73	77	791018	18.844	ng/ul 97
21) Isophorone	9.26	82	1571835	19.195	ng/ul 100
23) 2-Nitrophenol	9.44	139	395665	21.317	ng/ul 97
24) 2,4-Dimethylphenol	9.52	107	830968	19.590	ng/ul 97
25) Bis(2-Chloroethoxy)methane	9.75	93	952725	18.539	ng/ul 99
27) 2,4-Dichlorophenol	9.97	162	737890	20.161	ng/ul 99
28) Naphthalene	10.37	128	2193073	18.811	ng/ul 99
30) 4-Chloroaniline	10.48	127	850764	19.809	ng/ul 100
31) Hexachlorobutadiene	10.66	225	447681	18.115	ng/ul 97
32) Caprolactam	11.27	113	178317	15.034	ng/ul 94
33) 4-Chloro-3-methylphenol	11.63	107	836379	20.588	ng/ul 99
34) 2-Methylnaphthalene	11.99	142	1715285	19.233	ng/ul 98

Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM111219\
 Data File : BM023688.D
 Acq On : 13 Nov 2019 05:49
 Operator : JU
 Sample : SSTDCCC020EC
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD02014

Quant Time: Nov 13 07:58:48 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM11119MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 13 07:53:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
35) 1-Methylnaphthalene	12.22	142	1686312	19.172	ng/ul	99
37) 1,2,4,5-Tetrachlorobenzene	12.37	216	971761	18.559	ng/ul	99
38) Hexachlorocyclopentadiene	12.35	237	678746	24.747	ng/ul	100
39) 2,4,6-Trichlorophenol	12.62	196	659688	20.433	ng/ul	98
40) 2,4,5-Trichlorophenol	12.69	196	721234	20.512	ng/ul	99
41) 1,1'-Biphenyl	13.02	154	2335211	18.890	ng/ul	100
42) 2-Chloronaphthalene	13.06	162	1807565	18.997	ng/ul	99
43) 2-Nitroaniline	13.27	65	508135	22.452	ng/ul	95
45) Dimethylphthalate	13.67	163	2346727	19.162	ng/ul	100
46) 2,6-Dinitrotoluene	13.78	165	495071	22.195	ng/ul	98
48) Acenaphthylene	13.92	152	2786556	19.409	ng/ul	99
49) 3-Nitroaniline	14.11	138	464485	21.912	ng/ul	89
50) Acenaphthene	14.26	153	1942910	19.069	ng/ul	99
51) 2,4-Dinitrophenol	14.33	184	253944	18.058	ng/ul	96
53) 4-Nitrophenol	14.44	109	323924	18.816	ng/ul	94
54) Dibenzofuran	14.60	168	2845284	18.865	ng/ul	99
55) 2,4-Dinitrotoluene	14.57	165	741717	21.756	ng/ul	98
56) 2,3,4,6-Tetrachlorophenol	14.83	232	658241	19.779	ng/ul	99
57) Diethylphthalate	15.05	149	2391571	19.901	ng/ul	99
59) Fluorene	15.25	166	2350693	19.211	ng/ul	99
60) 4-Chlorophenyl-phenylether	15.26	204	1237753	18.745	ng/ul	97
61) 4-Nitroaniline	15.28	138	558935	21.528	ng/ul	96
64) 4,6-Dinitro-2-methylphenol	15.35	198	419456	17.725	ng/ul	99
65) N-Nitrosodiphenylamine	15.47	169	2137612	19.365	ng/ul	99
66) 4-Bromophenyl-phenylether	16.15	248	869476	18.910	ng/ul	98
67) Hexachlorobenzene	16.26	284	1024407	18.911	ng/ul	100
68) Atrazine	16.43	200	815877	19.731	ng/ul	99
69) Pentachlorophenol	16.60	266	473882	14.754	ng/ul	99
70) Phenanthrene	16.99	178	3920320	18.917	ng/ul	100
72) Anthracene	17.08	178	4025607	19.319	ng/ul	99
73) 1,2,3,4-Tetrachlorobenzene	12.99	216	969776	18.182	ng/uL	100
74) Pentachlorobenzene	14.52	250	1087490	18.453	ng/uL	99
75) Carbazole	17.36	167	3519582	20.262	ng/ul	100
76) Di-n-butylphthalate	17.93	149	4126032	21.541	ng/ul	99
77) Fluoranthene	19.02	202	4644556	20.919	ng/ul	99
80) Pyrene	19.38	202	4672906	21.598	ng/ul	99
81) Butylbenzylphthalate	20.30	149	1864549	27.209	ng/ul	99
82) 3,3'-Dichlorobenzidine	21.07	252	1400740	19.940	ng/ul	100
83) Benzo(a)anthracene	21.13	228	4316668	19.551	ng/ul	100
84) Bis(2-ethylhexyl)phthalate	21.09	149	2605992	25.187	ng/ul	100
85) Chrysene	21.19	228	3990611	19.199	ng/ul	99
87) Di-n-octyl phthalate	21.95	149	4017755	27.745	ng/ul	100
88) Benzo(b)fluoranthene	22.67	252	3729821	19.748	ng/ul	99
89) Benzo(k)fluoranthene	22.72	252	3496491	19.216	ng/ul	99
91) Benzo(a)pyrene	23.22	252	3309524	19.141	ng/ul	99
92) Indeno(1,2,3-cd)pyrene	25.46	276	3592057	19.429	ng/ul	99
93) Dibenzo(a,h)anthracene	25.47	278	3005203	19.260	ng/ul	99
94) Benzo(g,h,i)perylene	26.12	276	2953079	19.711	ng/ul	99


Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM111219\
Data File : BM023688.D
Acq On : 13 Nov 2019 05:49
Operator : JU
Sample : SSTDCCC020EC
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD02014

Quant Time: Nov 13 07:58:48 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM111119MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Nov 13 07:53:58 2019
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed						



#2

1,4-Dioxane

Concen: 6.520 ng/uL

RT: 3.10 min Scan# 37

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49

Instrument :

BNA_M

ClientSampleId :

SSTD02014


Tgt Ion: 88 Resp: 80094

Ion Ratio Lower Upper

88 100

43 27.2 18.9 28.3




58 61.4 49.1 73.7



Abundance Ion 88.00 (87.70 to 88.70): BM023679.D (-30) (-)

Ion 43.00 (42.70 to 43.70): BM023679.D (-30) (-)

Ion 58.00 (57.70 to 58.70): BM023679.D (-30) (-)



#4

Benzaldehyde

Concen: 20.866 ng/uL

RT: 6.69 min Scan# 647

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49


Tgt Ion: 77 Resp: 527340

Ion Ratio Lower Upper

77 100

105 95.1 75.9 113.9



106 91.8 70.2 105.4




Abundance Ion 77.00 (76.70 to 77.70): BM023679.D (-640) (-)

Ion 105.00 (104.70 to 105.70): BM023679.D (-640) (-)

Ion 106.00 (105.70 to 106.70): BM023679.D (-640) (-)





#6

Phenol

Concen: 19.272 ng/ul

RT: 6.76 min Scan# 658

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49

Instrument :

BNA_M

ClientSampleId :

SSTD02014


Tgt Ion: 94 Resp: 855268

Ion Ratio Lower Upper

94 100

65 26.5 21.4 32.2

66 36.5 30.7 46.1






Abundance Ion 94.00 (93.70 to 94.70): BM023679.D (-650) (-)

Ion 65.00 (64.70 to 65.70): BM023679.D (-650) (-)

Ion 66.00 (65.70 to 66.70): BM023679.D (-650) (-)

Time-->



#8

Bis(2-Chloroethyl)ether

Concen: 18.663 ng/ul

RT: 6.98 min Scan# 696

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49


Tgt Ion: 93 Resp: 634321

Ion Ratio Lower Upper

93 100

63 70.3 55.4 83.0

95 33.0 26.2 39.4






Abundance Ion 93.00 (92.70 to 93.70): BM023679.D (-689) (-)

Ion 63.00 (62.70 to 63.70): BM023679.D (-689) (-)

Ion 95.00 (94.70 to 95.70): BM023679.D (-689) (-)

Time-->





#10

2-Chlorophenol

Concen: 19.900 ng/ul

RT: 7.11 min Scan# 718

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49

Instrument :

BNA_M

ClientSampleId :

SSTD02014


Tgt Ion:128 Resp: 640451

Ion Ratio Lower Upper

128 100

64 45.8 38.1 57.1




130 32.2 26.4 39.6



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 64.00 (63.70 to 64.70): BM

Ion 130.00 (129.70 to 130.70): E



#11

2-Methylphenol

Concen: 19.709 ng/ul

RT: 7.99 min Scan# 868

Delta R.T. 0.00 min

Lab File: BM023688.D


Acq: 13 Nov 2019 05:49

Tgt Ion:108 Resp: 652844

Ion Ratio Lower Upper



108 100


107 90.1 71.4 107.2



Abundance Ion 108.00 (107.70 to 108.70): E

Ion 107.00 (106.70 to 107.70): E









#12
 2,2'-oxybis(1-Chloropropane)
 Concen: 18.959 ng/ul
 RT: 8.08 min Scan# 882
 Delta R.T. -0.01 min
 Lab File: BM023688.D
 Acq: 13 Nov 2019 05:49




Instrument :
 BNA_M
 ClientSampleId :
 SSTD02014


Tgt Ion: 45 Resp: 628667
 Ion Ratio Lower Upper
 45 100
 77 20.7 17.4 26.0
 79 16.4 13.9 20.9



#14
 Acetophenone
 Concen: 19.638 ng/ul
 RT: 8.36 min Scan# 931
 Delta R.T. 0.00 min
 Lab File: BM023688.D
 Acq: 13 Nov 2019 05:49

Tgt Ion: 105 Resp: 1052514
 Ion Ratio Lower Upper
 105 100
 77 79.2 63.8 95.6
 51 20.2 16.3 24.5









#15
N-Nitroso-di-n-propylamine
Concen: 20.099 ng/ul
RT: 8.36 min Scan# 930
Delta R.T. 0.00 min
Lab File: BM023688.D
Acq: 13 Nov 2019 05:49




Instrument :
BNA_M
ClientSampleId :
SSTD02014


Tgt Ion: 70 Resp: 565931
Ion Ratio Lower Upper
70 100
42 40.9 32.1 48.1
101 10.0 7.4 11.2
130 21.7 15.8 23.6



#16
4-Methylphenol
Concen: 19.939 ng/ul
RT: 8.32 min Scan# 924
Delta R.T. 0.00 min
Lab File: BM023688.D
Acq: 13 Nov 2019 05:49

Tgt Ion: 108 Resp: 731320
Ion Ratio Lower Upper
108 100
107 111.5 89.8 134.6





#17

Hexachloroethane

Concen: 19.120 ng/ul

RT: 8.61 min Scan# 973

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49

Instrument :

BNA_M

ClientSampleId :

SSTD02014


Tgt Ion:117 Resp: 244122

Ion Ratio Lower Upper

117 100

201 114.5 90.0 135.0




199 70.4 57.0 85.4



Abundance Ion 117.00 (116.70 to 117.70): E

Ion 201.00 (200.70 to 201.70): E

Ion 199.00 (198.70 to 199.70): E



#20

Nitrobenzene

Concen: 18.844 ng/ul

RT: 8.73 min Scan# 994

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49


Tgt Ion: 77 Resp: 791018

Ion Ratio Lower Upper

77 100

123 43.8 33.1 49.7



65 12.4 9.8 14.8




Abundance Ion 77.00 (76.70 to 77.70): BM0

Ion 123.00 (122.70 to 123.70): E

Ion 65.00 (64.70 to 65.70): BM0





#21

Isophorone

Concen: 19.195 ng/ul

RT: 9.26 min Scan# 1084

Delta R.T. -0.01 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49

Instrument :

BNA_M

ClientSampleId :

SSTD02014


Tgt Ion: 82 Resp: 1571835

Ion Ratio Lower Upper

82 100

95 7.4 5.8 8.8




138 17.0 13.4 20.2



Abundance Ion 82.00 (81.70 to 82.70): BM023679.D (-1076) (-)

Ion 95.00 (94.70 to 95.70): BM023679.D (-1076) (-)

Ion 138.00 (137.70 to 138.70): BM023679.D (-1076) (-)



#23

2-Nitrophenol

Concen: 21.317 ng/ul

RT: 9.44 min Scan# 1114

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49


Tgt Ion: 139 Resp: 395665

Ion Ratio Lower Upper

139 100

65 50.3 41.7 62.5



109 31.1 26.1 39.1




Abundance Ion 139.00 (138.70 to 139.70): BM023679.D (-1107) (-)

Ion 65.00 (64.70 to 65.70): BM023679.D (-1107) (-)

Ion 109.00 (108.70 to 109.70): BM023679.D (-1107) (-)









#24
2,4-Dimethylphenol
Concen: 19.590 ng/ul
RT: 9.52 min Scan# 1127
Delta R.T. 0.00 min
Lab File: BM023688.D
Acq: 13 Nov 2019 05:49




Instrument :
BNA_M
ClientSampleId :
SSTD02014


Tgt Ion	Ratio	Lower	Upper
107	100		
121	50.5	38.6	57.8
122	84.9	65.9	98.9



#25
Bis(2-Chloroethoxy)methane
Concen: 18.539 ng/ul
RT: 9.75 min Scan# 1166
Delta R.T. -0.01 min
Lab File: BM023688.D
Acq: 13 Nov 2019 05:49

Tgt Ion	Ratio	Lower	Upper
93	100		
95	32.5	25.7	38.5
123	10.7	8.3	12.5









#27
2,4-Dichlorophenol
Concen: 20.161 ng/ul
RT: 9.97 min Scan# 1205
Delta R.T. 0.00 min
Lab File: BM023688.D
Acq: 13 Nov 2019 05:49




Instrument :
BNA_M
ClientSampleId :
SSTD02014


Tgt Ion	Ratio	Lower	Upper
162	100		
164	64.3	51.9	77.9
98	36.7	29.9	44.9



#28
Naphthalene
Concen: 18.811 ng/ul
RT: 10.37 min Scan# 1272
Delta R.T. 0.00 min
Lab File: BM023688.D
Acq: 13 Nov 2019 05:49

Tgt Ion	Ratio	Lower	Upper
128	100		
129	10.8	8.7	13.1
127	12.8	10.6	16.0









#30
4-Chloroaniline
Concen: 19.809 ng/ul
RT: 10.48 min Scan# 1291
Delta R.T. 0.00 min
Lab File: BM023688.D
Acq: 13 Nov 2019 05:49




Instrument :
BNA_M
ClientSampleId :
SSTD02014


Tgt Ion:127 Resp: 850764
Ion Ratio Lower Upper
127 100
129 32.3 26.1 39.1



#31
Hexachlorobutadiene
Concen: 18.115 ng/ul
RT: 10.66 min Scan# 1321
Delta R.T. -0.01 min
Lab File: BM023688.D
Acq: 13 Nov 2019 05:49

Tgt Ion:225 Resp: 447681
Ion Ratio Lower Upper
225 100
223 60.8 50.5 75.7
227 62.6 51.9 77.9





#32

Caprolactam

Concen: 15.034 ng/ul

RT: 11.27 min Scan# 1426

Delta R.T. 0.01 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49

Instrument :

BNA_M

ClientSampleId :

SSTD02014


Tgt Ion:113 Resp: 178317

Ion Ratio Lower Upper

113 100

55 125.9 105.8 158.6




56 97.2 83.0 124.6



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BM0

Ion 56.00 (55.70 to 56.70): BM0



#33

4-Chloro-3-methylphenol

Concen: 20.588 ng/ul

RT: 11.63 min Scan# 1486

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49


Tgt Ion:107 Resp: 836379

Ion Ratio Lower Upper

107 100

144 25.9 20.6 30.8



142 79.2 62.9 94.3




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E





#34

2-Methylnaphthalene

Concen: 19.233 ng/ul

RT: 11.99 min Scan# 1548

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49

Instrument :

BNA_M

ClientSampleId :


SSTD02014

Tgt Ion:142 Resp: 1715285

Ion Ratio Lower Upper




142 100

141 89.0 69.8 104.8



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E



#35

1-Methylnaphthalene

Concen: 19.172 ng/ul

RT: 12.22 min Scan# 1586

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49


Tgt Ion:142 Resp: 1686312

Ion Ratio Lower Upper

142 100

141 91.9 74.6 112.0



116 3.7 3.2 4.8




Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 116.00 (115.70 to 116.70): E









#37
 1,2,4,5-Tetrachlorobenzene
 Concen: 18.559 ng/ul
 RT: 12.37 min Scan# 1612
 Delta R.T. 0.00 min
 Lab File: BM023688.D
 Acq: 13 Nov 2019 05:49

Instrument :
 BNA_M
 ClientSampleId :
 SSTD02014

Tgt Ion:	216	Resp:	971761
Ion	Ratio	Lower	Upper
216	100		
214	78.0	61.7	92.5
179	20.1	16.5	24.7
108	15.5	12.7	19.1




Abundance Ion 216.00 (215.70 to 216.70): E
 Ion 214.00 (213.70 to 214.70): E
 Ion 179.00 (178.70 to 179.70): E
 Ion 108.00 (107.70 to 108.70): E






#38
 Hexachlorocyclopentadiene
 Concen: 24.747 ng/ul
 RT: 12.35 min Scan# 1609
 Delta R.T. 0.00 min
 Lab File: BM023688.D
 Acq: 13 Nov 2019 05:49

Tgt Ion:	237	Resp:	678746
Ion	Ratio	Lower	Upper
237	100		
235	61.9	49.8	74.8
272	13.6	10.8	16.2



Abundance Ion 237.00 (236.70 to 237.70): E
 Ion 235.00 (234.70 to 235.70): E
 Ion 272.00 (271.70 to 272.70): E





#39

2,4,6-Trichlorophenol

Concen: 20.433 ng/ul

RT: 12.62 min Scan# 1655

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49

Instrument :

BNA_M

ClientSampleId :

SSTD02014


Tgt Ion:196 Resp: 659688

Ion Ratio Lower Upper

196 100

198 96.3 78.5 117.7




200 30.3 25.6 38.4



Abundance Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E

Ion 200.00 (199.70 to 200.70): E



#40

2,4,5-Trichlorophenol

Concen: 20.512 ng/ul

RT: 12.69 min Scan# 1667

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49


Tgt Ion:196 Resp: 721234

Ion Ratio Lower Upper

196 100

198 94.9 75.3 112.9



200 30.6 25.0 37.6




Abundance Ion 196.00 (195.70 to 196.70): E

Ion 198.00 (197.70 to 198.70): E

Ion 200.00 (199.70 to 200.70): E









#41
 1,1'-Biphenyl
 Concen: 18.890 ng/ul
 RT: 13.02 min Scan# 1723
 Delta R.T. -0.01 min
 Lab File: BM023688.D
 Acq: 13 Nov 2019 05:49




Instrument :
 BNA_M
 ClientSampleId :
 SST02014


Tgt Ion:154 Resp: 2335211
 Ion Ratio Lower Upper
 154 100
 153 40.4 32.3 48.5
 76 13.3 10.6 15.8



#42
 2-Chloronaphthalene
 Concen: 18.997 ng/ul
 RT: 13.06 min Scan# 1730
 Delta R.T. 0.00 min
 Lab File: BM023688.D
 Acq: 13 Nov 2019 05:49

Tgt Ion:162 Resp: 1807565
 Ion Ratio Lower Upper
 162 100
 164 33.0 26.7 40.1
 127 37.2 29.5 44.3









#43
2-Nitroaniline
Concen: 22.452 ng/ul
RT: 13.27 min Scan# 1766
Delta R.T. 0.00 min
Lab File: BM023688.D
Acq: 13 Nov 2019 05:49




Instrument :
BNA_M
ClientSampleId :
SSTD02014


Tgt Ion	Ratio	Lower	Upper
65	100		
92	76.0	60.4	90.6
138	112.9	83.8	125.6



#45
Dimethylphthalate
Concen: 19.162 ng/ul
RT: 13.67 min Scan# 1833
Delta R.T. 0.00 min
Lab File: BM023688.D
Acq: 13 Nov 2019 05:49

Tgt Ion	Ratio	Lower	Upper
163	100		
194	4.6	3.4	5.2
164	10.0	8.0	12.0









#46
2,6-Dinitrotoluene
Concen: 22.195 ng/ul
RT: 13.78 min Scan# 1852
Delta R.T. 0.00 min
Lab File: BM023688.D
Acq: 13 Nov 2019 05:49




Instrument :
BNA_M
ClientSampleId :
SSTD02014


Tgt Ion:165 Resp: 495071
Ion Ratio Lower Upper
165 100
89 53.0 43.8 65.8
121 20.3 17.0 25.4



#48
Acenaphthylene
Concen: 19.409 ng/ul
RT: 13.92 min Scan# 1875
Delta R.T. 0.00 min
Lab File: BM023688.D
Acq: 13 Nov 2019 05:49

Tgt Ion:152 Resp: 2786556
Ion Ratio Lower Upper
152 100
151 19.6 16.1 24.1
153 13.1 10.4 15.6





#49

3-Nitroaniline

Concen: 21.912 ng/ul

RT: 14.11 min Scan# 1908

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49

Instrument :

BNA_M

ClientSampleId :

SSTD02014


Tgt Ion:138 Resp: 464485

Ion Ratio Lower Upper

138 100

108 11.3 10.2 15.2




92 117.5 105.4 158.0



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BM0



#50

Acenaphthene

Concen: 19.069 ng/ul

RT: 14.26 min Scan# 1934

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49


Tgt Ion:153 Resp: 1942910

Ion Ratio Lower Upper

153 100

152 46.7 37.7 56.5



154 90.1 70.6 106.0




Abundance Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E

Ion 154.00 (153.70 to 154.70): E









#51
 2,4-Dinitrophenol
 Concen: 18.058 ng/ul
 RT: 14.33 min Scan# 1945
 Delta R.T. 0.01 min
 Lab File: BM023688.D
 Acq: 13 Nov 2019 05:49




Instrument :
 BNA_M
 ClientSampleId :
 SSTD02014


Tgt Ion:184 Resp: 253944
 Ion Ratio Lower Upper
 184 100
 63 49.9 43.4 65.0
 154 57.9 47.2 70.8



#53
 4-Nitrophenol
 Concen: 18.816 ng/ul
 RT: 14.44 min Scan# 1964
 Delta R.T. 0.01 min
 Lab File: BM023688.D
 Acq: 13 Nov 2019 05:49

Tgt Ion:109 Resp: 323924
 Ion Ratio Lower Upper
 109 100
 139 140.8 103.4 155.0
 65 129.7 101.3 151.9





#54

Dibenzenofuran

Concen: 18.865 ng/ul

RT: 14.60 min Scan# 1991

Delta R.T. -0.01 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49

Instrument :

BNA_M

ClientSampleId :


SSTD02014

Tgt Ion:168 Resp: 2845284

Ion Ratio Lower Upper




168 100

139 36.2 29.3 43.9



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E



#55

2,4-Dinitrotoluene

Concen: 21.756 ng/ul

RT: 14.57 min Scan# 1987

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49


Tgt Ion:165 Resp: 741717

Ion Ratio Lower Upper

165 100

63 37.5 31.0 46.6



182 2.9 2.4 3.6




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BM0

Ion 182.00 (181.70 to 182.70): E









#56
2,3,4,6-Tetrachlorophenol
Concen: 19.779 ng/ul
RT: 14.83 min Scan# 2031
Delta R.T. 0.00 min
Lab File: BM023688.D
Acq: 13 Nov 2019 05:49

Instrument :
BNA_M
ClientSampleId :
SSTD02014

Tgt Ion:	232	Resp:	658241
Ion	Ratio	Lower	Upper
232	100		
131	37.7	30.1	45.1
130	2.1	1.5	2.3
166	28.8	22.5	33.7




Abundance Ion 232.00 (231.70 to 232.70): E
Ion 131.00 (130.70 to 131.70): E
Ion 130.00 (129.70 to 130.70): E
Ion 166.00 (165.70 to 166.70): E






#57
Diethylphthalate
Concen: 19.901 ng/ul
RT: 15.05 min Scan# 2067
Delta R.T. 0.00 min
Lab File: BM023688.D
Acq: 13 Nov 2019 05:49

Tgt Ion:	149	Resp:	2391571
Ion	Ratio	Lower	Upper
149	100		
177	23.2	18.9	28.3
150	12.6	9.9	14.9



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 177.00 (176.70 to 177.70): E
Ion 150.00 (149.70 to 150.70): E





#59

Fluorene

Concen: 19.211 ng/ul

RT: 15.25 min Scan# 2102

Delta R.T. -0.01 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49

Instrument :

BNA_M

ClientSampleId :

SSTD02014


Tgt Ion:166 Resp: 2350693

Ion Ratio Lower Upper

166 100

165 97.3 77.2 115.8




167 13.5 10.9 16.3



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#60

4-Chlorophenyl-phenylether

Concen: 18.745 ng/ul

RT: 15.26 min Scan# 2103

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49


Tgt Ion:204 Resp: 1237753

Ion Ratio Lower Upper

204 100

206 32.4 25.8 38.6



141 52.9 44.5 66.7




Abundance Ion 204.00 (203.70 to 204.70): E

Ion 206.00 (205.70 to 206.70): E

Ion 141.00 (140.70 to 141.70): E









#61
4-Nitroaniline
Concen: 21.528 ng/ul
RT: 15.28 min Scan# 2107
Delta R.T. 0.00 min
Lab File: BM023688.D
Acq: 13 Nov 2019 05:49




Instrument :
BNA_M
ClientSampleId :
SSTD02014


Tgt Ion	Ratio	Lower	Upper
138	100		
92	53.5	44.6	66.8
108	78.2	66.2	99.4



#64
4,6-Dinitro-2-methylphenol
Concen: 17.725 ng/ul
RT: 15.35 min Scan# 2118
Delta R.T. 0.00 min
Lab File: BM023688.D
Acq: 13 Nov 2019 05:49

Tgt Ion	Ratio	Lower	Upper
198	100		
182	4.3	3.6	5.4
77	21.7	17.1	25.7





#65

N-Nitrosodiphenylamine

Concen: 19.365 ng/ul

RT: 15.47 min Scan# 2139

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49

Instrument :

BNA_M

ClientSampleId :

SSTD02014


Tgt Ion:169 Resp: 2137612

Ion Ratio Lower Upper

169 100

168 66.7 53.2 79.8




167 35.9 28.0 42.0



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E



#66

4-Bromophenyl-phenylether

Concen: 18.910 ng/ul

RT: 16.15 min Scan# 2254

Delta R.T. -0.01 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49


Tgt Ion:248 Resp: 869476

Ion Ratio Lower Upper

248 100

250 97.9 76.6 114.8



141 63.7 51.3 76.9




Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E





#67

Hexachlorobenzene

Concen: 18.911 ng/ul

RT: 16.26 min Scan# 2273

Delta R.T. -0.01 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49

Instrument :

BNA_M

ClientSampled :

SSTD02014


Tgt Ion:284 Resp: 1024407

Ion Ratio Lower Upper

284 100

142 27.2 21.8 32.6




249 29.0 23.1 34.7



Abundance Ion 284.00 (283.70 to 284.70): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E



#68

Atrazine

Concen: 19.731 ng/ul

RT: 16.43 min Scan# 2303

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49


Tgt Ion:200 Resp: 815877

Ion Ratio Lower Upper

200 100

173 26.0 20.6 30.8



215 48.7 39.7 59.5




Abundance Ion 200.00 (199.70 to 200.70): E

Ion 173.00 (172.70 to 173.70): E

Ion 215.00 (214.70 to 215.70): E









#69
 Pentachlorophenol
 Concen: 14.754 ng/ul
 RT: 16.60 min Scan# 2332
 Delta R.T. 0.00 min
 Lab File: BM023688.D
 Acq: 13 Nov 2019 05:49

Instrument :
 BNA_M
 ClientSampleId :
 SST02014

Tgt Ion:266 Resp: 473882
 Ion Ratio Lower Upper
 266 100
 264 62.5 50.6 75.8
 268 63.8 50.9 76.3




Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 264.00 (263.70 to 264.70): E
 Ion 268.00 (267.70 to 268.70): E






#70
 Phenanthrene
 Concen: 18.917 ng/ul
 RT: 16.99 min Scan# 2398
 Delta R.T. 0.00 min
 Lab File: BM023688.D
 Acq: 13 Nov 2019 05:49

Tgt Ion:178 Resp: 3920320
 Ion Ratio Lower Upper
 178 100
 179 15.5 12.3 18.5
 176 19.1 15.0 22.6



Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 179.00 (178.70 to 179.70): E
 Ion 176.00 (175.70 to 176.70): E





#72

Anthracene

Concen: 19.319 ng/ul

RT: 17.08 min Scan# 2413

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49

Instrument :

BNA_M

ClientSampleId :

SSTD02014


Tgt Ion:178 Resp: 4025607

Ion Ratio Lower Upper

178 100

179 15.1 12.3 18.5




176 18.1 14.7 22.1



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E



#73

1,2,3,4-Tetrachlorobenzene

Concen: 18.182 ng/uL

RT: 12.99 min Scan# 1717

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49


Tgt Ion:216 Resp: 969776

Ion Ratio Lower Upper

216 100

214 79.0 63.1 94.7



218 48.1 38.6 57.8




Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 218.00 (217.70 to 218.70): E





#74

Pentachlorobenzene

Concen: 18.453 ng/uL

RT: 14.52 min Scan# 1978

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49

Instrument :

BNA_M

ClientSampleId :

SSTD02014


Tgt Ion:250 Resp: 1087490

Ion Ratio Lower Upper

250 100

248 62.1 50.4 75.6




252 63.8 51.0 76.4



Abundance Ion 250.00 (249.70 to 250.70): E

Ion 248.00 (247.70 to 248.70): E

Ion 252.00 (251.70 to 252.70): E



#75

Carbazole

Concen: 20.262 ng/uL

RT: 17.36 min Scan# 2460

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49


Tgt Ion:167 Resp: 3519582

Ion Ratio Lower Upper

167 100

166 21.5 17.1 25.7



139 12.1 9.8 14.6




Abundance Ion 167.00 (166.70 to 167.70): E

Ion 166.00 (165.70 to 166.70): E

Ion 139.00 (138.70 to 139.70): E









#76
Di-n-butylphthalate
Concen: 21.541 ng/ul
RT: 17.93 min Scan# 2558
Delta R.T. -0.01 min
Lab File: BM023688.D
Acq: 13 Nov 2019 05:49




Instrument :
BNA_M
ClientSampleId :
SSTD02014


Tgt Ion:149 Resp: 4126032
Ion Ratio Lower Upper
149 100
150 9.2 7.1 10.7
104 4.8 4.2 6.2



#77
Fluoranthene
Concen: 20.919 ng/ul
RT: 19.02 min Scan# 2742
Delta R.T. 0.00 min
Lab File: BM023688.D
Acq: 13 Nov 2019 05:49

Tgt Ion:202 Resp: 4644556
Ion Ratio Lower Upper
202 100
101 10.3 7.7 11.5
100 7.5 6.0 9.0









#80
 Pyrene
 Concen: 21.598 ng/ul
 RT: 19.38 min Scan# 2804
 Delta R.T. 0.00 min
 Lab File: BM023688.D
 Acq: 13 Nov 2019 05:49

Instrument :
 BNA_M
 ClientSampleId :
 SSTD02014

Tgt Ion: 202 Resp: 4672906




Ion	Ratio	Lower	Upper
202	100		
101	11.8	9.8	14.6
100	9.8	8.1	12.1




#81
 Butylbenzylphthalate
 Concen: 27.209 ng/ul
 RT: 20.30 min Scan# 2961
 Delta R.T. 0.00 min
 Lab File: BM023688.D
 Acq: 13 Nov 2019 05:49

Tgt Ion: 149 Resp: 1864549

Ion	Ratio	Lower	Upper
149	100		
91	67.7	54.9	82.3
206	23.8	19.3	28.9









#82
 3,3'-Dichlorobenzidine
 Concen: 19.940 ng/ul
 RT: 21.07 min Scan# 3092
 Delta R.T. 0.00 min
 Lab File: BM023688.D
 Acq: 13 Nov 2019 05:49




Instrument :
 BNA_M
 ClientSampleId :
 SSTD02014


Tgt Ion:252 Resp: 1400740
 Ion Ratio Lower Upper
 252 100
 254 64.7 51.6 77.4
 126 8.8 6.5 9.7



#83
 Benzo(a)anthracene
 Concen: 19.551 ng/ul
 RT: 21.13 min Scan# 3102
 Delta R.T. 0.00 min
 Lab File: BM023688.D
 Acq: 13 Nov 2019 05:49

Tgt Ion:228 Resp: 4316668
 Ion Ratio Lower Upper
 228 100
 229 19.8 15.6 23.4
 226 26.9 21.4 32.2





#84

Bis(2-ethylhexyl)phthalate

Concen: 25.187 ng/ul

RT: 21.09 min Scan# 3095

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49

Instrument :

BNA_M

ClientSampleId :

SSTD02014


Tgt Ion:149 Resp: 2605992

Ion Ratio Lower Upper

149 100

167 29.4 23.4 35.2




279 6.3 5.0 7.4



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#85

Chrysene

Concen: 19.199 ng/ul

RT: 21.19 min Scan# 3111

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49


Tgt Ion:228 Resp: 3990611

Ion Ratio Lower Upper

228 100

226 30.3 23.8 35.8



229 19.9 15.8 23.6




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#87
 Di-n-octyl phthalate
 Concen: 27.745 ng/ul
 RT: 21.95 min Scan# 3241
 Delta R.T. 0.00 min
 Lab File: BM023688.D
 Acq: 13 Nov 2019 05:49




Instrument :
 BNA_M
 ClientSampleId :
 SSTD02014


Tgt Ion:149 Resp: 4017755



#88
 Benzo(b)fluoranthene
 Concen: 19.748 ng/ul
 RT: 22.67 min Scan# 3363
 Delta R.T. 0.00 min
 Lab File: BM023688.D
 Acq: 13 Nov 2019 05:49

Tgt Ion:252 Resp: 3729821
 Ion Ratio Lower Upper
 252 100
 253 22.0 17.8 26.8
 125 7.9 6.5 9.7





#89

Benzo(k)fluoranthene

Concen: 19.216 ng/ul

RT: 22.72 min Scan# 3371

Delta R.T. 0.01 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49

Instrument :

BNA_M

ClientSampleId :

SSTD02014


Tgt Ion:252 Resp: 3496491

Ion Ratio Lower Upper

252 100

253 22.0 17.2 25.8




125 7.6 5.9 8.9



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#91

Benzo(a)pyrene

Concen: 19.141 ng/ul

RT: 23.22 min Scan# 3457

Delta R.T. 0.01 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49


Tgt Ion:252 Resp: 3309524

Ion Ratio Lower Upper

252 100

253 22.4 17.3 25.9



125 8.8 6.9 10.3




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#92

Indeno(1,2,3-cd)pyrene

Concen: 19.429 ng/ul

RT: 25.46 min Scan# 3838

Delta R.T. 0.00 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49

Instrument :


BNA_M

ClientSampleId :

SSTD02014

Tgt Ion:276 Resp: 3592057




Ion	Ratio	Lower	Upper
276	100		
138	18.2	15.0	22.4
277	24.9	20.0	30.0



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E



#93

Dibenzo(a,h)anthracene

Concen: 19.260 ng/ul

RT: 25.47 min Scan# 3840


Delta R.T. 0.01 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49

Tgt Ion:278 Resp: 3005203



Ion	Ratio	Lower	Upper
278	100		
139	11.7	10.0	15.0
279	23.6	19.0	28.6




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#94

Benzo(g,h,i)perylene

Concen: 19.711 ng/ul

RT: 26.12 min Scan# 3949

Delta R.T. 0.01 min

Lab File: BM023688.D

Acq: 13 Nov 2019 05:49

Instrument :

BNA_M

ClientSampleId :

SSTD02014


Tgt Ion:276 Resp: 2953079

Ion Ratio Lower Upper

276 100

138 15.8 13.2 19.8

277 23.9 18.6 28.0



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E

