




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111222\
Data File : BM037487.D
Acq On : 12 Nov 2022 10:54
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 14 08:16:45 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM111222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 14 01:42:00 2022
Response via : Initial Calibration



TIC: BM037487.D\data.ms

(70) Pentachlorophenol (C)




16.845min (+ 0.000) 702026.83 ng

response 467513

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.20	64.02
264.00	61.90	61.73
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111222\
Data File : BM037487.D
Acq On : 12 Nov 2022 10:54
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 14 08:16:45 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM111222.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 14 01:42:00 2022
Response via : Initial Calibration



TIC: BM037487.D\data.ms

(77) Benzidine**19.451min (+ 0.001) 4630.22 ng****response 2438364**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	13.90	14.10
183.00	11.30	11.38
0.00	0.00	0.00