

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111321\
 Data File : BM033061.D
 Acq On : 14 Nov 2021 03:14
 Operator : CG/JU
 Sample : M4413-16
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BG375

Quant Time: Nov 15 03:31:08 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM110921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 15 03:11:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.434	152	2113	0.400	ng/ul	0.00
4) Naphthalene-d8	10.208	136	8574	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.091	164	5652	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.838	188	11381	0.400	ng/ul	0.00
17) Chrysene-d12	21.033	240	7993	0.400	ng/ul	0.00
23) Perylene-d12	23.134	264	6831	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.843	96	11089	4.110	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	11.804	152	4211	0.346	ng/ul	0.00
18) Fluoranthene-d10	18.869	212	11366	0.469	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111321\
Data File : BM033061.D
Acq On : 14 Nov 2021 03:14
Operator : CG/JU
Sample : M4413-16
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BG375

Quant Time: Nov 15 03:31:08 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM110921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 15 03:11:31 2021
Response via : Initial Calibration

