

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111321\
 Data File : BM033077.D
 Acq On : 14 Nov 2021 13:22
 Operator : CG/JU
 Sample : M4446-01
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BG378

Quant Time: Nov 15 03:50:54 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM110921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 15 03:11:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.431	152	2011	0.400	ng/ul	0.00
4) Naphthalene-d8	10.199	136	7993	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.091	164	5069	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.834	188	10467	0.400	ng/ul	0.00
17) Chrysene-d12	21.033	240	6923	0.400	ng/ul	0.00
23) Perylene-d12	23.134	264	5892	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.840	96	9117	3.550	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	11.804	152	3391	0.299	ng/ul	0.00
18) Fluoranthene-d10	18.866	212	9234	0.440	ng/ul	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111321\
Data File : BM033077.D
Acq On : 14 Nov 2021 13:22
Operator : CG/JU
Sample : M4446-01
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BG378

Quant Time: Nov 15 03:50:54 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM110921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 15 03:11:31 2021
Response via : Initial Calibration

