

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111321\
 Data File : BM033080.D
 Acq On : 14 Nov 2021 15:10
 Operator : CG/JU
 Sample : M4445-11
 Misc :
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BG385

Quant Time: Nov 15 03:52:59 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM110921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 15 03:11:31 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.428	152	2194	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.199	136	8771	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.091	164	5597	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.834	188	11120	0.400	ng/ul	0.00
17) Chrysene-d12	21.030	240	7118	0.400	ng/ul	0.00
23) Perylene-d12	23.132	264	6099	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.839	96	9989	3.565	ng/ul	-0.01
6) 2-Methylnaphthalene-d10	11.804	152	3253	0.261	ng/ul	0.00
18) Fluoranthene-d10	18.866	212	8474	0.393	ng/ul	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111321\
Data File : BM033080.D
Acq On : 14 Nov 2021 15:10
Operator : CG/JU
Sample : M4445-11
Misc :
ALS Vial : 44 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BG385

Quant Time: Nov 15 03:52:59 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM110921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 15 03:11:31 2021
Response via : Initial Calibration

