

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM111516\
 Data File : BM007902.D
 Acq On : 15 Nov 2016 20:44
 Operator : UM/SJ
 Sample : H5612-05
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 F2H05

Manual Integrations
 APPROVED

Umangi
 11/16/2016 6:02:20 PM

Quant Time: Nov 16 08:06:26 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM102516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 16 07:43:19 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.74	152	894	0.40	ng/ul	0.00
2) Naphthalene-d8	10.51	136	2880	0.40	ng/ul	-0.01
6) Acenaphthene-d10	14.37	164	1775	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.11	188	3042m	0.40	ng/ul	-0.02
16) Chrysene-d12	21.31	240	3430	0.40	ng/ul	0.00
20) Perylene-d12	23.56	264	2788m	0.40	ng/ul	-0.01
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.11	152	1202	0.35	ng/ul	-0.01
14) Fluoranthene-d10	19.16	212	2807m	0.41	ng/ul	0.00
Target Compounds						
15) Fluoranthene	19.18	202	467	0.05	ng/ul#	57
18) Benzo(a)anthracene	21.29	228	286	0.03	ng/ul#	60
19) Chrysene	21.35	228	286	0.03	ng/ul#	65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM111516\
Data File : BM007902.D
Acq On : 15 Nov 2016 20:44
Operator : UM/SJ
Sample : H5612-05
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
F2H05

Manual Integrations
APPROVED

Umangi
11/16/2016 6:02:20 PM

Quant Time: Nov 16 08:06:26 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM102516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 16 07:43:19 2016
Response via : Initial Calibration

