

Data Path : Z:\HPCHEM1\BNA_M\Data\BM111516\
 Data File : BM007895.D
 Acq On : 15 Nov 2016 16:36
 Operator : UM/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.462

Quant Time: Nov 16 01:23:06 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM102516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 16 01:19:46 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	83	0.00
2 I	Naphthalene-d8	0.400	0.400	0.0	85	0.00
3	Naphthalene	0.400	0.411	-2.7	85	0.00
4 SURR	2-Methylnaphthalene-d10	0.400	0.441	-10.2	91	0.00
5	2-Methylnaphthalene	0.400	0.427	-6.7	88	0.00
6 I	Acenaphthene-d10	0.400	0.400	0.0	104	0.00
7	Acenaphthylene	0.400	0.387	3.3	95	0.00
8 C	Acenaphthene	0.400	0.371	7.3	90	0.00
9	Fluorene	0.400	0.378	5.5	94	0.00
10 I	Phenanthrene-d10	0.400	0.400	0.0	95	0.00
11	Pentachlorophenol	0.400	0.501	-25.2#	131	0.00
12	Phenanthrene	0.400	0.409	-2.2	92	0.00
13	Anthracene	0.400	0.428	-7.0	100	0.00
14 SURR	Fluoranthene-d10	0.400	0.459	-14.8	104	0.00
15 C	Fluoranthene	0.400	0.426	-6.5	96	0.00
16 I	Chrysene-d12	0.400	0.400	0.0	100	0.00
17	Pyrene	0.400	0.418	-4.5	94	0.00
18	Benzo(a)anthracene	0.400	0.438	-9.5	101	0.00
19	Chrysene	0.400	0.407	-1.7	88	0.00
20 I	Perylene-d12	0.400	0.400	0.0	96	0.00
21	Benzo(b)fluoranthene	0.400	0.435	-8.7	91	0.00
22	Benzo(k)fluoranthene	0.400	0.400	0.0	91	0.00
23 C	Benzo(a)pyrene	0.400	0.438	-9.5	96	0.00
24	Indeno(1,2,3-cd)pyrene	0.400	0.433	-8.2	92	0.00
25	Dibenzo(a,h)anthracene	0.400	0.426	-6.5	91	0.00
26	Benzo(g,h,i)perylene	0.400	0.419	-4.7	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0