

Data Path : Z:\HPCHEM1\BNA_M\Data\BM111516\
 Data File : BM007910.D
 Acq On : 16 Nov 2016 01:27
 Operator : UM/SJ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.463

Quant Time: Nov 16 07:46:19 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM102516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 16 07:43:19 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	78	0.00
2 I	Naphthalene-d8	0.400	0.400	0.0	80	0.00
3	Naphthalene	0.400	0.418	-4.5	81	0.00
4 SURR	2-Methylnaphthalene-d10	0.400	0.425	-6.2	82	0.01
5	2-Methylnaphthalene	0.400	0.420	-5.0	81	0.01
6 I	Acenaphthene-d10	0.400	0.400	0.0	91	0.00
7	Acenaphthylene	0.400	0.400	0.0	86	0.00
8 C	Acenaphthene	0.400	0.382	4.5	81	0.00
9	Fluorene	0.400	0.385	3.8	84	0.02
10 I	Phenanthrene-d10	0.400	0.400	0.0	85	0.00
11	Pentachlorophenol	0.400	0.482	-20.5#	112	0.02
12	Phenanthrene	0.400	0.407	-1.7	82	0.00
13	Anthracene	0.400	0.425	-6.2	89	0.00
14 SURR	Fluoranthene-d10	0.400	0.453	-13.2	91	0.00
15 C	Fluoranthene	0.400	0.430	-7.5	87	0.01
16 I	Chrysene-d12	0.400	0.400	0.0	87	0.00
17	Pyrene	0.400	0.431	-7.7	84	0.00
18	Benzo(a)anthracene	0.400	0.432	-8.0	87	0.01
19	Chrysene	0.400	0.426	-6.5	80	0.00
20 I	Perylene-d12	0.400	0.400	0.0	85	0.00
21	Benzo(b)fluoranthene	0.400	0.423	-5.7	77	0.00
22	Benzo(k)fluoranthene	0.400	0.424	-6.0	84	0.00
23 C	Benzo(a)pyrene	0.400	0.449	-12.2	87	0.00
24	Indeno(1,2,3-cd)pyrene	0.400	0.445	-11.2	83	0.00
25	Dibenzo(a,h)anthracene	0.400	0.438	-9.5	82	0.00
26	Benzo(g,h,i)perylene	0.400	0.441	-10.2	81	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0