

Data Path : Z:\HPCHEM1\BNA M\DATA\BM111615\
 Data File : BM003136.D
 Acq On : 16 Nov 2015 10:47
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Nov 17 13:42:35 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM110415.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 17 12:50:51 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	-0.02
2	1,4-Dioxane	0.510	0.447	12.4	83	-0.02
3	n-Nitrosodimethylamine	0.469	0.440	6.2	85	-0.02
4 S	2-Fluorophenol	1.058	0.956	9.6	83	-0.02
5 S	Phenol-d6	1.234	1.107	10.3	83	-0.02
6	bis(2-Chloroethyl)ether	1.183	1.125	4.9	86	-0.03
7 I	Naphthalene-d8	1.000	1.000	0.0	87	-0.01
8 S	Nitrobenzene-d5	0.231	0.221	4.3	88	-0.02
9	Nitrobenzene	0.249	0.238	4.4	88	-0.02
10	Naphthalene	1.075	1.064	1.0	88	-0.01
11	Hexachlorobutadiene	0.166	0.166	0.0	88	-0.01
12	2-Methylnaphthalene	0.695	0.680	2.2	87	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	84	0.00
14 S	2,4,6-Tribromophenol	0.129	0.126	2.3	83	0.00
15 S	2-Fluorobiphenyl	1.576	1.630	-3.4	88	-0.02
16	Acenaphthylene	1.946	1.830	6.0	82	0.00
17	Acenaphthene	1.339	1.410	-5.3	96	-0.03
18	Fluorene	1.426	1.731	-21.4	100	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	88	0.00
20	4-Bromophenyl-phenylether	0.173	0.274	-58.4#	106	0.00
21	Hexachlorobenzene	0.247	0.287	-16.2	83	0.00
22	Pentachlorophenol	0.076	0.088	-15.8	93	0.00
23	Phenanthrene	1.134	1.709	-50.7#	103	0.00
24	Anthracene	1.060	1.030	2.8	85	0.00
25	Fluoranthene	1.127	1.342	-19.1	96	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	101	0.00
27	Benzydine	0.454	0.422	7.0	97	0.00
28	Pvrene	1.452	1.366	5.9	97	0.00
29 S	Terphenyl-d14	0.721	0.657	8.9	89	0.00
30	Benzo(a)anthracene	1.406	1.269	9.7	102	0.00
31	3,3'-Dichlorobenzidine	0.367	0.329	10.4	91	0.00
32	Chrysene	1.562	1.246	20.2	82	0.00
33	Bis(2-ethylhexyl)phthalate	0.562	0.356	36.7#	74	0.00
34	Indeno(1,2,3-cd)pyrene	1.385	1.267	8.5	92	-0.02
35 I	Perylene-d12	1.000	1.000	0.0	92	0.00
36	Benzo(b)fluoranthene	1.707	1.357	20.5	84	0.00
37	Benzo(k)fluoranthene	1.542	1.335	13.4	91	0.00
38 C	Benzo(a)pyrene	1.283	1.162	9.4	92	0.00
39	Dibenzo(a,h)anthracene	1.249	1.194	4.4	93	0.00
40	Benzo(a,h,i)perylene	1.349	1.244	7.8	91	-0.02

Data Path : Z:\HPCHEM1\BNA M\DATA\BM111615\
Data File : BM003136.D
Acq On : 16 Nov 2015 10:47
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Nov 17 13:42:35 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM110415.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 17 12:50:51 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0