

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM111616\
 Data File : BM007931.D
 Acq On : 16 Nov 2016 21:00
 Operator : UM/SJ
 Sample : H5622-24
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 A4WM1

Quant Time: Nov 17 04:44:41 2016
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM102516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 17 04:42:29 2016
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.74	152	982	0.40	ng/ul	-0.02
2) Naphthalene-d8	10.51	136	3161	0.40	ng/ul	-0.02
6) Acenaphthene-d10	14.37	164	1868	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.21	188	266957	0.40	ng/ul	0.09
16) Chrysene-d12	21.32	240	3304	0.40	ng/ul	0.00
20) Perylene-d12	23.40	264	231303	0.40	ng/ul	-0.17
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.12	152	1386	0.37	ng/ul	-0.01
14) Fluoranthene-d10	0.00	212	0	0.00	ng/ul	
Target Compounds						
17) Pyrene	19.51	202	611	0.05	ng/ul#	1




(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM111616\
Data File : BM007931.D
Acq On : 16 Nov 2016 21:00
Operator : UM/SJ
Sample : H5622-24
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
A4WM1

Quant Time: Nov 17 04:44:41 2016
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM102516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 17 04:42:29 2016
Response via : Initial Calibration





#17

Pyrene

Concen: 0.05 ng/ul

RT: 19.51 min Scan# 1178

Delta R.T. -0.04 min

Lab File: BM007931.D

Acq: 16 Nov 2016 21:00

Instrument :

BNA_M

ClientSampleId :

A4WM1

Tgt Ion: 202 Resp: 611

Ion Ratio Lower Upper

202 100

200 17.1 18.2 27.4#

203 199.7 15.6 23.4#

