


Data Path : Z:\HPCHEM1\BNA_M\DATA\BM111616\
Data File : BM007911.D
Acq On : 16 Nov 2016 09:09
Operator : UM/SJ
Sample : DFTPP50
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP50

Integration File: LSCINT.P

Method : Z:\HPCHEM1\BNA_M\METHODS\SOM02.2-EPA-BM041416.M
Title : SVOA CALIBRATION
Last Update : Sat Apr 23 01:00:10 2016



AutoFind: Scans 2543, 2544, 2545; Background Corrected with Scan 2538

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.0	47410	PASS
68	69	0.00	2	2.0	1064	PASS
69	198	0.00	100	36.5	52490	PASS
70	69	0.00	2	0.5	287	PASS
127	198	10	80	43.3	62168	PASS
197	198	0.00	2	0.5	788	PASS
198	198	100	100	100.0	143709	PASS
199	198	5	9	7.0	10100	PASS
275	198	10	60	30.7	44138	PASS
365	198	1	100	4.5	6520	PASS
441	443	0.01	100	79.8	19008	PASS
442	198	50	100	84.9	121938	PASS
443	442	15	24	19.5	23808	PASS

Scan 2538 (17.515 min): BM007911.D
DFTPP50

m/z	Abundance
40.00	361.0
44.00	686.0
73.00	473.0
147.00	307.0
207.00	1606.0
207.90	485.0
281.20	528.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

m/z	Abundance
38.10	691.0
39.10	3653.0
39.90	658.0
43.90	742.0
49.00	499.0
50.10	15707.0
51.10	55248.0
52.10	3157.0
55.00	346.0
56.00	1504.0
57.00	3930.0
61.00	992.0
62.10	938.0
63.10	2449.0
64.10	434.0
65.00	1252.0
68.10	1181.0
69.00	60936.0
70.00	501.0
73.00	1113.0
74.10	6520.0
75.10	11263.0
76.00	3328.0
77.10	70272.0
78.10	4738.0
79.00	6056.0
80.00	3643.0
81.00	5864.0
82.10	1023.0
83.10	970.0
83.90	317.0
85.10	918.0
86.00	1577.0
87.00	883.0
91.00	1449.0
92.10	1347.0
93.10	9652.0
94.10	543.0
96.10	428.0
98.00	7045.0
99.00	6307.0
100.10	515.0
101.00	2587.0
103.00	768.0
104.10	1831.0
105.00	1799.0
106.10	800.0
107.00	21472.0
108.00	3036.0
109.00	742.0
110.00	37112.0
111.00	6807.0
111.90	459.0
116.10	1490.0
117.00	20712.0
118.00	1457.0
122.00	2280.0
123.00	2499.0
124.00	1235.0
125.00	834.0
126.20	642.0
127.00	70176.0
128.00	5651.0
129.00	28968.0
130.00	3292.0
130.90	549.0
133.00	345.0
134.00	757.0
135.00	2065.0
136.00	979.0
137.10	1664.0
141.00	4188.0
141.90	1119.0
143.10	966.0
145.80	605.0
147.00	2489.0
148.00	4989.0
148.90	1307.0
149.90	342.0
151.20	587.0
153.10	1001.0
154.10	1017.0
155.00	2287.0
156.00	3405.0
157.00	452.0
158.00	793.0
158.90	922.0
160.00	1298.0
161.10	1712.0
162.00	421.0
164.00	301.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

165.00	1247.0
166.00	1220.0
167.00	8289.0
168.00	5618.0
169.00	920.0
170.20	322.0
171.00	563.0
172.00	805.0
173.00	866.0
174.00	1882.0
175.10	3170.0
176.10	1135.0
177.00	1463.0
177.90	555.0
179.00	6453.0
180.00	4713.0
181.00	1990.0
182.10	473.0
184.00	494.0
185.00	3020.0
186.00	23160.0
187.00	7514.0
188.10	1133.0
189.00	1603.0
191.00	1126.0
192.00	2220.0
193.00	2065.0
193.90	429.0
196.00	3270.0
197.10	939.0
198.00	152256.0
199.00	10280.0
200.00	841.0
201.60	778.0
203.00	1394.0
204.10	5793.0
205.00	9693.0
206.10	43560.0
207.10	7117.0
208.00	1868.0
208.90	554.0
210.00	692.0
211.00	1514.0
215.00	403.0
215.90	749.0
217.00	11899.0
217.90	1765.0
220.20	382.0
221.00	4962.0
221.90	1786.0
223.00	2458.0
224.10	23912.0
225.00	6139.0
226.00	785.0
227.00	11066.0
228.00	1613.0
229.00	2218.0
229.80	332.0
231.00	771.0
233.90	720.0
235.00	1035.0
235.90	627.0
237.00	1206.0
239.00	456.0
239.90	313.0
240.90	503.0
242.00	1397.0
243.00	1299.0
244.10	17456.0
245.00	2139.0
246.00	4290.0
247.00	872.0
248.90	764.0
254.10	472.0
255.00	98960.0
256.00	14212.0
257.00	1245.0
258.00	6411.0
259.00	984.0
265.00	2326.0
271.00	324.0
273.10	2902.0
274.00	6933.0
275.00	42376.0
276.00	5187.0
277.00	4625.0
278.00	768.0
280.90	696.0
282.90	618.0
284.90	791.0
293.00	1060.0
293.90	301.0
296.00	13161.0
297.00	1555.0
303.00	2108.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

303.90	433.0
313.90	510.0
315.00	1582.0
316.00	1156.0
321.10	444.0
323.00	4227.0
324.00	617.0
327.00	1094.0
328.00	389.0
331.90	376.0
333.00	560.0
334.10	2517.0
335.00	597.0
341.00	464.0
346.00	742.0
352.00	1106.0
353.00	617.0
354.00	1006.0
355.00	371.0
365.00	5732.0
366.10	567.0
372.10	1906.0
383.00	414.0
401.90	559.0
403.10	858.0
404.00	555.0
420.90	632.0
421.90	658.0
423.10	6060.0
424.00	2115.0
441.10	15626.0
442.00	97680.0
443.10	20880.0
444.00	1488.0

Instrument :

BNA_M

ClientSampleId :

DFTPP50

m/z	Abundance
38.00	913.0
39.10	4176.0
40.00	611.0
43.90	770.0
49.10	751.0
50.10	14916.0
51.10	55920.0
52.10	2872.0
54.90	426.0
56.00	2085.0
57.00	3794.0
61.00	760.0
61.90	621.0
63.10	2650.0
64.00	486.0
65.00	1000.0
68.10	1392.0
69.00	61368.0
69.90	361.0
73.00	1218.0
74.10	6869.0
75.00	10405.0
76.00	3448.0
77.10	68000.0
78.10	4353.0
79.00	5392.0
80.00	3951.0
81.00	5400.0
82.00	1436.0
83.00	1053.0
84.00	482.0
84.90	909.0
86.00	1590.0
87.00	696.0
88.00	373.0
91.00	1235.0
92.00	1708.0
93.00	9933.0
94.00	662.0
96.10	543.0
98.00	6873.0
99.00	7001.0
100.10	665.0
101.00	2636.0
103.00	835.0
104.00	2315.0
105.00	1837.0
105.90	709.0
107.00	23096.0
108.00	4118.0
109.20	677.0
110.00	39440.0
111.00	6088.0
112.10	1053.0
116.10	1665.0
117.00	22744.0
118.00	1899.0
118.90	475.0
119.90	313.0
122.00	1449.0
123.00	2507.0
124.00	1023.0
125.00	1020.0
126.20	571.0
127.00	69672.0
128.10	5444.0
129.00	29352.0
130.00	2550.0
131.00	566.0
134.00	611.0
135.00	2870.0
136.10	951.0
137.10	1595.0
140.00	561.0
141.00	4731.0
142.00	1545.0
143.00	916.0
145.10	301.0
146.00	774.0
147.00	2395.0
148.00	4970.0
149.00	1497.0
151.00	445.0
151.90	371.0
153.00	1478.0
154.00	1095.0
155.00	2331.0
156.10	3771.0
157.00	744.0
158.00	707.0
159.00	698.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

160.00	1570.0
161.00	2130.0
162.00	606.0
164.00	342.0
165.00	1938.0
166.10	1296.0
167.10	9644.0
168.00	5736.0
169.00	687.0
170.10	331.0
170.90	436.0
172.00	792.0
173.00	1118.0
174.00	1895.0
175.00	3912.0
176.00	1035.0
177.00	1717.0
178.10	625.0
179.00	6992.0
180.00	5206.0
181.00	2973.0
182.10	438.0
183.00	386.0
184.10	624.0
185.00	3941.0
186.00	24792.0
187.00	6514.0
188.00	977.0
188.90	1849.0
190.00	313.0
190.90	744.0
192.10	2082.0
193.00	2950.0
193.90	693.0
195.00	367.0
196.00	4061.0
197.00	839.0
198.00	166336.0
199.00	12074.0
200.00	898.0
201.40	625.0
202.90	1099.0
204.00	6788.0
205.00	11944.0
206.10	48208.0
207.10	7800.0
208.00	2154.0
208.90	752.0
210.00	923.0
211.00	2202.0
214.90	532.0
216.10	1265.0
217.00	13479.0
218.00	1615.0
221.00	5673.0
222.00	1385.0
223.00	2847.0
224.00	27848.0
225.10	7631.0
226.00	1056.0
227.00	12531.0
228.00	2385.0
229.00	2773.0
230.00	443.0
231.00	1100.0
234.00	897.0
235.00	985.0
235.90	647.0
237.00	1587.0
239.00	533.0
239.90	552.0
241.00	865.0
242.00	1573.0
243.00	1768.0
244.10	21880.0
245.00	2623.0
246.00	4640.0
247.00	1006.0
248.90	764.0
251.10	371.0
252.90	359.0
254.10	703.0
255.00	121424.0
256.00	17096.0
257.00	1673.0
258.10	7176.0
258.90	1138.0
259.90	343.0
265.00	3157.0
266.00	552.0
273.00	3582.0
274.00	10068.0
275.00	54352.0
276.10	7698.0
277.00	5321.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

278.00	847.0
281.00	601.0
283.00	523.0
285.10	784.0
291.90	302.0
293.00	967.0
295.20	431.0
296.00	17608.0
297.00	2656.0
301.10	300.0
303.00	1951.0
304.10	400.0
308.00	350.0
314.00	1052.0
315.00	2019.0
316.00	1146.0
321.00	609.0
323.00	4551.0
324.00	952.0
327.00	1156.0
328.00	632.0
332.00	437.0
332.90	465.0
334.00	3065.0
335.00	751.0
341.00	443.0
346.00	760.0
352.00	1427.0
353.00	1459.0
354.00	1736.0
354.90	421.0
365.00	7505.0
366.10	1129.0
370.00	303.0
371.00	525.0
372.00	2230.0
373.00	659.0
382.90	691.0
384.10	308.0
390.00	351.0
402.10	1034.0
403.10	1218.0
403.90	602.0
421.00	1276.0
422.00	1215.0
423.00	8605.0
424.00	2514.0
441.00	23024.0
442.00	148288.0
443.00	27976.0
444.00	2829.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

m/z	Abundance
38.10	580.0
39.10	2442.0
40.00	393.0
44.00	681.0
50.00	7905.0
51.10	31064.0
52.10	1900.0
56.00	841.0
57.00	1891.0
61.00	487.0
62.00	531.0
63.00	1567.0
65.00	803.0
68.00	621.0
69.00	35168.0
73.00	1027.0
74.10	4006.0
75.00	6002.0
76.10	2124.0
77.10	41576.0
78.10	2452.0
79.00	3724.0
80.00	2590.0
81.00	2982.0
82.00	736.0
83.00	609.0
84.90	490.0
85.90	1263.0
87.10	396.0
91.00	682.0
91.90	892.0
93.00	5433.0
94.00	375.0
96.00	322.0
98.00	4732.0
99.00	3965.0
100.00	317.0
101.00	1751.0
102.90	861.0
104.00	1556.0
105.00	1309.0
106.10	372.0
107.00	15019.0
108.00	1974.0
109.10	362.0
110.00	24768.0
111.00	3374.0
111.90	638.0
116.10	856.0
117.00	14682.0
118.00	1131.0
121.90	975.0
123.00	1709.0
124.00	827.0
125.00	735.0
127.00	46656.0
128.00	3312.0
129.00	19488.0
130.00	1947.0
130.80	349.0
132.90	336.0
133.90	760.0
135.00	1440.0
135.90	728.0
136.90	1012.0
140.10	354.0
141.00	2392.0
142.10	1069.0
143.10	620.0
146.00	533.0
147.00	1848.0
148.00	3944.0
149.10	762.0
150.90	325.0
153.00	850.0
154.00	618.0
155.10	1581.0
156.00	2297.0
156.90	863.0
157.90	541.0
158.90	494.0
160.00	835.0
161.10	931.0
161.90	419.0
165.00	1183.0
166.00	817.0
167.00	5667.0
168.00	3872.0
169.00	393.0
172.00	514.0
173.10	603.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

174.00	947.0
175.10	1688.0
176.00	714.0
177.00	1246.0
179.00	5202.0
180.00	3343.0
181.00	1482.0
184.10	313.0
185.00	2302.0
186.00	17160.0
187.00	4823.0
188.00	622.0
189.00	1585.0
191.10	887.0
192.00	1847.0
193.00	2085.0
194.00	428.0
196.00	2893.0
196.90	588.0
198.00	112536.0
199.00	7948.0
200.00	505.0
201.70	377.0
203.00	1331.0
204.10	4552.0
205.00	8716.0
206.10	31368.0
207.10	5357.0
208.00	1488.0
209.00	786.0
210.10	647.0
211.10	1695.0
215.10	561.0
216.10	698.0
217.00	9214.0
218.00	1342.0
221.10	4014.0
221.90	955.0
223.10	2056.0
224.10	20072.0
225.10	4815.0
226.10	471.0
227.00	8150.0
228.00	1494.0
229.10	1975.0
231.10	927.0
233.00	306.0
234.00	624.0
235.00	717.0
236.00	554.0
237.00	1205.0
238.90	572.0
240.00	377.0
241.00	348.0
242.00	1058.0
243.00	1010.0
244.10	15773.0
245.00	2229.0
246.00	3399.0
247.10	569.0
249.00	510.0
251.00	366.0
253.90	534.0
254.20	564.0
255.00	86328.0
256.00	12358.0
257.00	1047.0
258.00	5798.0
259.00	801.0
265.00	3032.0
266.00	324.0
270.90	369.0
273.00	2907.0
274.00	6697.0
275.00	35688.0
276.00	5274.0
277.00	3526.0
278.00	572.0
281.10	488.0
282.90	642.0
284.00	433.0
285.00	692.0
293.10	936.0
294.00	391.0
296.00	12180.0
296.90	1539.0
303.00	1112.0
304.10	379.0
313.90	547.0
315.00	1216.0
316.00	1108.0
320.90	649.0
323.10	3636.0
324.00	517.0
327.00	724.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50

327.90	357.0
333.00	332.0
334.10	2556.0
335.00	496.0
341.00	358.0
346.00	1110.0
352.00	1249.0
353.10	796.0
354.00	1216.0
355.00	352.0
365.00	6324.0
366.00	997.0
370.90	300.0
372.00	2093.0
383.10	389.0
402.10	889.0
402.90	1311.0
404.10	408.0
421.00	832.0
422.10	828.0
423.10	6533.0
424.10	2416.0
441.00	18376.0
442.10	119848.0
443.00	22568.0
444.00	2474.0

Instrument :
BNA_M
ClientSampleId :
DFTPP50