

Data Path : Z:\HPCHEM1\BNA M\DATA\BM111715\
 Data File : BM003157.D
 Acq On : 16 Nov 2015 23:50
 Operator : UM/IZ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleID :
 SSTD02043

Quant Time: Nov 17 02:32:31 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM103015.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 17 00:18:48 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.75	152	177706	20.00	ng/ul	0.00
18) Naphthalene-d8	10.54	136	825981	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.39	164	483205	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.14	188	996429	20.00	ng/ul	0.00
78) Chrysene-d12	21.33	240	840520	20.00	ng/ul	0.00
86) Perylene-d12	23.59	264	830648	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.20	96	24105	6.39	ng/uL	0.00
5) Phenol-d5	6.91	99	245438	17.47	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.07	67	135168	16.46	ng/ul	0.00
9) 2-Chlorophenol-d4	7.27	132	209517	17.60	ng/ul	0.00
13) 4-Methylphenol-d8	8.45	113	208438	18.40	ng/ul	0.00
19) Nitrobenzene-d5	8.90	128	102910	20.43	ng/ul	0.00
22) 2-Nitrophenol-d4	9.62	143	110545	21.66	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.16	165	211350	18.47	ng/ul	0.00
29) 4-Chloroaniline-d4	10.67	131	271986	18.37	ng/ul	0.00
44) Dimethylphthalate-d6	13.80	166	634904	17.47	ng/ul	0.00
47) Acenaphthylene-d8	14.09	160	784704	17.03	ng/ul	0.00
52) 4-Nitrophenol-d4	14.59	143	114206	18.64	ng/ul	0.00
58) Fluorene-d10	15.39	176	538114	17.12	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.50	200	85158	18.45	ng/ul	0.00
71) Anthracene-d10	17.23	188	762607	17.26	ng/ul	0.00
79) Pyrene-d10	19.53	212	730109	17.65	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.44	264	699392	17.72	ng/ul	0.00

Target Compounds

					Ovalue
2) 1,4-Dioxane	3.24	88	25353	6.16	ng/ul 92
4) Benzaldehyde	6.89	77	107349	14.87	ng/ul 96
6) Phenol	6.93	94	251763	17.41	ng/ul 93
8) Bis(2-Chloroethyl)ether	7.17	93	194623	16.87	ng/ul 100
10) 2-Chlorophenol	7.31	128	214136	17.42	ng/ul 99
11) 2-Methylphenol	8.18	108	202311	17.94	ng/ul 98
12) 2,2'-oxybis(1-Chloropropan	8.28	45	219700	15.49	ng/ul# 92
14) Acetophenone	8.56	105	317187	18.44	ng/ul 96
15) N-Nitroso-di-n-propylamine	8.55	70	154459	18.13	ng/ul 93
16) 4-Methylphenol	8.51	108	223051	18.31	ng/ul 98
17) Hexachloroethane	8.82	117	80846	17.28	ng/ul 95
20) Nitrobenzene	8.95	77	223258	18.29	ng/ul 95
21) Isophorone	9.46	82	436998	17.65	ng/ul 96
23) 2-Nitrophenol	9.65	139	121101	20.43	ng/ul# 88
24) 2,4-Dimethylphenol	9.71	107	245511	17.73	ng/ul 92
25) Bis(2-Chloroethoxy)methane	9.95	93	275047	17.20	ng/ul 98
27) 2,4-Dichlorophenol	10.18	162	214192	18.24	ng/ul 99
28) Naphthalene	10.59	128	701380	16.92	ng/ul 100
30) 4-Chloroaniline	10.69	127	282404	18.20	ng/ul 99
31) Hexachlorobutadiene	10.87	225	130818	17.99	ng/ul 99
32) Caprolactam	11.45	113	66875	20.12	ng/ul 95
33) 4-Chloro-3-methylphenol	11.82	107	229327	18.48	ng/ul 96
34) 2-Methylnaphthalene	12.20	142	521154	17.44	ng/ul 100

Data Path : Z:\HPCHEM1\BNA M\DATA\BM111715\
 Data File : BM003157.D
 Acq On : 16 Nov 2015 23:50
 Operator : UM/IZ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleID :
 SSTD02043

Quant Time: Nov 17 02:32:31 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM103015.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Nov 17 00:18:48 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
35) 1-Methylnaphthalene	12.20	142	521154	17.79	ng/ul	95
37) 1,2,4,5-Tetrachlorobenzene	12.57	216	256177	16.57	ng/ul	99
38) Hexachlorocyclopentadiene	12.56	237	127450	14.01	ng/ul	97
39) 2,4,6-Trichlorophenol	12.82	196	168830	17.89	ng/ul	100
40) 2,4,5-Trichlorophenol	12.89	196	183692	18.10	ng/ul	99
41) 1,1'-Biphenyl	13.22	154	675442	16.59	ng/ul	99
42) 2-Chloronaphthalene	13.26	162	505198	16.68	ng/ul	99
43) 2-Nitroaniline	13.47	65	126279	20.74	ng/ul	93
45) Dimethylphthalate	13.85	163	636071	17.17	ng/ul	100
46) 2,6-Dinitrotoluene	13.97	165	126594	20.91	ng/ul#	86
48) Acenaphthylene	14.11	152	830300	16.88	ng/ul	100
49) 3-Nitroaniline	14.30	138	130472	20.37	ng/ul	97
50) Acenaphthene	14.46	153	540655	16.32	ng/ul	99
51) 2,4-Dinitrophenol	14.50	184	60005	17.84	ng/ul	95
53) 4-Nitrophenol	14.60	109	85105	18.27	ng/ul#	74
54) Dibenzofuran	14.79	168	768893	16.60	ng/ul	94
55) 2,4-Dinitrotoluene	14.76	165	179088	21.02	ng/ul	93
56) 2,3,4,6-Tetrachlorophenol	15.02	232	154354	18.43	ng/ul	98
57) Diethylphthalate	15.22	149	644495	17.64	ng/ul	99
59) Fluorene	15.44	166	601825	16.72	ng/ul	98
60) 4-Chlorophenyl-phenylether	15.44	204	295412	17.48	ng/ul	98
61) 4-Nitroaniline	15.46	138	130498	19.02	ng/ul#	79
64) 4,6-Dinitro-2-methylphenol	15.52	198	91218	18.24	ng/ul#	99
65) N-Nitrosodiphenylamine	15.65	169	525689	17.57	ng/ul	99
66) 4-Bromophenyl-phenylether	16.33	248	180397	18.60	ng/ul	97
67) Hexachlorobenzene	16.44	284	203879	17.34	ng/ul	99
68) Atrazine	16.60	200	190785	18.46	ng/ul	97
69) Pentachlorophenol	16.79	266	115965	17.59	ng/ul	98
70) Phenanthrene	17.18	178	920267	16.31	ng/ul	99
72) Anthracene	17.27	178	923838	16.76	ng/ul	99
73) 1,2,3,4-Tetrachlorobenzene	13.19	216	253514	16.78	ng/uL	98
74) Pentachlorobenzene	14.71	250	245296	17.17	ng/uL	99
75) Carbazole	17.54	167	796045	16.91	ng/ul	99
76) Di-n-butylphthalate	18.10	149	1019080	18.80	ng/ul	99
77) Fluoranthene	19.20	202	927944	16.27	ng/ul	96
80) Pvrene	19.56	202	949591	17.27	ng/ul#	93
81) Butylbenzylphthalate	20.46	149	398799	24.58	ng/ul	93
82) 3,3'-Dichlorobenzidine	21.24	252	238692	20.32	ng/ul	99
83) Benzo(a)anthracene	21.31	228	818447	17.73	ng/ul	99
84) Bis(2-ethylhexyl)phthalate	21.24	149	590223	23.26	ng/ul	100
85) Chrysene	21.36	228	763544	16.72	ng/ul	99
87) Di-n-octyl phthalate	22.13	149	982876	22.98	ng/ul	100
88) Benzo(b)fluoranthene	22.91	252	805616	16.84	ng/ul#	97
89) Benzo(k)fluoranthene	22.95	252	787501	17.64	ng/ul#	96
91) Benzo(a)pyrene	23.49	252	777545	17.16	ng/ul#	96
92) Indeno(1,2,3-cd)pyrene	25.87	276	928841	18.64	ng/ul#	93
93) Dibenzo(a,h)anthracene	25.89	278	785496	20.00	ng/ul#	93
94) Benzo(a,h,i)perylene	26.57	276	779018	17.38	ng/ul#	93


Data Path : Z:\HPCHEM1\BNA M\DATA\BM111715\
Data File : BM003157.D
Acq On : 16 Nov 2015 23:50
Operator : UM/IZ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD02043

Quant Time: Nov 17 02:32:31 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-BM103015.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Nov 17 00:18:48 2015
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed						



#2

1,4-Dioxane

Concen: 6.16 ng/uL

RT: 3.24 min Scan# 26

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50

Instrument :

BNA_M

ClientSampleId :

SSTD02043


Tot Ion: 88 Resp: 25353

Ion Ratio Lower Upper

88 100

43 24.0 23.9 35.9




58 66.1 57.1 85.7



Abundance Ion 88.00 (87.70 to 88.70): BM003141.D (-21) (-)

Ion 43.00 (42.70 to 43.70): BM003141.D (-21) (-)

Ion 58.00 (57.70 to 58.70): BM003141.D (-21) (-)



#4

Benzaldehyde

Concen: 14.87 ng/uL

RT: 6.89 min Scan# 646

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50


Tgt Ion: 77 Resp: 107349

Ion Ratio Lower Upper

77 100

105 104.2 87.8 131.8



106 101.5 83.4 125.2






Abundance Ion 77.00 (76.70 to 77.70): BM003141.D (-640) (-)

Ion 105.00 (104.70 to 105.70): BM003141.D (-640) (-)

Ion 106.00 (105.70 to 106.70): BM003141.D (-640) (-)





#6

Phenol

Concen: 17.41 ng/ul

RT: 6.93 min Scan# 654

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50

Instrument :

BNA_M

ClientSampleId :

SSTD02043





Tot Ion: 94 Resp: 251763

Ion Ratio Lower Upper

94 100

65 26.1 18.7 28.1

66 34.6 24.2 36.2



#8

Bis(2-Chloroethyl)ether

Concen: 16.87 ng/ul

RT: 7.17 min Scan# 694

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50


Tgt Ion: 93 Resp: 194623


Ion Ratio Lower Upper

93 100

63 69.9 56.0 84.0

95 32.3 26.2 39.2





#10

2-Chlorophenol

Concen: 17.42 ng/ul

RT: 7.31 min Scan# 718

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50

Instrument :

BNA_M

ClientSampleId :

SSTD02043


Tot Ion:128 Resp: 214136

Ion Ratio Lower Upper

128 100

64 39.6 32.2 48.4




130 32.4 26.7 40.1



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 64.00 (63.70 to 64.70): BM0

Ion 130.00 (129.70 to 130.70): E



#11

2-Methylphenol

Concen: 17.94 ng/ul

RT: 8.18 min Scan# 866

Delta R.T. 0.00 min

Lab File: BM003157.D


Acq: 16 Nov 2015 23:50

Tgt Ion:108 Resp: 202311

Ion Ratio Lower Upper



108 100


107 88.1 72.3 108.5



Abundance Ion 108.00 (107.70 to 108.70): E

Ion 107.00 (106.70 to 107.70): E









#12
 2,2'-oxvbis(1-Chloropropane)
 Concen: 15.49 ng/ul
 RT: 8.28 min Scan# 883
 Delta R.T. 0.00 min
 Lab File: BM003157.D
 Acq: 16 Nov 2015 23:50




Instrument :
 BNA_M
 ClientSampleId :
 SSTD02043




Tot Ion: 45 Resp: 219700
 Ion Ratio Lower Upper
 45 100
 77 17.3 11.0 16.6#
 79 14.1 9.0 13.6#



#14
 Acetophenone
 Concen: 18.44 ng/ul
 RT: 8.56 min Scan# 931
 Delta R.T. 0.00 min
 Lab File: BM003157.D
 Acq: 16 Nov 2015 23:50

Tgt Ion: 105 Resp: 317187
 Ion Ratio Lower Upper
 105 100
 77 73.5 55.9 83.9
 51 22.5 17.0 25.4









#15
N-Nitroso-di-n-propylamine
Concen: 18.13 ng/ul
RT: 8.55 min Scan# 929
Delta R.T. 0.00 min
Lab File: BM003157.D
Acq: 16 Nov 2015 23:50


Instrument :
BNA_M
ClientSampleId :
SSTD02043


Tot Ion:	70	Resp:	154459
Ion	Ratio	Lower	Upper
70	100		
42	48.3	43.5	65.3
101	10.6	9.0	13.4
130	23.8	21.4	32.2



#16
4-Methylphenol
Concen: 18.31 ng/ul
RT: 8.51 min Scan# 922
Delta R.T. 0.00 min
Lab File: BM003157.D
Acq: 16 Nov 2015 23:50

Tgt Ion:	108	Resp:	223051
Ion	Ratio	Lower	Upper
108	100		
107	112.2	88.4	132.6





#17

Hexachloroethane

Concen: 17.28 ng/ul

RT: 8.82 min Scan# 975

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50

Instrument :

BNA_M

ClientSampleId :

SSTD02043


Tot Ion:117 Resp: 80846

Ion Ratio Lower Upper

117 100

201 102.4 78.0 117.0




199 64.3 48.1 72.1



Abundance Ion 117.00 (116.70 to 117.70): E

Ion 201.00 (200.70 to 201.70): E

Ion 199.00 (198.70 to 199.70): E



#20

Nitrobenzene

Concen: 18.29 ng/ul

RT: 8.95 min Scan# 996

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50


Tgt Ion: 77 Resp: 223258

Ion Ratio Lower Upper

77 100

123 50.7 44.4 66.6



65 12.8 10.2 15.2




Abundance Ion 77.00 (76.70 to 77.70): BM

Ion 123.00 (122.70 to 123.70): E

Ion 65.00 (64.70 to 65.70): BM





#21

Isophorone

Concen: 17.65 ng/ul

RT: 9.46 min Scan# 1084

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50

Instrument :

BNA_M

ClientSampleId :

SSTD02043


Tot Ion: 82 Resp: 436998

Ion Ratio Lower Upper

82 100

95 7.3 5.3 7.9




138 19.1 17.2 25.8



Abundance Ion 82.00 (81.70 to 82.70): BM003141.D (-1077) (-)

Ion 95.00 (94.70 to 95.70): BM003141.D (-1077) (-)

Ion 138.00 (137.70 to 138.70): BM003141.D (-1077) (-)



#23

2-Nitrophenol

Concen: 20.43 ng/ul

RT: 9.65 min Scan# 1116

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50


Tgt Ion: 139 Resp: 121101

Ion Ratio Lower Upper

139 100

65 45.0 31.0 46.6



109 29.9 18.5 27.7#




Abundance Ion 139.00 (138.70 to 139.70): BM003141.D (-1110) (-)

Ion 65.00 (64.70 to 65.70): BM003141.D (-1110) (-)

Ion 109.00 (108.70 to 109.70): BM003141.D (-1110) (-)





#24

2,4-Dimethylphenol

Concen: 17.73 ng/ul

RT: 9.71 min Scan# 1126

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50

Instrument :

BNA_M

ClientSampleId :

SSTD02043


Tot Ion:107 Resp: 245511

Ion Ratio Lower Upper

107 100

121 51.3 45.4 68.0




122 87.7 76.6 115.0



Abundance Ion 107.00 (106.70 to 107.70): E

Ion 121.00 (120.70 to 121.70): E

Ion 122.00 (121.70 to 122.70): E



#25

Bis(2-Chloroethoxy)methane

Concen: 17.20 ng/ul

RT: 9.95 min Scan# 1167

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50


Tgt Ion: 93 Resp: 275047

Ion Ratio Lower Upper

93 100

95 33.0 25.8 38.6



123 11.9 10.2 15.2




Abundance Ion 93.00 (92.70 to 93.70): BM0

Ion 95.00 (94.70 to 95.70): BM0

Ion 123.00 (122.70 to 123.70): E





#27

2,4-Dichlorophenol

Concen: 18.24 ng/ul

RT: 10.18 min Scan# 1206

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50

Instrument :

BNA_M

ClientSampleId :

SSTD02043


Tot Ion:162 Resp: 214192

Ion Ratio Lower Upper

162 100

164 66.1 51.9 77.9




98 35.3 28.2 42.2



Abundance Ion 162.00 (161.70 to 162.70): E

Ion 164.00 (163.70 to 164.70): E

Ion 98.00 (97.70 to 98.70): BM



#28

Naphthalene

Concen: 16.92 ng/ul

RT: 10.59 min Scan# 1275

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50


Tgt Ion:128 Resp: 701380

Ion Ratio Lower Upper

128 100

129 11.0 8.8 13.2



127 13.2 10.5 15.7




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E









#30
4-Chloroaniline
Concen: 18.20 ng/ul
RT: 10.69 min Scan# 1293
Delta R.T. 0.00 min
Lab File: BM003157.D
Acq: 16 Nov 2015 23:50




Instrument :
BNA_M
ClientSampleId :
SSTD02043




Tot Ion:127 Resp: 282404
Ion Ratio Lower Upper
127 100
129 31.7 25.6 38.4



#31
Hexachlorobutadiene
Concen: 17.99 ng/ul
RT: 10.87 min Scan# 1324
Delta R.T. 0.00 min
Lab File: BM003157.D
Acq: 16 Nov 2015 23:50

Tgt Ion:225 Resp: 130818
Ion Ratio Lower Upper
225 100
223 62.0 50.2 75.4
227 63.3 50.5 75.7





#32

Caprolactam

Concen: 20.12 ng/ul

RT: 11.45 min Scan# 1422

Delta R.T. 0.01 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50

Instrument :

BNA_M

ClientSampleId :

SSTD02043





Tot Ion:113 Resp: 66875

Ion Ratio Lower Upper

113 100

55 131.4 112.9 169.3

56 102.3 82.7 124.1



#33

4-Chloro-3-methylphenol

Concen: 18.48 ng/ul

RT: 11.82 min Scan# 1485

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50


Tgt Ion:107 Resp: 229327


Ion Ratio Lower Upper

107 100

144 26.8 22.8 34.2

142 82.8 69.5 104.3





#34

2-Methylnaphthalene

Concen: 17.44 ng/ul

RT: 12.20 min Scan# 1550

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50

Instrument :

BNA_M

ClientSampleId :





SSTD02043

Tot Ion:142 Resp: 521154

Ion Ratio Lower Upper

142 100

141 88.4 70.5 105.7



#35

1-Methylnaphthalene

Concen: 17.79 ng/ul

RT: 12.20 min Scan# 1550

Delta R.T. -0.22 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50




Tgt Ion:142 Resp: 521154


Ion Ratio Lower Upper

142 100

141 88.4 74.6 112.0

116 3.7 2.6 3.8









#37
 1,2,4,5-Tetrachlorobenzene
 Concen: 16.57 ng/ul
 RT: 12.57 min Scan# 1613
 Delta R.T. 0.00 min
 Lab File: BM003157.D
 Acq: 16 Nov 2015 23:50

Instrument :
 BNA_M
 ClientSampleId :
 SSTD02043

Tot Ion:216	Resp:	256177
Ion Ratio	Lower	Upper
216	100	
214	79.1	63.4 95.0
179	21.4	17.4 26.2
108	17.9	13.0 19.6




Abundance Ion 216.00 (215.70 to 216.70): E
 Ion 214.00 (213.70 to 214.70): E
 Ion 179.00 (178.70 to 179.70): E
 Ion 108.00 (107.70 to 108.70): E






#38
 Hexachlorocyclopentadiene
 Concen: 14.01 ng/ul
 RT: 12.56 min Scan# 1610
 Delta R.T. 0.00 min
 Lab File: BM003157.D
 Acq: 16 Nov 2015 23:50

Tgt Ion:237	Resp:	127450
Ion Ratio	Lower	Upper
237	100	
235	61.9	47.5 71.3
272	12.6	10.3 15.5



Abundance Ion 237.00 (236.70 to 237.70): E
 Ion 235.00 (234.70 to 235.70): E
 Ion 272.00 (271.70 to 272.70): E









#39
 2,4,6-Trichlorophenol
 Concen: 17.89 ng/ul
 RT: 12.82 min Scan# 1654
 Delta R.T. 0.00 min
 Lab File: BM003157.D
 Acq: 16 Nov 2015 23:50




Instrument :
 BNA_M
 ClientSampleID :
 SSTD02043


Tot Ion:196 Resp: 168830
 Ion Ratio Lower Upper
 196 100
 198 95.6 76.2 114.2
 200 30.7 25.0 37.4



#40
 2,4,5-Trichlorophenol
 Concen: 18.10 ng/ul
 RT: 12.89 min Scan# 1666
 Delta R.T. 0.01 min
 Lab File: BM003157.D
 Acq: 16 Nov 2015 23:50

Tgt Ion:196 Resp: 183692
 Ion Ratio Lower Upper
 196 100
 198 95.3 76.7 115.1
 200 30.1 25.5 38.3









#41
 1,1'-Biphenyl
 Concen: 16.59 ng/ul
 RT: 13.22 min Scan# 1723
 Delta R.T. 0.00 min
 Lab File: BM003157.D
 Acq: 16 Nov 2015 23:50




Instrument :
 BNA_M
 ClientSampleId :
 SSTD02043


Tot Ion:154 Resp: 675442
 Ion Ratio Lower Upper
 154 100
 153 39.6 31.9 47.9
 76 11.7 9.8 14.6



#42
 2-Chloronaphthalene
 Concen: 16.68 ng/ul
 RT: 13.26 min Scan# 1730
 Delta R.T. 0.00 min
 Lab File: BM003157.D
 Acq: 16 Nov 2015 23:50

Tgt Ion:162 Resp: 505198
 Ion Ratio Lower Upper
 162 100
 164 32.6 25.8 38.6
 127 39.9 31.7 47.5









#43
2-Nitroaniline
Concen: 20.74 ng/ul
RT: 13.47 min Scan# 1765
Delta R.T. 0.00 min
Lab File: BM003157.D
Acq: 16 Nov 2015 23:50




Instrument :
BNA_M
ClientSampled :
SSTD02043


Tot Ion: 65 Resp: 126279
Ion Ratio Lower Upper
65 100
92 77.5 63.4 95.2
138 127.6 111.0 166.4



#45
Dimethylphthalate
Concen: 17.17 ng/ul
RT: 13.85 min Scan# 1830
Delta R.T. 0.01 min
Lab File: BM003157.D
Acq: 16 Nov 2015 23:50

Tgt Ion: 163 Resp: 636071
Ion Ratio Lower Upper
163 100
194 4.0 3.2 4.8
164 9.8 7.8 11.8









#46
2,6-Dinitrotoluene
Concen: 20.91 ng/ul
RT: 13.97 min Scan# 1850
Delta R.T. 0.00 min
Lab File: BM003157.D
Acq: 16 Nov 2015 23:50




Instrument :
BNA_M
ClientSampleId :
SSTD02043


Tot Ion:165 Resp: 126594
Ion Ratio Lower Upper
165 100
89 54.5 34.6 51.8#
121 21.4 15.4 23.0



#48
Acenaphthylene
Concen: 16.88 ng/ul
RT: 14.11 min Scan# 1874
Delta R.T. 0.00 min
Lab File: BM003157.D
Acq: 16 Nov 2015 23:50

Tgt Ion:152 Resp: 830300
Ion Ratio Lower Upper
152 100
151 19.7 15.8 23.6
153 13.2 10.3 15.5





#49

3-Nitroaniline

Concen: 20.37 ng/ul

RT: 14.30 min Scan# 1906

Delta R.T. 0.01 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50

Instrument :

BNA_M

ClientSampleId :

SSTD02043


Tot Ion:138 Resp: 130472

Ion Ratio Lower Upper

138 100

108 10.3 8.2 12.4




92 104.7 81.0 121.6



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BM



#50

Acenaphthene

Concen: 16.32 ng/ul

RT: 14.46 min Scan# 1933

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50


Tgt Ion:153 Resp: 540655

Ion Ratio Lower Upper

153 100

152 47.3 37.0 55.6



154 89.0 71.0 106.4




Abundance Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E

Ion 154.00 (153.70 to 154.70): E









#51
 2,4-Dinitrophenol
 Concen: 17.84 ng/ul
 RT: 14.50 min Scan# 1941
 Delta R.T. 0.01 min
 Lab File: BM003157.D
 Acq: 16 Nov 2015 23:50




Instrument :
 BNA_M
 ClientSampleId :
 SSTD02043


Tot Ion:184 Resp: 60005
 Ion Ratio Lower Upper
 184 100
 63 52.3 41.1 61.7
 154 63.8 46.0 69.0



#53
 4-Nitrophenol
 Concen: 18.27 ng/ul
 RT: 14.60 min Scan# 1957
 Delta R.T. 0.00 min
 Lab File: BM003157.D
 Acq: 16 Nov 2015 23:50

Tgt Ion:109 Resp: 85105
 Ion Ratio Lower Upper
 109 100
 139 138.6 150.3 225.5#
 65 118.5 110.4 165.6





#54

Dibenzofuran

Concen: 16.60 ng/ul

RT: 14.79 min Scan# 1990

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50

Instrument :

BNA_M

ClientSampleId :





SSTD02043

Tot Ion:168 Resp: 768893

Ion Ratio Lower Upper

168 100

139 38.8 28.3 42.5



#55

2,4-Dinitrotoluene

Concen: 21.02 ng/ul

RT: 14.76 min Scan# 1984

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50




Tgt Ion:165 Resp: 179088


Ion Ratio Lower Upper

165 100

63 36.9 26.2 39.2

182 2.6 2.2 3.4









#56
2,3,4,6-Tetrachlorophenol
Concen: 18.43 ng/ul
RT: 15.02 min Scan# 2028
Delta R.T. 0.00 min
Lab File: BM003157.D
Acq: 16 Nov 2015 23:50

Instrument :
BNA_M
ClientSampleId :
SSTD02043

Tot Ion:232 Resp: 154354
Ion Ratio Lower Upper
232 100
131 48.7 37.0 55.4
130 2.5 1.8 2.6
166 33.0 26.6 39.8




Abundance Ion 232.00 (231.70 to 232.70): E
Ion 131.00 (130.70 to 131.70): E
Ion 130.00 (129.70 to 130.70): E
Ion 166.00 (165.70 to 166.70): E






#57
Diethylphthalate
Concen: 17.64 ng/ul
RT: 15.22 min Scan# 2062
Delta R.T. 0.00 min
Lab File: BM003157.D
Acq: 16 Nov 2015 23:50

Tgt Ion:149 Resp: 644495
Ion Ratio Lower Upper
149 100
177 21.5 17.6 26.4
150 12.6 9.8 14.6



Abundance Ion 149.00 (148.70 to 149.70): E
Ion 177.00 (176.70 to 177.70): E
Ion 150.00 (149.70 to 150.70): E





#59

Fluorene

Concen: 16.72 ng/ul

RT: 15.44 min Scan# 2100

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50

Instrument :

BNA_M

ClientSampleId :

SSTD02043


Tot Ion:166 Resp: 601825

Ion Ratio Lower Upper

166 100

165 98.4 77.2 115.8




167 13.9 10.2 15.4



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#60

4-Chlorophenyl-phenylether

Concen: 17.48 ng/ul

RT: 15.44 min Scan# 2100

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50


Tgt Ion:204 Resp: 295412

Ion Ratio Lower Upper

204 100

206 32.5 26.1 39.1



141 61.7 47.2 70.8




Abundance Ion 204.00 (203.70 to 204.70): E

Ion 206.00 (205.70 to 206.70): E

Ion 141.00 (140.70 to 141.70): E





#61

4-Nitroaniline

Concen: 19.02 ng/ul

RT: 15.46 min Scan# 2104

Delta R.T. 0.01 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50

Instrument :

BNA_M

ClientSampleId :

SSTD02043


Tot Ion:138 Resp: 130498

Ion Ratio Lower Upper

138 100

92 47.9 35.1 52.7




108 77.0 43.2 64.8#



Abundance Ion 138.00 (137.70 to 138.70): E

Ion 92.00 (91.70 to 92.70): BM0

Ion 108.00 (107.70 to 108.70): E



#64

4,6-Dinitro-2-methylphenol

Concen: 18.24 ng/ul

RT: 15.52 min Scan# 2114

Delta R.T. 0.01 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50


Tgt Ion:198 Resp: 91218

Ion Ratio Lower Upper

198 100

182 3.7 3.8 5.6#



77 21.9 17.4 26.0




Abundance Ion 198.00 (197.70 to 198.70): E

Ion 182.00 (181.70 to 182.70): E

Ion 77.00 (76.70 to 77.70): BM0





#65

N-Nitrosodiphenylamine

Concen: 17.57 ng/ul

RT: 15.65 min Scan# 2136

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50

Instrument :

BNA_M

ClientSampleId :

SSTD02043


Tot Ion:169 Resp: 525689

Ion Ratio Lower Upper

169 100

168 66.2 52.2 78.4




167 36.0 28.2 42.2



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E



#66

4-Bromophenyl-phenylether

Concen: 18.60 ng/ul

RT: 16.33 min Scan# 2252

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50


Tgt Ion:248 Resp: 180397

Ion Ratio Lower Upper

248 100

250 95.3 79.0 118.6



141 75.6 62.8 94.2




Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E





#67

Hexachlorobenzene

Concen: 17.34 ng/ul

RT: 16.44 min Scan# 2271

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50

Instrument :

BNA_M

ClientSampleId :

SSTD02043


Tot Ion:284 Resp: 203879

Ion Ratio Lower Upper

284 100

142 36.0 28.3 42.5


249 31.6 25.2 37.8





Abundance Ion 284.00 (283.70 to 284.70): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E



Time-->



#68

Atrazine

Concen: 18.46 ng/ul

RT: 16.60 min Scan# 2298

Delta R.T. 0.01 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50


Tgt Ion:200 Resp: 190785

Ion Ratio Lower Upper

200 100

173 25.7 23.1 34.7


215 47.3 38.9 58.3




Abundance Ion 200.00 (199.70 to 200.70): E


Ion 173.00 (172.70 to 173.70): E

Ion 215.00 (214.70 to 215.70): E



Time-->









#69
 Pentachlorophenol
 Concen: 17.59 ng/ul
 RT: 16.79 min Scan# 2329
 Delta R.T. 0.00 min
 Lab File: BM003157.D
 Acq: 16 Nov 2015 23:50

Instrument :
 BNA_M
 ClientSampleId :
 SSTD02043

Tot Ion:266 Resp: 115965
 Ion Ratio Lower Upper
 266 100
 264 63.0 52.0 78.0
 268 63.8 49.7 74.5




Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 264.00 (263.70 to 264.70): E
 Ion 268.00 (267.70 to 268.70): E






#70
 Phenanthrene
 Concen: 16.31 ng/ul
 RT: 17.18 min Scan# 2396
 Delta R.T. 0.00 min
 Lab File: BM003157.D
 Acq: 16 Nov 2015 23:50

Tgt Ion:178 Resp: 920267
 Ion Ratio Lower Upper
 178 100
 179 15.1 12.2 18.4
 176 19.0 14.7 22.1



Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 179.00 (178.70 to 179.70): E
 Ion 176.00 (175.70 to 176.70): E





#72

Anthracene

Concen: 16.76 ng/uL

RT: 17.27 min Scan# 2411

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50

Instrument :

BNA_M

ClientSampleId :

SSTD02043


Tot Ion:178 Resp: 923838

Ion Ratio Lower Upper

178 100

179 15.3 12.3 18.5




176 18.7 14.4 21.6



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E



#73

1,2,3,4-Tetrachlorobenzene

Concen: 16.78 ng/uL

RT: 13.19 min Scan# 1717

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50


Tgt Ion:216 Resp: 253514

Ion Ratio Lower Upper

216 100

214 78.4 61.8 92.8



218 49.1 38.1 57.1




Abundance Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 218.00 (217.70 to 218.70): E









#74
 Pentachlorobenzene
 Concen: 17.17 ng/uL
 RT: 14.71 min Scan# 1976
 Delta R.T. 0.00 min
 Lab File: BM003157.D
 Acq: 16 Nov 2015 23:50

Instrument :
 BNA_M
 ClientSampleId :
 SSTD02043

Tot Ion:250 Resp: 245296
 Ion Ratio Lower Upper
 250 100
 248 64.0 51.2 76.8
 252 63.2 49.6 74.4




Abundance Ion 250.00 (249.70 to 250.70): E
 Ion 248.00 (247.70 to 248.70): E
 Ion 252.00 (251.70 to 252.70): E






#75
 Carbazole
 Concen: 16.91 ng/uL
 RT: 17.54 min Scan# 2457
 Delta R.T. 0.00 min
 Lab File: BM003157.D
 Acq: 16 Nov 2015 23:50

Tgt Ion:167 Resp: 796045
 Ion Ratio Lower Upper
 167 100
 166 21.1 17.0 25.4
 139 13.0 9.9 14.9



Abundance Ion 167.00 (166.70 to 167.70): E
 Ion 166.00 (165.70 to 166.70): E
 Ion 139.00 (138.70 to 139.70): E





#76

Di-n-butylphthalate

Concen: 18.80 ng/ul

RT: 18.10 min Scan# 2553

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50

Instrument :

BNA_M

ClientSampleId :

SSTD02043


Tot Ion:149 Resp: 1019080

Ion Ratio Lower Upper

149 100

150 9.1 7.4 11.0




104 4.8 3.2 4.8



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#77

Fluoranthene

Concen: 16.27 ng/ul

RT: 19.20 min Scan# 2739

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50


Tgt Ion:202 Resp: 927944

Ion Ratio Lower Upper

202 100

101 11.1 10.4 15.6



100 8.2 7.8 11.6




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): B









#80
Pvrene
Concen: 17.27 ng/ul
RT: 19.56 min Scan# 2801
Delta R.T. 0.00 min
Lab File: BM003157.D
Acq: 16 Nov 2015 23:50




Instrument :
BNA_M
ClientSampleId :
SSTD02043


Tot Ion	Ratio	Lower	Upper
202	100		
101	12.7	12.6	18.8
100	10.2	10.6	15.8



#81
Butylbenzylphthalate
Concen: 24.58 ng/ul
RT: 20.46 min Scan# 2954
Delta R.T. -0.01 min
Lab File: BM003157.D
Acq: 16 Nov 2015 23:50

Tgt Ion	Ratio	Lower	Upper
149	100		
91	69.3	49.7	74.5
206	20.3	15.9	23.9









#82
 3,3'-Dichlorobenzidine
 Concen: 20.32 ng/ul
 RT: 21.24 min Scan# 3087
 Delta R.T. -0.01 min
 Lab File: BM003157.D
 Acq: 16 Nov 2015 23:50

Instrument :
 BNA_M
 ClientSampleId :
 SSTD02043

Tot Ion	Ratio	Lower	Upper
252	100		
254	64.4	51.3	76.9
126	12.2	10.9	16.3




Abundance Ion 252.00 (251.70 to 252.70): E
 Ion 254.00 (253.70 to 254.70): E
 Ion 126.00 (125.70 to 126.70): E






#83
 Benzo(a)anthracene
 Concen: 17.73 ng/ul
 RT: 21.31 min Scan# 3098
 Delta R.T. -0.01 min
 Lab File: BM003157.D
 Acq: 16 Nov 2015 23:50

Tgt Ion	Ratio	Lower	Upper
228	100		
229	19.7	15.6	23.4
226	26.8	21.0	31.6



Abundance Ion 228.00 (227.70 to 228.70): E
 Ion 229.00 (228.70 to 229.70): E
 Ion 226.00 (225.70 to 226.70): E









#84
 Bis(2-ethylhexyl)phthalate
 Concen: 23.26 ng/ul
 RT: 21.24 min Scan# 3087
 Delta R.T. -0.01 min
 Lab File: BM003157.D
 Acq: 16 Nov 2015 23:50




Instrument :
 BNA_M
 ClientSampleId :
 SSTD02043


Tot Ion:149 Resp: 590223
 Ion Ratio Lower Upper
 149 100
 167 27.5 22.0 33.0
 279 4.1 3.2 4.8



#85
 Chrysene
 Concen: 16.72 ng/ul
 RT: 21.36 min Scan# 3107
 Delta R.T. -0.01 min
 Lab File: BM003157.D
 Acq: 16 Nov 2015 23:50

Tgt Ion:228 Resp: 763544
 Ion Ratio Lower Upper
 228 100
 226 29.9 23.3 34.9
 229 19.6 15.8 23.6









#87
 Di-n-octyl phthalate
 Concen: 22.98 ng/ul
 RT: 22.13 min Scan# 3237
 Delta R.T. -0.01 min
 Lab File: BM003157.D
 Acq: 16 Nov 2015 23:50

Instrument :
 BNA_M
 ClientSampleId :
 SSTD02043

Tgt Ion:149 Resp: 982876




Abundance Ion 149.00 (148.70 to 149.70): E






#88
 Benzo(b)fluoranthene
 Concen: 16.84 ng/ul
 RT: 22.91 min Scan# 3370
 Delta R.T. 0.00 min
 Lab File: BM003157.D
 Acq: 16 Nov 2015 23:50

Tgt Ion:252 Resp: 805616
 Ion Ratio Lower Upper
 252 100
 253 21.6 17.4 26.2
 125 10.2 10.7 16.1#



Abundance Ion 252.00 (251.70 to 252.70): E





#89

Benzo(k)fluoranthene

Concen: 17.64 ng/ul

RT: 22.95 min Scan# 3377

Delta R.T. -0.01 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50

Instrument :

BNA_M

ClientSampleId :

SSTD02043


Tot Ion:252 Resp: 787501

Ion Ratio Lower Upper

252 100

253 22.0 17.4 26.0




125 10.2 11.1 16.7#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#91

Benzo(a)pyrene

Concen: 17.16 ng/ul

RT: 23.49 min Scan# 3469

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50


Tgt Ion:252 Resp: 777545

Ion Ratio Lower Upper

252 100

253 21.7 17.1 25.7



125 10.8 11.9 17.9#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#92

Indeno(1,2,3-cd)pyrene

Concen: 18.64 ng/ul

RT: 25.87 min Scan# 3874

Delta R.T. -0.01 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50

Instrument :

BNA_M

ClientSampleId :

SSTD02043


Tot Ion:276 Resp: 928841

Ion Ratio Lower Upper

276 100

138 26.4 26.6 40.0#




277 24.8 20.2 30.4



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E



#93

Dibenzo(a,h)anthracene

Concen: 20.00 ng/ul

RT: 25.89 min Scan# 3876

Delta R.T. -0.01 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50


Tgt Ion:278 Resp: 785496

Ion Ratio Lower Upper

278 100

139 16.6 18.8 28.2#



279 23.7 19.3 28.9




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#94

Benzo(a,h,i)perylene

Concen: 17.38 ng/ul

RT: 26.57 min Scan# 3993

Delta R.T. 0.00 min

Lab File: BM003157.D

Acq: 16 Nov 2015 23:50

Instrument :

BNA_M

ClientSampleId :

SSTD02043

Tot Ion: 276 Resp: 779018

Ion Ratio Lower Upper

276 100

138 22.9 23.8 35.6#

277 24.1 19.4 29.2

