

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111722\
 Data File : BM037527.D
 Acq On : 17 Nov 2022 15:11
 Operator : CG/JU
 Sample : N5622-07
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BGCA9

Quant Time: Nov 17 15:52:57 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM111622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 17 00:46:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.140	152	6487	0.400	ng/ul	0.00
4) Naphthalene-d8	10.963	136	19064	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.760	164	12066	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.495	188	25681	0.400	ng/ul	0.00
17) Chrysene-d12	21.655	240	18203	0.400	ng/ul	0.00
23) Perylene-d12	24.154	264	21472	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.453	96	31985	4.241	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.536	152	10649	0.337	ng/ul	0.00
18) Fluoranthene-d10	19.508	212	25144	0.418	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.491	88	1517	0.193	ng/ul#	63

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111722\
Data File : BM037527.D
Acq On : 17 Nov 2022 15:11
Operator : CG/JU
Sample : N5622-07
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BGCA9

Quant Time: Nov 17 15:52:57 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM111622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 17 00:46:50 2022
Response via : Initial Calibration

