

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM111722\  
 Data File : BM037536.D  
 Acq On : 17 Nov 2022 21:17  
 Operator : CG/JU  
 Sample : N5622-11  
 Misc :  
 ALS Vial : 20 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 BGCB3

Quant Time: Nov 17 22:52:40 2022  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\SFAM-EPA-SIM-BM111622.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Thu Nov 17 00:46:50 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc  | Units  | Dev(Min) |
|-----------------------------|--------|------|----------|-------|--------|----------|
| -----                       |        |      |          |       |        |          |
| Internal Standards          |        |      |          |       |        |          |
| 1) 1,4-Dichlorobenzene-d4   | 8.141  | 152  | 5949     | 0.400 | ng/ul  | 0.00     |
| 4) Naphthalene-d8           | 10.963 | 136  | 17410    | 0.400 | ng/ul  | # 0.00   |
| 9) Acenaphthene-d10         | 14.760 | 164  | 11078    | 0.400 | ng/ul  | 0.00     |
| 13) Phenanthrene-d10        | 17.491 | 188  | 22982    | 0.400 | ng/ul  | 0.00     |
| 17) Chrysene-d12            | 21.649 | 240  | 15986    | 0.400 | ng/ul  | # 0.00   |
| 23) Perylene-d12            | 24.145 | 264  | 19112    | 0.400 | ng/ul  | 0.00     |
| System Monitoring Compounds |        |      |          |       |        |          |
| 3) 1,4-Dioxane-d8           | 3.453  | 96   | 42351    | 6.123 | ng/ul  | 0.00     |
| 6) 2-Methylnaphthalene-d10  | 12.536 | 152  | 10149    | 0.352 | ng/ul  | 0.00     |
| 18) Fluoranthene-d10        | 19.503 | 212  | 22826    | 0.432 | ng/ul  | 0.00     |
| Target Compounds            |        |      |          |       |        |          |
|                             |        |      |          |       |        | Qvalue   |
| 2) 1,4-Dioxane              | 3.491  | 88   | 1962     | 0.273 | ng/ul# | 60       |
| -----                       |        |      |          |       |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM111722\  
Data File : BM037536.D  
Acq On : 17 Nov 2022 21:17  
Operator : CG/JU  
Sample : N5622-11  
Misc :  
ALS Vial : 20 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
BGCB3

Quant Time: Nov 17 22:52:40 2022  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\SFAM-EPA-SIM-BM111622.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Thu Nov 17 00:46:50 2022  
Response via : Initial Calibration

