

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111722\
 Data File : BM037522.D
 Acq On : 17 Nov 2022 12:10
 Operator : CG/JU
 Sample : N5622-02
 Misc :
 ALS Vial : 6 Sample Multiplier: 1


Instrument :
 BNA_M
 ClientSampleId :
 BGCA4

Quant Time: Nov 17 12:54:15 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM111622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 17 00:46:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.145	152	6578	0.400	ng/ul	0.00
4) Naphthalene-d8	10.963	136	19458	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.761	164	12483	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.496	188	26233	0.400	ng/ul	0.00
17) Chrysene-d12	21.654	240	18439	0.400	ng/ul	0.00
23) Perylene-d12	24.156	264	21259	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.453	96	33322	4.357	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.541	152	9464	0.293	ng/ul	0.00
18) Fluoranthene-d10	19.509	212	24830	0.407	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.495	88	1795	0.226	ng/ul#	48

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 0.226 ng/ul
 RT: 3.495 min Scan# 98
 Delta R.T. 0.000 min
 Lab File: BM037522.D
 Acq: 17 Nov 2022 12:10

Instrument :
 BNA_M
 ClientSampleId :
 BGCA4



Tgt Ion: 88 Resp: 1795

Ion	Ratio	Lower	Upper
88	100		
43	56.1	102.8	154.2#
58	52.3	67.3	100.9#

