

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111722\
 Data File : BM037530.D
 Acq On : 17 Nov 2022 17:01
 Operator : CG/JU
 Sample : N5622-10
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BGCB2

Quant Time: Nov 17 18:15:10 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM111622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 17 00:46:50 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.140	152	6475	0.400	ng/ul	0.00
4) Naphthalene-d8	10.963	136	19148	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.761	164	12175	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.491	188	25719	0.400	ng/ul	0.00
17) Chrysene-d12	21.654	240	18101	0.400	ng/ul	0.00
23) Perylene-d12	24.153	264	20984	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.453	96	37253	4.948	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.536	152	9888	0.311	ng/ul	0.00
18) Fluoranthene-d10	19.504	212	25352	0.424	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.491	88	1873	0.239	ng/ul#	66


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111722\
Data File : BM037530.D
Acq On : 17 Nov 2022 17:01
Operator : CG/JU
Sample : N5622-10
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BGCB2


Quant Time: Nov 17 18:15:10 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM111622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 17 00:46:50 2022
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 0.239 ng/ul
 RT: 3.491 min Scan# 91
 Delta R.T. -0.004 min
 Lab File: BM037530.D
 Acq: 17 Nov 2022 17:01

Instrument :
 BNA_M
 ClientSampleId :
 BGCB2



Tgt Ion:	88	Resp:	1873
Ion	Ratio	Lower	Upper
88	100		
43	74.6	102.8	154.2#
58	70.8	67.3	100.9

