

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111722\
 Data File : BM037540.D
 Acq On : 17 Nov 2022 23:42
 Operator : CG/JU
 Sample : N5622-15
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BGCA0

Quant Time: Nov 18 00:34:08 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM111622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 17 23:00:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.141	152	7514	0.400	ng/ul	0.00
4) Naphthalene-d8	10.958	136	21882	0.400	ng/ul	#-0.01
9) Acenaphthene-d10	14.756	164	13653	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.491	188	28486	0.400	ng/ul	0.00
17) Chrysene-d12	21.649	240	19774	0.400	ng/ul	0.00
23) Perylene-d12	24.139	264	23544	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.453	96	45646	5.225	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.536	152	12330	0.340	ng/ul	0.00
18) Fluoranthene-d10	19.503	212	29076	0.445	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.491	88	2200	0.242	ng/ul#	48


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111722\
Data File : BM037540.D
Acq On : 17 Nov 2022 23:42
Operator : CG/JU
Sample : N5622-15
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BGCA0


Quant Time: Nov 18 00:34:08 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM111622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 17 23:00:56 2022
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 0.242 ng/ul
 RT: 3.491 min Scan# 91
 Delta R.T. -0.004 min
 Lab File: BM037540.D
 Acq: 17 Nov 2022 23:42

Instrument :
 BNA_M
 ClientSampleId :
 BGCA0



Tgt Ion:	88	Resp:	2200
Ion	Ratio	Lower	Upper
88	100		
43	52.4	102.8	154.2#
58	55.2	67.3	100.9#

