

Data Path : Z:\HPCHEM1\BNA M\DATA\BM111815\
 Data File : BM003172.D
 Acq On : 18 Nov 2015 22:23
 Operator : UM/IZ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD0.451

Quant Time: Nov 19 03:19:58 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM02.2-EPA-SIM-BM111815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 19 03:17:25 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2 SURR	1,4-Dioxane-d8	0.427	0.445	-4.2	105	0.00
3	1,4-Dioxane	0.486	0.490	-0.8	103	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
5	Naphthalene	1.027	1.037	-1.0	103	0.00
6 SURR	2-Methylnaphthalene-d10	0.544	0.550	-1.1	102	0.00
7	2-Methylnaphthalene	0.669	0.680	-1.6	103	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
9	Acenaphthylene	1.665	1.623	2.5	100	0.00
10 C	Acenaphthene	1.406	1.421	-1.1	102	0.00
11	Fluorene	1.521	1.544	-1.5	102	0.00
12 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
13	Pentachlorophenol	0.094	0.095	-1.1	106	0.00
14	Phenanthrene	1.161	1.172	-0.9	101	0.00
15	Anthracene	0.972	0.960	1.2	101	0.00
16 SURR	Fluoranthene-d10	0.980	0.977	0.3	102	0.00
17 C	Fluoranthene	1.287	1.279	0.6	102	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
19	Pyrene	1.371	1.366	0.4	103	0.00
20	Benzo(a)anthracene	1.160	1.164	-0.3	108	0.00
21	Chrysene	1.306	1.348	-3.2	106	0.00
22 I	Perylene-d12	1.000	1.000	0.0	109	0.00
23	Benzo(b)fluoranthene	1.521	1.602	-5.3	116	0.00
24	Benzo(k)fluoranthene	1.438	1.287	10.5	93	0.00
25 C	Benzo(a)pyrene	1.353	1.386	-2.4	111	0.00
26	Indeno(1,2,3-cd)pyrene	1.612	1.745	-8.3	116	0.00
27	Dibenzo(a,h)anthracene	1.303	1.408	-8.1	116	0.00
28	Benzo(q,h,i)perylene	1.416	1.452	-2.5	110	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0